



Columbia Blood Agar Base

M144

Intended Use:

Columbia Blood Agar Base is used as an efficient base for preparation of blood agar, chocolate agar and for preparation of various selective and identification media and isolation of organisms from clinical and non clinical samples.

Composition**

Ingredients	Gms / Litre
Peptone, special	23.000
Corn starch	1.000
Sodium chloride	5.000
Agar	15.000
Final pH (at 25°C)	7.3±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 44.0 grams of in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C before adding heat sensitive compounds.

For Blood Agar: Add 5% v/v sterile defibrinated sheep blood to sterile cool base.

For Chocolate Agar: Add 10% v/v sterile defibrinated sheep blood to sterile cool base. Heat to 80°C for 10 minutes with constant agitation.

The medium can be made selective by adding different antimicrobials to sterile base.

For *Brucella* species: Add rehydrated contents of 1 vial of Brucella Selective Supplement (FD005) to 500 ml sterile molten base.

For *Campylobacter* species: Add rehydrated contents of 1 vial of Campylobacter Supplement- I (Blaser-Wang) (FD006) or Campylobacter Supplement- II, (Butzler) (FD007) or Campylobacter Supplement- III (Skirrow) (FD008) or Campylobacter Selective Supplement (FD090) or Campylobacter Supplement- VI (Butzler) (FD106) to 500 ml sterile molten base along with rehydrated contents of 1 vial of Campylobacter Growth Supplement (FD009) and 5-7% v/v horse or sheep blood.

For *Gardnerella* species: Add rehydrated contents of 1 vial of G.Vaginalis Selective Supplement (FD056) to 500 ml sterile molten base.

For Cocci: Add rehydrated contents of 1 vial of Staph-Strepto Supplement (FD030) or Strepto Supplement (FD031) or Streptococcus Selective Supplement (FD119) to 500 ml sterile molten base.

Principle And Interpretation

Columbia Blood Agar Base was devised by Ellner et al (4). This medium contains special peptone which supports rapid and luxuriant growth of fastidious and non-fastidious organisms. Also, this medium promotes typical colonial morphology; better pigment production and more sharply defined haemolytic reactions. Fildes found that Nutrient Agar supplemented with a digest of sheep blood supplied both of these factors and the medium would support the growth of *H. influenzae* (5,6). The inclusion of bacitracin makes the enriched Columbia Agar Medium selective for the isolation of *Haemophilus* species from clinical specimens, especially from upper respiratory tract (3). Columbia Agar Base is used as the base for the media containing blood and for selective media formulations in which different combinations of antimicrobial agents are used as additives.

Corn starch serves as an energy source and also neutralizes toxic metabolites. Sheep blood permits the detection of haemolysis and also provides heme (X factor) which is required for the growth of many bacteria. However it is devoid of V factor (Nicotinamide adenine dinucleotide) and hence *Haemophilus influenzae* which needs both the X and V factors, will not grow on this medium.

Columbia Agar Base with added sterile serum provides an efficient medium for *Corynebacterium diphtheriae* virulence test medium. After following the established technique for *C. diphtheriae*, lines of toxin-antitoxin precipitation are clearly visible in 48 hours. Many pathogens require carbon dioxide; therefore, plates may be incubated in an atmosphere containing approximately 3-10% CO₂.

Precaution: *Brucella* cultures are highly infective and must be handled carefully; incubate in 5-10% CO₂. *Campylobacter* species are best grown at 42°C in a micro aerophilic atmosphere. Plates with *Gardenerella* supplements plates should be incubated at 35°C for 48 hours containing 7% CO₂ (5).

Type of specimen

Clinical samples : blood, respiratory exudates.

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (9,10).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. Certain fastidious organisms like *Haemophilus influenzae* may not grow on the medium, blood supplementation may be required.
2. As this medium have a relatively high carbohydrate content, beta-haemolytic Streptococci may exhibit a greenish haemolytic reaction which may be mistaken for the alpha haemolysis.
3. Carry out confirmatory tests of all the colonies.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Basal medium: Light amber coloured clear to slightly opalescent gel.

After addition of 5%w/v sterile defibrinated blood : Cherry red coloured opaque gel forms in Petri plates

Reaction

Reaction of 4.4% w/v aqueous solution at 25°C. pH : 7.3±0.2

pH

7.10-7.50

Growth Promotion Test

In accordance with the harmonized method of USP/EP/BP/JP.

Cultural Response

Cultural characteristics observed with added 5% w/v sterile defibrinated blood, after an incubation at 35-37°C for 24-48 hours.

Cultural Response

Organism	Growth	Inoculum (CFU)	Recovery	Haemolysis
Cultural Response				
<i>Neisseria meningitidis</i> ATCC 13090	luxuriant	50-100	≥70%	none
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 25923 (00034*)	luxuriant	50-100	≥70%	beta / gamma
<i>Staphylococcus epidermidis</i> ATCC 12228 (00036*)	luxuriant	50-100	≥70%	gamma
<i>Staphylococcus aureus</i> subsp. <i>aureus</i> ATCC 6538 (00032*)	luxuriant	50-100	≥70%	beta / gamma
<i>Streptococcus pneumoniae</i> ATCC 6303	luxuriant	50-100	≥70%	alpha
<i>Streptococcus pyogenes</i> ATCC 19615	luxuriant	50-100	≥70%	beta
<i>Clostridium sporogenes</i> ATCC 19404 (00008*)	luxuriant	50-100	≥50 %	
<i>Clostridium sporogenes</i> ATCC 11437	good-luxuriant	50-100	≥50 %	
<i>Clostridium perfringens</i> ATCC 13124 (00007*)	luxuriant	50-100	≥50 %	
<i>Clostridium perfringens</i> ATCC 12934	luxuriant	50-100	≥50 %	

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store below 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (7,8).

Reference


- Bailey R. K., Voss J. L. and Smith R. F., 1979, J. Clin. Microbiol., 9 ; 65-71
- Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
- Chapin K. C. and Doern G. V., 1983, J. Clin. Microbiol., 17:1163.
- Ellner P. P., Stoessel C. J., Drakeford E. and Vasi F., 1966, Am. J. Clin. Pathol., 45:502.
- Fildes P., 1920, Br. J. Exp. Pathol., 1:129.
- Fildes P., 1921, Br. J. Exp. Pathol., 2:16.
- Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
- Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
- Food and Drug Administration, 2001, "4237. "Eqo r gpf kwo "qh"O gyj qf u"lqt"y g"O letqdkmqi kecn"Gzco kpcvqp"qh" Foods, 5th Ed., American Public Health Association, Washington, D.C.
- Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.