

CERTIFICATE of ANALYSIS
LYOPHILISED & ABSORBED ANTI-RABIES
NUCLEOCAPSID CONJUGATE : Code 3572112

Lot n° : 3K0078

Control date : 2023/10/31

Manufacturing date : 2023/10/20

Expiration date : 2025/10/19

I COMPOSITION OF THE KIT :

Reagent	Number of vials	Lot n°
Rabbit IgG anti-rabies nucleocapsid coupled with fluorescein isothiocyanate (FITC) – Lyophilised, absorbed on mice brain cells	4x3 ml	3K0078
Insert	1	/

II QUALITY CONTROL

Control	Norm	Result
<u>Specificity</u>		
- with infected brain cells	Specific fluorescence	Conform
- with non-infected brain cells	No fluorescence	Conform
<u>Sensitivity</u>		
- Intensity of fluorescence with rabies infected cells	Expected intensity	Conform
- Stability after rehydration and incubation 4 days at 37°C	No free fluorescence detected	Conform

Date : 2023/12/18

Signature :



K. DEBRUYNE

Accepted ☒ Refused ☐

