

**Specific details and technical standards**

**LOT: “Disposable FFP2/N95 NR Filtering half Mask with Exhalation Valve (Particulate respirators), to protect against particles for the National Reference Laboratory in TB Microbiology”**

№	Name of product	Minimal technical requirements	Total Quantity	Packing	Manufacturer (name, country); Model number; Description; Compliance
1	<b>Disposable FFP2/N95 NR Filtering half Mask with Exhalation Valve, (Particulate respirators), to protect against particles.</b>	<ul style="list-style-type: none"> <li>• To be used for medical purposes for the protection of personnel working in Tuberculosis laboratories;</li> <li>• Can help protect against airborne micro-organisms such as mycobacterium tuberculosis, and SARS;</li> <li>• Stretchable <b>headbands</b> easy to put on and provide a secure fit;</li> <li>• Adjustable nose clip;</li> <li>• Protection Level – FFP2 (N95); minimum 94-95% according to “Chapter 7.2, WHO Tuberculosis Laboratory Biosafety Manual. (the Year 2012)”.</li> <li>• Assigned Protection Factor - min. 10;</li> <li>• Includes an exhalation valve fitted to the front of the respirator which minimizes heat and moisture build-up.</li> <li>• <b>At least 3 samples of the offered product to conduct Fit Test (sealing property test). Will be considered only the products that pass the Fit Test.</b></li> </ul>	5200 pcs.	15 pcs/pack	<p><b>Manufacturer: Name - BLS SRL; Country –Italy</b>  <b>Model number - BLS 129BW FFP2 R D:</b></p> <ul style="list-style-type: none"> <li>• <b>ELASTICS Thermoplastic elastomer (TPE)</b></li> <li>• <b>NOSE CLIP Reinforced Polypropylene (PP)</b></li> <li>• <b>VALVE Polypropylene (PP); Rubber</b></li> <li>• <b>GASKET Foamed polymer</b></li> <li>• <b>FILTERING MATERIAL Polypropylene (PP)</b></li> </ul> <p>- <b>Certified according to European Regulation 2016/425 (Personal Protective Equipments)</b>  - <b>Certified as PPE of III category, in presumption of conformity to harmonized standard EN 149:2001+A1:2009</b>  - <b>Certified and controlled according to Annex D by Italcert S.r.l. (Notified Body n°0426)</b>  - <b>CE marked</b></p>

	<p><b>Mandatory requirements (this certificates/letters must be included in the offer)</b></p>	<ul style="list-style-type: none"> <li>• Manufacturer's ISO 9001 certificate;</li> <li>• CE standard EN149:2001+A1:2009;</li> <li>• Written confirmation (certificate) that the offered products are intended for use in Tuberculosis laboratories;</li> <li>• All presented certificates should be valid;</li> <li>• Original manufacturer packing for all items;</li> <li>• Each product packed individually;</li> <li>• Warranty: the remaining shelf life should be not less than 80% from the total one.</li> </ul>
--	--	--

- All goods will be tested by the beneficiary. The goods will be accepted and the acts of receipt will be signed upon positive testing performed by the beneficiary.

NAME OF SUPPLIER Absolutely Safety SRL

Authorized Signature and Stamp Dubic Vera, Director



Place: mun. Chisinau, Moldova

Date: 11.08.2022

