



- Front loaded
- 2 body capacity with single door system
- 304 grade stainless steel inner and outer surface
- Cooling unit and evaporator
- Hermetic type condensing unit with ventilation system
- Drainage system
- High quality walls with sandwich panel in order to preserve temperature
- Polyurethane panels with 80mm thick for insulation
- Tray with handle made of 304 grade stainless steel
- Tray is sliding on a special rail system to prevent tip over
- Digital control unit and alphanumeric temperature display
- Programmable temperature control
- Working temperature +/- 10 °C
- Alarm life-sensor / sound signal system
- Visual interior lighting
- Lockable swivel castors

TECHNICAL DETAILS	
Length:	236 cm
Depth:	87 cm
Height:	160 cm



Client Name and address

Müşteri ismi ve adresi

## OPTİUM MEDİKAL LİMİTED ŞİRKETİ

KÖRKÜN MAH. HİDAYET CAD. NO:23/1 OĞUZELİ GAZİANTEP TÜRKİYE

Standard

Standard

## ISO 13485:2016 Quality Management Systems-Medical Devices

INSPECT assessed and approved that related firm meets the requirements of the designated standard at related scope  
INSPECT ilgili kuruluşun yukarıda belirtilen standardın gereklerine uygunluğunu ilgili kapsamda tetkik etmiş olup onaylamaktadır.

Scope

Kapsam

HASTANE EKİPMANLARI VE MOBİLYALARI ÜRETİMİ, AR-GE VE SATIŞI

MANUFACTURING, R&amp;D AND SELLING OF HOSPITAL FURNITURES AND EQUIPMENTS

EA/Category Code: 23

Initial issue date: 16.Apr.2021

Issue date of this certificate: 08.Apr.2022

Validity of this certificate: 15.Apr.2023

Recertification date: 15.Apr.2024

Certificate No: 21.02.337-MD

  
General ManagerTo verify the validity of this certificate please visit: [www.inspect.com.tr](http://www.inspect.com.tr)Bu sertifikanın geçerliliğini kontrol etmek için lütfen [www.inspect.com.tr](http://www.inspect.com.tr)'yi ziyaret ediniz

The validity of this certificate is depend on the success of surveillance assessments


Bu belgenin geçerliliği ara denetimlerin başarılı olmasına bağlıdır.

INSPECT Uluslararası Belgelendirme ve Gözetim Hiz. Tic. Ltd. Şti. / TURKEY  
[www.inspect.com.tr](http://www.inspect.com.tr) - Belge periyodu 3 yıldır. / Certificate period is 3 years.

## EC CONFORMITY DECLARATION

Date of issue: 22.01.2022

Conformity declaration issued by:

Commercial name	Optium Medikal Ltd.Sti.
Registered address	Körkün Mah. Hidayet Cad. No:23 OĞUZELİ GAZİANTEP TURKEY
Telephone	+90 850 807 6284
Fax	+90 850 807 6284
E-mail	<a href="mailto:info@optium.com.tr">info@optium.com.tr</a>
Brand	
Products	INOXY 155 Mortuary Unit Single INOXY 153 Mortuary Unit Double INOXY 154 Mortuary Unit Triple
Description and function designation:	Mortuary Units, including all applicable accessories
Classification of the product as the medical device:	Class I nonsterile, without measuring function

### A) Declaration

I declare that the said products are safe under the conditions of common use in compliance with the instructions and that measures have been taken to ensure the conformity of all the products brought to market with basic requirements of directives related thereto, stated in paragraph B.

### B) Fulfilled technical requirements

These products' characteristics comply with the technical parameters related to it and stated in MDD 93/42/EEC as amended, which stipulates the technical parameters for healthcare products.

### C) Means of assessing conformity

Conformity was assessed by the procedure stated MDD 93/42/EEC, Annex VII

### D) Used standards

The said products fulfill the requirements of these harmonized technical standards which were used for assessing of conformity.



optium MEDİKAL LTD. ŞTİ.  
Körkün Mah. Hidayet Cad. No: 23/1  
OĞUZELİ GAZİANTEP  
Tic.Sic.No: 55211 Mersis No: 0644067549100001  
SUSURCU V.D. : 644 067 54 91