

Specificații tehnice

Numărul procedurii de achiziție nr. 21560117/ ocids-b3wdp1-MD-1770110932848 din 16.02.2026						
Obiectul achiziției: Reactivi de laborator pentru anul 2026						
Denumirea bunurilor/serviciilor	Denumirea modelului bunului/serviciului	Țara de origine	Produ-cătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant	Standarde de referință
1	2	3	4	5	6	7
Lotul 5. Reactivi si consumabile necesare pentru Aparat de biochimie urinară Uryxxon Relax						
Medi-Test Combi 11	Medi-test Combi11, 93060	Germania	Macherey-Ngel	teste penrtu determinarea in urina: BLD, UBG, BIL,PRO, NIT, KET, ASC, GLU, pH, SG, LEU. Compatibili cu apatarul URYXXON Relax. Ambalaj: 1 set* 100 teste	teste penrtu determinarea in urina: BLD, UBG, BIL,PRO, NIT, KET, ASC, GLU, pH, SG, LEU. Compatibili cu apatarul URYXXON Relax. Ambalaj: 1 set* 100 teste	CE, ISO
Hirtie pentru aparat URYXXON Relax	14988	Italia	Ceracarta		Termohirtie pentru aparat de biochimie urinară Uryxxon Relax	CE, ISO
Lotul 6. Reactivi si consumabile necesare pentru Stat Profile Prime-Plus						
Micro Senzor Card p/u 1/100 teste compatibil cu analizatorul Stat Profile Prime Plus	61604	SUA	Nova Biomedical	Stat Profile Prime-Plus Sensor Card: w/ COOx (standart) Ambalaj: 1 buc/100 teste pH, pCO2, pO2, SO2, Hct, tHb, Na, K, Cl, iCa, iMg, Glu, Lac, O2Hb, COHb, MetHb,HHb, tBil, HbF, TCO2	Stat Profile Prime-Plus Sensor Card: w/ COOx (standart) Ambalaj: 1 buc/100 teste pH, pCO2, pO2, SO2, Hct, tHb, Na, K, Cl, iCa, iMg, Glu, Lac, O2Hb, COHb, MetHb,HHb, tBil, HbF, TCO2	CE, ISO
Cartus de Calibrare compatibil cu analizatorul Stat Profile Prime Plus	57825	SUA	Nova Biomedical	Stat Profile Prime Plus® Calibrator Cartridge pentru 100 teste	Stat Profile Prime Plus® Calibrator Cartridge 1buc/ 100 teste	CE, ISO
Ampuled Controls BG	57844	SUA	Nova Biomedical	Stat Profile Prime Plus® Ampuled Controls BG, COOX Levels 1, 2, 3; Anbalaj :10amp. x 3ml;	Stat Profile Prime Plus® Ampuled Controls BG, COOX Levels 1, 2, 3; 10amp. x 3ml.	CE, ISO
Ampuled Controls Chemistry	57845	SUA	Nova Biomedical	Stat Profile Prime Plus® Ampuled Controls Chemistry Levels 4,5; 10amp. x 3ml;	Stat Profile Prime Plus® Ampuled Controls Chemistry Levels 4,5; 10amp. x 3ml	CE, ISO
Sensor de referinta (Pompa-Tub)	57823	SUA	Nova Biomedical	Reference Sensor- Prime-Plus (w/ W/R Pump Tubing) (CLINICAL). Stat Profile Prime-Plus	Reference Sensor- Prime-Plus (w/ W/R Pump Tubing) (CLINICAL)	CE, ISO
Cartus blank (BUN/crea)	58379	SUA	Nova Biomedical	Stat Profile Prime Plus®	Prime-Plus Sensor Card- BLANK RENAL Sensor Card: (Clinical)	CE, ISO
Rola hirtie	49200	SUA	Nova Biomedical	Printer Paper termosensibila compatibil cu Analizatorul Stat Profile Prime Plus	Printer Paper termosensibila compatibil cu Analizatorul Stat Profile Prime Plus	CE, ISO

Lotul 8 Reactivi si consumabile necesare pentru Stat Fax 1904						
Lampa penrtu aparat	Lampa 10 W 6V G4	SUA	Awareness Tehnology	10W 6V Stat Fax 1904, 3300 Awareness Tehnology	10W 6V Stat Fax 1904, 3300	ISO
Lotul 10						
Lampa penrtu microscop	Lampa p-u microscop, 20W6V	Germania	Osram	20W 6V	20W 6V	ISO

Semnat: _____ Numele, Prenumele: Tudor Ceaicovschi În calitate de: Director general

Ofertantul: "GBG-MLD" SRL Adresa: MD-2005, mun. Chişinău, str. Albişoara 64/2