


Test Report

– Eppendorf PCR clean –

Protein LoBind Tubes 5.0 mL



100 EPPENDORF TUBES®
PCR clean
centrifugation up to 25 000 x g

LOT M212994M

2028-07-28

Order no.: 0030 108.302
Eppendorf SE
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030108302
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

• **Human DNA**

Method: *real-time* PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• **DNase**

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable (LOD: 1.0×10^{-6} Kunitz units)

• **PCR inhibition**

Method: *real-time* PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• **RNase**

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 08.09.23

Testing of the above described lot showed conformity within the limits of detection.

12.09.23
Date of release

epp 2703-3736
Registration number

i. V. B. Weigel
Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1