

TO WHOM IT MAY CONCERN

Saint Jean, May 1st, 2024

Our ref.: LET-2405-1548

MANUFACTURER NOISE STATEMENT – STERIPLUS™ & STERISHRED®

We undersigned, the company **Tesalys**, 7 rue du Cassé, 31240 Saint-Jean (Toulouse), France, represented by Mr. Miquel Lozano, President, by this letter **STATES** that:

The noise measurements performed at 1 m. distance during the treatment cycle of our **STERIPLUS™ and STERISHRED®** shredder-sterilizers for infectious waste treatment show the following results:

Average noise level during the cycle: < 60 dBa

Name: *Miquel Lozano*
Address: *7, rue du Cassé, 31240 Saint-Jean*
Title: *President*
Date: *01/05/2024*



Tesalys
Making biomedical waste safe
P.A. du Cassé - 7 rue du Cassé
31240 SAINT JEAN
Tél : +33 5 62 10 18 91
www.tesalys.fr
Siret: 750 475 527 00032 - TVA Intracom. FR 42 750 475 527