

Sphygmomanometer LD-71



Professional Sphygmomanometer With Phonendoscope.

Features:

- Metal Aneroid Manometer.
- **Stethoscope Included.**
- Nylon Cuff for Upper Arm Circumference 25-36 cm.
- Check Valve Screen Filter from Clogging Up with Dust.
- Metal Needle Air Valve.
- The cuff has size marks and a metal fixing ring for convenient use.

Features	
Model	LD-71
Stethoscope included	OK
Stethoscope Attached to the Cuff	No
Manometer material	Metal
Manometer dial diameter	45 mm
Adaptable for left or right hand	No
Valve type	Metal Needle Air Valve
Measuring Range (pressure)	0-300 mmHg
BLOWER WITH NON-RETURN VALVE	
Check valve filter	ok
CUFF	
Cuff model	LD Cuff N2AR
Cuff size	25-36 cm
With a Metal fixing Ring	OK
Number of tubes	2
Cuff material	Nylon
Cuff case material	TPU
MEASUREMENT	
Measuring Range (pressure)	0-300 mmHg
Measuring Accuracy (cuff pressure)	+/- 3 mmHg
Operating Conditions (temperature)	10 °C to 40 °C
Operating Conditions (relative humidity)	85% and below
OTHER	
Weight	328 g
Dimensions	115 x 185 x 75 mm
Aneroid Gauge, Nylon D-Ring Cuff (Arm Circumference 25-36 cm), Bulb with Valve, Stethoscope, Carrying Case, Operating Instructions, Warranty Card.	



Declaration of Conformity

Manufacturer: Little Doctor Electronic (Nantong) Co., Ltd.
No. 8, Tongxing Road, Economic and Technical Development
Area, 226010 Nantong, Jiangsu, People's Republic of China

Facility: Little Doctor Electronic (Nantong) Co., Ltd.
No. 8, Tongxing Road, Economic and Technical Development
Area, 226010 Nantong, Jiangsu, People's Republic of China

EU Representative: Little Doctor Europe Sp. z o. o.
Zawila Str. 57G, 30-390, Krakow, Poland

Product name: Aneroid Sphygmomanometers

Model number: LD-71

UMDNS Code 16156

Classification Class I with measuring function (MDR, Annex VIII Rule 1)

UDI-DI 697368616001X-697368616999X

Conforming Assessment Route Annex IX

We herewith declare that the above mentioned products meet the provisions of the following EC Council Regulation and Standards. All supporting documentations are retained under the premises of the manufacturer. We are exclusively responsible for the Declaration of Conformity.

General applicable regulation: Medical Device regulation: COUNCIL REGULATION
(EU)2017/745

Notified Body: TÜV SÜD Product Service GmbH,
Certification Body, Ridlerstraße 65, 80339 Munich, Germany

Identification No: 0123

EC Certificate: G2M 071993 0016 Rev.00

Expiration date of the Certificate: 2024-05-26

Data CE mark was affixed: 2006-07

Place: People's Republic of China

Date: 2021-06-15

Signature:

Name: GAO JIAWEN

Position: Director





Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Production Quality Assurance System
Directive 93/42/EEC on Medical Devices (MDD), Annex V
(Devices in Class IIa, IIb or III)

No. G2 071993 0017 Rev. 00

Manufacturer:

**Little Doctor Electronic
(Nantong) Co., Ltd.**

No.8, Tongxing Road
Economic & Technical Development Area
226010 Nantong, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Facility(ies):

Little Doctor Electronic (Nantong) Co., Ltd.
No.8, Tongxing Road, Economic & Technical Development Area,
226010 Nantong, Jiangsu, PEOPLE'S REPUBLIC OF CHINA

Product Category(ies):

**Blood Pressure Monitor,
Digital Thermometer, Ultrasonic Nebulizer,
Compressor Nebulizer , Dental Oral Irrigator**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex V. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class IIb and III devices an additional Annex III certificate is mandatory. See also notes overleaf.

Report No.: SH1940401

Valid from: 2020-02-25
Valid until: 2024-05-26

Date, 2020-02-25

Christoph Dicks
Head of Certification/Notified Body



Product Service

Certificate

No. Q5 071993 0018 Rev. 01

Holder of Certificate: **Little Doctor Electronic
(Nantong) Co., Ltd.**

No.8, Tongxing Road
Economic & Technical Development Area
226010 Nantong, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Facility(ies):

Little Doctor Electronic (Nantong) Co., Ltd.
No.8, Tongxing Road, Economic & Technical Development Area,
226010 Nantong, Jiangsu, PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate:

Design and Development,
Production and Distribution of
Aneroid sphygmomanometers,
Blood pressure monitor, Nebulizer,
Stethoscopes, Dental oral irrigator, Examination light
Production and distribution of Digital thermometer

Applied Standard(s):

EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see:
www.tuvsud.com/ps-cert?q=cert:Q5 071993 0018 Rev. 01

Report No.: SH2040401

Valid from: 2020-10-12

Valid until: 2023-09-30

Date, 2020-10-13

Christoph Dicks
Head of Certification/Notified Body