



EC DECLARATION OF CONFORMITY

MANUFACTURER:

DIESSE DIAGNOSTICA SENESE SPA
VIA DELLE ROSE 10
53035 MONTERIGGIONI (SI),
ITALY



EUROPEAN REPRESENTATIVE:

//

GENERIC NAME

AUTOMATIC ANALYSIS INSTRUMENT FOR IMMUNOMETRY

PRODUCT:

CHORUS TRIO RAS

CODE:

81220

TECHNICAL DATA:

110-220 Vac (50-60 Hz); Pwr: 350 VA

CLASSIFICATION:

IVD NOT IN ANNEX II OR SELF-TESTING IVD

CONFORMITY ASSESSMENT ROUTE:

ANNEX APPLIED N° III EXCLUDING (6)
ESSENTIAL REQUIREMENTS OF ANNEX I

WE HEREWITH DECLARE THAT THE DESIGN, TYPE OF MANUFACTURE OF THE IN VITRO MEDICAL DIAGNOSTIC DEVICE DESCRIBED ABOVE AND THE VERSION DISTRIBUTED ON THE MARKET, CONFORMS TO THE 98/79/EEC DIRECTIVE RELEVANT TO THE IN VITRO MEDICAL-DIAGNOSTICS DEVICES (IVD)

THIS CERTIFICATE WILL LOSE ITS VALIDITY IN THE EVENT OF:

- MODIFICATIONS MADE TO THE MACHINE IN QUESTION WITHOUT OUR AUTHORIZATION
- INCORRECT USE OF THE INSTRUMENT
- TECHNICAL INTERVENTIONS PERFORMED BY UNAUTHORIZED PERSONNEL
- INSTALLATION OF NON-ORIGINAL SPARE PARTS.

THE PRODUCT CONFORMS, AS A WHOLE AND IN ITS PARTS, WITH THE FOLLOWING STANDARDS AND THEIR AMENDMENTS:

EN 61010-1:2010 SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 1: GENERAL REQUIREMENTS

EN 61010-2-081:2015 SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-081: PARTICULAR REQUIREMENTS FOR AUTOMATIC AND SEMI-AUTOMATIC LABORATORY EQUIPMENT FOR ANALYSIS AND OTHER PURPOSES

EN 61010-2-101:2002 SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - PART 2-101: SAFETY REQUIREMENTS FOR IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

EN 61326-1:2013

ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 1: GENERAL REQUIREMENTS

EN 61326-2-6:2006

ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE - EMC REQUIREMENTS - PART 2-6: PARTICULAR REQUIREMENTS - IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT

AND THEREFORE MEETS THE ESSENTIAL REQUIREMENTS OF THE FOLLOWING COMMUNITY DIRECTIVES AND THEIR AMENDMENTS:

LOW VOLTAGE DIRECTIVE (2014/35/EU)

ELECTROMAGNETIC COMPATIBILITY DIRECTIVE (2014/30/EU)

RESTRICTION OF HAZARDOUS SUBSTANCES IN ELECTRICAL AND ELECTRONIC EQUIPMENT DIRECTIVE – ROHS2 (2011/65/EU)

NOTIFIED BODY:

NOT NECESSARY

(EC) CERTIFICATE:

N.A.

START OF CE-MARKING:

JUNE 2018

REVISION:

3

PLACE, DATE OF ISSUE:

MONTERIGGIONI, 5 FEBRUARY 2021

EXPIRY DATE:

25 MAY 2022

THE PRESENT EC DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE AND EXCLUSIVE RESPONSIBILITY OF THE MANUFACTURER.

SIGNATURE:



CHIARA MUZZI
QUALITY & REGULATORY MANAGER

THIS IS THE CERTIFIED COPY OF THE ORIGINAL DOCUMENT STORED IN ARCHIVE OF DIESSE DIAGNOSTICA SENESE SPA

ISSUED: MONTERIGGIONI,

05/02/2021



MAGDALENA STOCZKO
REGULATORY SUPERVISOR