

S/N: 002007

**No.: ICR Polska/VC/MH210101/A1**replaces **ICR Polska/VC/MH210101****Name and address of Applicant**DÖRT-ATIP MALZEMELERİ SANAYİ İTHALAT İHRACAT TİCARET LİMİTED ŞİRKETİ  
BALIKHİSAR MH. KÖYÜÇİ SERPMELERİ NO:795/A AKYURT / ANKARA / TURKEY**Name and address of manufacturer:**DÖRT-ATIP MALZEMELERİ SANAYİ İTHALAT İHRACAT TİCARET LİMİTED ŞİRKETİ  
BALIKHİSAR MH. KÖYÜÇİ SERPMELERİ NO:795/A AKYURT / ANKARA / TURKEY**Product name:**

BOWIE&amp;DICK TEST STRİBİ VE PAKETİ / BOWIE&amp;DICK TEST STRIP AND PACKAGE

**Product types:**

n/a

**Product trademark:**  
4a medical®  
producing health for the world

This document confirms that the product sample meets the requirements of the following standards:

EN ISO 11140-4:2007

EN ISO 11140-1:2009

The assessment process has been carried out in accordance with individual rules and conditions agreed with the applicant. Evaluation has been carried out in accordance with:

**Test reports:**

partial report file

**Tests conducted by:**

Saniter Gıda – Çevre Bilimi Gözetim Ve Mühendislik Hiz. Tur. Tic. A.Ş.

**Issue date:**

13.10.2021

**Initial issue date**

09.02.2021

**Expiration date:**

08.02.2026


**Remarks:**

Document refers to the above mentioned product and its conformity in regards of above mentioned standard(s) was proven on test sample

Document was issued on voluntary basis and does not imply meeting all essential requirements, assessment of the series-production or any other restricted Notified Bodies conformity assessment procedure appropriate for the product  
Document status can be checked: <https://cert.icrpolska.com/>**CE marking remarks:**

CE mark is not sanctioned by the following document

CE mark given here as reference, can be only use by the manufacturer after applying all essential requirements from relevant directives

ICR Polska Co. Ltd.  
[www.icrpolska.com](http://www.icrpolska.com)  
[icrpolska@icrq.com](mailto:icrpolska@icrq.com)Director: Rafał Kalinowski  
Warsaw, 13.10.2021

Edition: 4.1.0.A of 18.01.2021