

Certificate of Analysis

Certificat d'Analyse

applied biosystems
by Thermo Fisher Scientific

VetMAX™ West Nile Virus Kit



50 Tests

LOT

WNPEXO-026



WNPEXO50



2024-06-24



2025-06-20

KIT CONFIGURATION CONFIGURATION DU KIT

COMPONENT DESCRIPTION <i>Description du composant</i>	REF	UNIT	LOT	
3 - Mix West Nile	MPEWN	2 x 500µL	MPEWN-026	2025-06-20
4a - EPC West Nile	EPCWN	90 µL	EPCWN-026	2025-06-24
5 - IPC West Nile	IPCWN	250 µL	IPCWN-026	2025-06-24

INSTRUCTIONS FOR USE

Notices d'utilisation



French MAN0008914 RevB.0
English MAN0008915 RevB.0

QUALITY CONTROL RESULTS RESULTATS DE CONTROLE QUALITE

EPC <i>EPC</i>	Already extracted EPC <i>EPC déjà extrait</i>	Thermocycler used for QC <i>Thermocycleur utilisé pour le CQ</i>
4a - EPC West Nile	C _{t QC} WN = 26	ABI 7500 (Applied Biosystems)

IPC <i>IPC</i>	Column extraction <i>Extraction colonne</i>	Thermocycler used for QC <i>Thermocycleur utilisé pour le CQ</i>
5 - IPC West Nile	C _{t QC} IPC = 29	ABI 7500 (Applied Biosystems)
Endogenous IPC <i>IPC endogène</i>	C _{t QC} IPC < 40	ABI 7500 (Applied Biosystems)

C_{t QC}: Indicative Ct of the Quality Control *C_t indicatif du Contrôle Qualité*

SUPPLEMENTAL INSTRUCTIONS - Available upon request to Eurotech@thermofisher.com

Notices d'utilisation complémentaires - Disponibles sur demande auprès de Eurotech@thermofisher.com

Amplification Protocol <i>Protocole d'amplification</i>	Nucleic acid purification Protocol <i>Protocole de purification d'acides nucléiques</i>
N/A	French MAN0008916 RevB.0



Laboratoire Service International
6 allée des Ecureuils
Parc Tertiaire du Bois-Dieu
69380 Lissieu - France
www.thermofisher.com/support

Quality Assurance

Date: 08 July 2024