





COMPLIANCE DOCUMENT

To whom may concern:	
Center for Centralized Public Procurement in Health	
& MINISTRY OF HEALTH AND LABOUR AND SOCIAL PROTECTION, REPUBLIC OF MOLDOVA	
We Modül Grup Mühendislik Elektronik Medikal İnşaat Eğitim ve Bilişim Sanayi Ticaret Limited Şirketi Alınteri Bulvarı Gül 86 Yapı Koop. C Blok No:1/75 Ostim, Yenimahalle, Ankara, Turkey	
Do confirm, that the offered transport ventilators	Brand:Oxivent Model: Oxivent Oxi4Plus
Offered in cooperation with ROMEDCOM SRL for the public procurement	# 21040135 Lot 8 3 Units
Complies with the following technical specification:	S UNITS
REQUIRED:	OFFERED:
Adult transport ventilator (basic features)	Comply
Parameter Specification	Comply
Adult patient type	Patient Type: Adult, Pediatric, Optional Neonatal
Game control / settings Total volume 100-1,000	Tidal volume: 50 to 2500 ml (1 to 2500 ml with optional neonatal kit) Inspiration flow: Adult / pediatric: 0-200 l/min
Inspiration flow> 60 L / min Pressure level yes	Comply
Pressure in ramps yes	Comply
Respiratory rate 0-60 rpm	Respiratory rate: 1 to 120 b/min
Inspired time 0-2 s	Inspiratory time: 0.10 to 12 s
Ratio I: E 1: 1 to 1: 3	I:E ratio: 1:10 to 4:1
FiO2, 40-100% PEEP, cm H2O 0-20	FiO2: 21% to 100% Peep: 0 to 25 cmH2O
PEEP, cm H2O 0-20 Pressure support yes	Pressure support: 0 to 60 cmH2O, added to PEEP/CPAP
Trigger mechanism Flow or pressure	Comply
Operating modes A / C mode A / C by respiratory volume yes	Comply
A / C after after respiratory pressure yes	Comply
SIMV module SIMV respiratory volume yes	Comply Comply
SIMV respiratory pressure yes SIMV pressure support yes	Comply
SIPAP / spontaneous CPAP mode support pressure yes	Comply
Non-invasive ventilation yes	Comply
Apnea-backup mode is	Comply
Monitors / displays Maximum inspiratory pressure yes	Comply
PEEP pressure yes Total volume yes	Comply Comply
Respiratory rate yes	Comply
Time inspires yes	Comply
Patient alarms Low inspiratory pressure yes	Comply
High pressure yes	Comply
High continuous pressure / occlusion yes	Comply
Disconnected respiratory circuit yes Equipment alarms No gas supply yes	Comply Comply
No power supply yes	Comply
System error yes	Comply
Discharged battery yes	Comply
Self-diagnosis failed yes	Comply
TFT or LED LCD display yes Power supply 220 V mains, 50 Hz yes	Comply, 7.1 inch LCD screen Power consumption: 100-240 V 50-60 Hz - 60 Watt
Internal battery yes	Comply
Operating time 2 h	Operating time: Standard 4 hours
Total weight with accessories ≤ 8 kg	Weight: 4 kg
Accessory	Standard Accessories: Disposable flow sensor – 1 pcs, disposable patient valve – 1 pcs, 1.5 m. oxygen
Reusable adult respirator 2 sets. Reusable adult respirator 2 sets.	hose with quick coupling with O2 probe – 1 pcs, 220V to 12 V adapter –1 pcs, O2 Cylinder connection
Reusable adult respirator 2 sets. Disposable adult antibacterial filters 100 pcs.	hose – 1 pcs, Carrying bag (blue) – 1 pcs, Disposable bacteria Filter - 1 pcs, disposable breathing circuit - 1 pcs.
Reusable adult flow sensor 2 pcs.	Extra accessories:
Reusable adult test lung 1 pc.	Reusable patient circuit
Support bracket for fixing ventilator to the stretcher yes	Disposable bacterial filter
Oxygen cylinder with reducer and manometer, min. 2 liters 2 pcs.	Reusable adult flow sensor
	Adult Test Lung- 1 liter Aluminum oxygen cylinder (2 liters)
Manufacturer's	
Representative Stamp and Signature:	
ALINTERİ BULVARI GÜL-86 SİTESİ 1151.SK. C BLOK	PHONE: +90 312 354 0373 www.oxivent.com.tr
NO:75 OSTÍM / ANKARA / TURKEY	FAX : +90 312 354 7545 info@oxivent.com.tr