

Expiration date : 2025/02/17

Lot n° : 3K0049

: 2023/08/22

(c) conforming
(nc) not conforming

C / NC

1) PACKAGING CONFORMANCE

Control of labels	C
Expiration date	C
Lot number	C

2) PACKAGE INSERT CONFORMANCE

With kit reagents	C
-------------------	-------	----------

3) KIT CONFORMANCE

C

	Qt	LOT	Exp. Date
Number of components per kit	19		
R1 Microplates	2	3H0084	2025/02/17
R2 Washing solution*	1	3J0115	2025/03/05
R3 Negative control	1	3J0071	2025/03/23
R4a 0,5 UE/ml Positive control	1	3J0048	2025/03/23
R4b 4 UE/ml Positive control	1	3J0048	2025/03/23
R6 Sample diluent	2	3J0072	2025/03/23
R7 Concentrated conjugate	1	3J0049	2025/03/24
R8 Substrate buffer WW2*	1	3J0631	2025/03/19
R9 Chromogen WW2*	1	3H0345	2025/02/24
R10 Stopping solution 1N*	1	3E0569	2025/05/22
Package insert	1		
Adhesive plate sealers	6		

* R2, R8, R9 and R10 are generic reagents

Expiration date : 2025/02/17

Lot n° : 3K0049

4) PHYSICAL PRODUCT CONTROL

C / NC

* All criteria and parameters (aspect, volumes, pH and conductivity) were checked and found conforming.

C

5) BACTERIOLOGICAL CONTROL

* Each reagent to be tested was checked and found conforming.

C

6) ACTIVITY CONTROL

C

The following results were obtained with generic reagents:

R2 : 3J0115 R8 : 3J0631
R9 : 3H0345 R10 : 3E0569

Batch of OIE reference used for calibrations: 2014-3

CONFORMING

~~NOT CONFORMING~~

Steenvoorde the 2023-10-23

K. DEBRUYNE

