# Sample Cup



REF	CONTENT	SYSTEM
INLI	CONTENT	
		COBAS INTEGRA® 400 plus analyzer
		MODULAR PRE-ANALYTICS EVO
		cobas c 111 analyzer
		cobas c 303 analytical unit
		cobas c 311 analyzer
		cobas e 402 analytical unit
		cobas e 411 analyzer
		cobas c 501 module
		cobas c 502 module
		cobas c 503 analytical unit
10394246001	5000 (20 x 250)	cobas e 601 module
		cobas e 602 module
		cobas c 701 module
		cobas c 702 module
		cobas c 703 analytical unit
		cobas e 801 module
		cobas e 801 analytical unit
		cobas 8000 ISE 900 module
		cobas 8000 ISE 1800 module
		cobas ISE neo analytical unit
		cobas pro ISE analytical unit

### **English**

### Intended use

The Sample Cup is an IVD accessory intended to be used with the following systems:

- COBAS INTEGRA® 400 plus analyzer
- MODULAR PRE-ANALYTICS EVO
- cobas c 111 analyzer
- cobas c 303 analytical unit
- cobas c 311 analyzer
- cobas e 402 analytical unit
- cobas e 411 analyzer
- cobas c 501 module
- cobas c 502 module
- cobas c 503 analytical unit
- cobas e 601 module
- cobas e 602 module
- cobas c 701 module
- cobas c 702 module
- cobas c 703 analytical unit
  - cobas e 801 module
  - cobas e 801 analytical unit
  - cobas 8000 ISE 900 module
  - cobas 8000 ISE 1800 module
  - cobas ISE neo analytical unit
  - cobas pro ISE analytical unit

For professional use only.

### **Combination of devices**

The Sample Cup is meant to be used <u>only</u> in combination with the following devices:

- COBAS INTEGRA® 400 plus analyzer
- MODULAR PRE-ANALYTICS EVO
- cobas c 111 analyzer
- cobas c 303 analytical unit
- cobas c 311 analyzer
- cobas e 402 analytical unit
- cobas e 411 analyzer
- cobas c 501 module
- cobas c 502 module
- cobas c 503 analytical unit
- cobas e 601 module
- cobas e 602 module
- cobas c 701 module
- cobas c 702 module
- cobas c 703 analytical unit
- cobas e 801 module
- cobas e 801 analytical unit
- cobas 8000 ISE 900 module
- cobas 8000 ISE 1800 module
- cobas ISE neo analytical unit
- cobas pro ISE analytical unit

### Summary

A small container used for sample, calibrator and control material. Depending on the system, the Sample Cup will be loaded directly onto a rack or sample disk or in a primary sample container. The Sample Cup can be used on all systems listed.

Please check the respective Operators Manual for specific dead volume specifications.

# Sample Cup



### **Precautions and warnings**

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all biological samples. Do not re-use the product.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

#### Storage and Stability

Do not expose the Sample Cup to UV-light. Store the Sample Cup in the box. Store at 2 - 32  $^{\circ}\text{C}.$ 

For further information, please refer to the appropriate Operators Manual for the analyzer concerned, and the Method Sheets of all necessary components.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

#### Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

SYSTEM Analyzers/Instruments on which reagents can be used

CONTENT Content of kit

GTIN Global Trade Item Number

Rx only For USA: Caution: Federal law restricts this device to

sale by or on the order of a physician.

## FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS and COBAS ISE NEO are trademarks of Roche.

All other product names and trademarks are the property of their respective owners Additions, deletions or changes are indicated by a change bar in the margin.

© 2024, Roche Diagnostics



For USA: Rx only



Roche Diagnostics GmbH Sandhofer Strasse 116 68305 Mannheim, Germany





Distributed in USA by: Roche Diagnostics, 9115 Hague Road Indianapolis, Indiana USA US Customer Technical Support 1-800-428-2336