

Concise form of the test report No.: 1/21.12.18./PhM-10

Information about laboratory

Name of testing laboratory	SIA Baltic Research Center
Date of issue of test report	21.12.2018.
Test report number	1/21.12.18./PhM-10
Laboratory contact name	Ingmārs Felcis
Signature, Head of Laboratory Ingmārs Felcis	Welm

⁻ Note: By signing this form, the signatory is attesting that the information on the form is correct and the same as on the original, complete test report(s).

Product information

Manufacturer	SIA "VIZULO"		
Brand name	MICRO MARTIN		
Model number	MRU 046 740 V22 F024 CSN DG1		
Type of luminaire	Street luminaire		
Luminaire length	51.7	cm	
Luminaire width	17.0	cm	


Electrical measurements

Goniophotometer output

Wattage	46	W
Current	0.208	Α
Voltage (AC)	230	٧
Power factor	0.96	

Photometric characteristics

Total lumen output	5589	Lm
Luminaire efficiency	121.5	Lm/W
Correlated colour temperature CCT	4007	K
Colour rendering index CRI	72	



BALTIC RESEARCH CENTER SIA

Test Report No.: 1/21.12.18./PhM-10

Ph: +371 26648433 www.brc-lab.com, E-mail: office@brc-lab.com R3-01.02.2016.

Page 1 of 1

The signatory also attests that all of the results on this form measured entirely in accordance with IES LM-79-08 and EN 13032:1 standards.