2,75 mm NONGUARDED SLIT KNIFE TECHNICAL SPECIFICATION

- 1. It is disposable, cutting edge is 2.75 mm. It has a 45 degree angle and a full handle.
- 2. Due to the quality of the incision, it is double-bevel.
- 3. It is suitable for making a lateral incision in the cornea.
- 4. The blade does not shine under the microscope.
- 5. The brand and model is written on the handle of the knife. (label is tyvek, not sticking label)
- 6. It has a transparent protection guard made of polycarbonate that protects the tip of the blade (to see the model of the blade). The handle of the knife is made of sterilization-resistant polycarbonate.
- 7. The blade is in a hard blister container to withstand impacts that may occur until the end of use.)
- 8. It has a special coating system that protects the surface of the blade. This feature preserves the sharpness of the blade and prevents deformation on its surface.
- 9. The blister container is sealed with Medical Grade Tyvek to complete sterilization..
- 10- Product information should not be sticking labels, must be printed by the original manufacturer.
- 11-The blades are produced in different colors to avoid mixing with different models (Slit, Stab, Mvr, Crescent) for ease of use during surgery.
- 12- Products are FDA certified.