

LUGOL'S SOLUTION

IVD In vitro diagnostic medical device

 ϵ

Aqueous solution of iodine and potassium iodide INSTRUCTIONS FOR USE

REF Product code: LUG-0T-100 (100 mL) LUG-0T-250 (250 mL) LUG-0T-500 (500 mL)

Introduction

Lugol's solution, named after French physician J. G. A. Lugol, is widely used in various applications. It is most commonly known as an antiseptic, disinfectant and starch indicator. Lugol's solution affects the carbohydrate molecules in the following way: iodine binds to complex carbohydrates (starch in plant organisms, glycogen in animals), staining them in the process. Lugol's solution is an aqueous solution of iodine and potassium iodide. It is used in parasitology for staining parasites in stool samples, in zoology for staining protozoic organisms in order to enhance the contrast of internal structures (nuclei and glycogen vacuoles) and for preserving the protozoa, and in microbiological testing Lugol's solution is one of the components of Gram staining.

Product description

• LUGOL'S SOLUTION - Aqueous solution of iodine and potassium iodide.

Usage of Lugol's solution:

- Parasitology (staining parasites in stool samples)
- Microbiology (Gram staining)
- Zoology (staining and preserving protozoic organisms)
- · Cytology (staining cellular nuclei)

Note:

Staining procedures are not standardized and they depend on standard operating procedures of individual laboratories and the experience of the personnel conducting the staining procedure. Intensity of staining depends on the period of immersion in the dye. Depending on personal requests and standard laboratory operating procedures, sample processing and staining can be carried out according to other protocols.

Preparing the sample and diagnostics

Use only appropriate instruments for collecting and preparing the samples. Process the samples with modern technology and mark them clearly. Follow the manufacturer's instructions for use. In order to avoid mistakes, the staining procedure and diagnostics should only be conducted by authorized and qualified personnel. Use only microscope according to standards of the medical diagnostic laboratory. In order to avoid an erroneous result, a positive and negative check is advised before application.

Safety at work and environmental protection

Handle the product in accordance with safety at work and environmental protection guidelines. Used solutions and out of date solutions should be taken care of as a special waste in accordance with national guidelines. Chemicals used in this procedure could pose danger to human health. Tested tissue specimens are potentially infectious. Necessary safety measures for protecting human health should be taken in accordance with good laboratory practice. Act in accordance with signs and warnings notices printed on the product's label, as well as in BioGnost's material safety data sheet which is available on demand.

Storing, stability and expiry date

Lugol's solution should be stored at temperature between 15 °C and 25 °C. Do not keep in cold places, do not freeze and avoid exposing to direct sunlight. Date of manufacture and expiry date are printed on the product's label.

References

- 1. Higdon, J. (2003): Micronutrient Information Center: Iodine, Linux Pauling Institute/Oregon State University.
- 2. Carson, F. L., Hladik, C. (2009): Histotechnology: A Self-Instructional Text, 3rd ed., Chicago: ASCP Press
- 3. Sankaranarayanan, R. et al. (2003): Test characteristics of visual inspection with 4% acetic acid (via) and Lugol's iodine (vili) in cervical cancer screening in Kerala, India, *Int. J. Cancer*, 106, pp 404-408.
- 4. Sargent, D. L. (1936): An improvement in staining technic for Protozoa, Biotechnic and Histochemistry, 11, pp 49-52.

LUG-0T-X, IFU V11-EN6, KB/IŠP

