



## VERIFICATION OF COMPLIANCE

This is to certify that the product listed below was (were) tested in the BTL EMC Laboratory to comply with the required criteria levels of the follow-mentioned Generic Standards or Product Family Standard(s) and/or Basic Standard(s) based-on the essential conformity requirements of EMC Directive 2014/30/EU.

**Equipment** Medical AIO Panel PC  
**Model Name** TME-5040-22 LVDP XXXXX (X=0~9,A~Z)  
**Brand Name** TEGUAR  
**Applicant** Teguar Corporation  
**Address** 4235 South Stream Boulevard. Suite L-130, Charlotte, NC 28217

**Standard(s)** EN 55032:2015+A11:2020 Class B  
BS EN 55032:2015+A11:2020 Class B  
AS/NZS CISPR 32:2015 Class B  
CISPR 32:2015+COR1:2016 Class B  
EN IEC 61000-3-2:2019 Class D  
BS EN IEC 61000-3-2:2019 Class D  
EN 61000-3-3:2013+A1:2019  
BS EN 61000-3-3:2013+A1:2019  
EN 55024:2010+A1:2015  
BS EN 55024:2010+A1:2015

**Report(s)** BTL-EMC-1-2104T063

The test data, data evaluation, and equipment configuration contained in our test report(s) above was (were) obtained utilizing the test procedures, test instruments, test sites that has been accredited by the Authority of TAF according to the ISO/IEC 17025 quality assessment standard and technical standard(s). The test data contained in the referenced test report relate only to the EUT sample and item(s) tested.

Pike Lee  
Authorized Signatory



### **BTL INC.**

No.18, Ln. 171, Sec. 2, Jiuzong Rd.,  
Neihu Dist., Taipei City 114, Taiwan  
TEL:+886-2-2657-3299  
FAX:+886-2-2657-3331

