Technical Specification No Fastened oxygen instalation

- Manufacturer: OztereyagoglubMedikal Gaz
- Country of Origin : Turkey
- Model: TY 0123
- Certification ISO, CE,
- •

Technical Specification	Yes	No	Remarks
Configuration			
Fastened oxygen installation:	X X		
 Oxygen cylinders: minimum 2 pieces of 10 litters each, 	X		
Pressure reducers endowed with manometers for each cylinder.			
Minimum 2 standard fast connections DIN for respiratory assistance devices, attached on the left lateral wall.	X		
Portable oxygen			
 oxygen cylinder 5 liters , with transport hose and bag 	Х		
Pressure regulator with Flowmeter with a maximum capacity of at least15 l/min., adjusting valve, humidifier, tubing and facial mask, a port for rapid connection to the patient ventilator must to be built in	х		
 regulator Pressure limiting system for preventing overpressure. 	Х		
 oxygen cylinder 2 liters , with transport hose and bag 			
Pressure regulator with Flowmeter with a			
maximum capacity of at least15 l/min., adjusting valve, humidifier, tubing and			
facial mask, a port for rapid connection to	Х		
the patient ventilator must to be built in	Х		
regulator			
Pressure limiting system for preventing			
overpressure.			
Certification:			
> ISO 9001	X		certificate attached
➢ ISO 13485	Х		certificate attached

>	CE	Х		certificate attached
\triangleright	CE Declaration of Conformity 93/42/EEC	Х		certificate attached
	Technical Specification	Yes	No	Remarks
Other	services included in offer			
>	transport to the final customer	X		bidder declaration attached
\triangleright	instalation and put in function	Х		bidder declaration attached
\succ	training for the technical staff for final customer	Х		bidder declaration attached
\succ	training for the medical staff for final customer	Х		bidder declaration attached
≻	warranty for 36 months	Х		bidder declaration attached
GUAR	ANTEE			
\triangleright	36 months guaranteed from signing the receiving	X		bidder declaration attached
	minutes			
SERV	ICE AND MAINTENANCE			
>	the bidder has technical facilities necessary for the	X		bidder declaration attached
	service both for ambulances as well as for medical			
	equipment according to the general guarantee			
	conditions and the user guide of the manufacturer			
\triangleright	Maximum time for technical intervention – 48	Х		bidder declaration attached
	hours.			
\succ	Maximum time for remedial measures, in total 72	Х		bidder declaration attached
	hours.			
\succ	The technical service and current repair will be	Х		bidder declaration attached
	performed without waiting in a queue.			
\triangleright	Temporary replacement of the equipment has to be	Х		bidder declaration attached
	ensured according to the periods mentioned above.			
\triangleright	During the guarantee period, at the reasonable			
	request of the user, repair, adjustments and			
	maintenance of medical equipment, according to			
	the specifications in the manufacturers' guides,			
	will be done free of charge.			
	The spare parts and workmanship are for free,	77		
	except from the consumables for vehicles	Х		bidder declaration attached
	established by the manufacturer			

Technical Specification	Yes	No	Remarks
AVAILABILITY OF SPARE PARTS			
 the bider will ensure the availability of spare parts, accessories and consumables for all offered positions on the market of the Republic of Moldova for at least 10 years from putting into operation, free of charge or against payment, as follows: spare parts free of charge, including the workmanship for the 24 months of guarantee 	X X		 bidder declaration attached bidder declaration attached
 period o for the rest of the period – against payment for at least 10 years period is fulfilled. 	Х		bidder declaration attached
The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary	Х		 bidder declaration attached
GUIDES			
 Technical Guide and User Guide. All guides will be available in Romanian /Russian. 	X X		bidder declaration attachedbidder declaration attached