

Declaration of Conformity V9.0

## Declaration of Conformity



**Manufacturer:** Shenzhen Mindray Bio-Medical Electronics Co., Ltd.  
Mindray Building, Keji 12th Road South, Hi-tech Industrial Park,  
Nanshan, Shenzhen, 518057, P. R. China

**EC-Representative:** Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80  
20537 Hamburg, Germany

**Product:** Defibrillator/Monitor (Including accessories and Vehicle Mount kit)

**Model:** BeneHeart D2/BeneHeart D3

**Classification:** IIb (According to Rule 9 of MDD Annex IX)

**Conformity Assessment Route:** MDD Annex II excluding (4)

**We herewith declare that the above mentioned products meet the provisions of the Council Directive 93/42/EEC concerning Medical Device, as amended by 2007/47/EC.**

**All supporting documentations are retained under the premises of the manufacturer.**

**Standards Applied:**

List of (harmonized) standards for which documented evidence for compliance can be provided as attachment.

**Notified Body:** TÜV SÜD Product Service GmbH  
Ridlerstraße 65  
80339 München, Germany.

**Notified Body No. :** 0123

**Start of CE-Marking:** 2010-12-10

**Place, Date of Issue:**

Shenzhen, 2017.9.30

**Signature:**

**Name of Authorized Signatory:** Mr. Wang Xinbing

**Position Held in Company:** Manager, Technical Regulation

## Applied Standards List

**Product:** Defibrillator/Monitor

**Model:** BeneHeart D2/BeneHeart D3

### Standards Applied:

EN ISO 14971: 2012	Medical devices - Application of risk management to medical devices
EN 1041: 2008	Information supplied by the manufacturer with medical devices
ISO 15223-1:2012	Medical devices-Symbols to be used with medical device labels, labeling and information to be supplied
EN ISO 10993-1: 2009	Biological evaluation of medical devices - Part 1: Evaluation and testing
EN 60601-1:2006/A1:2013	Medical electrical equipment--Part 1:General requirements for basic safety and essential performance
EN 60601-1-2: 2007/AC:2010	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
IEC 60601-1-6: 2013	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability
IEC 60601-1-8: 2012	Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems
IEC 60601-2-4: 2010	Medical electrical equipment - Part 2-4: Particular requirements for the safety of cardiac defibrillators
IEC 60601-2-27: 2011	Medical electrical equipment - Part 2-27: Particular requirements for the safety, including essential performance, of electrocardiographic monitoring equipment

<b>ISO 80601-2-30: 2013</b>	Medical electrical equipment - Part 2-30: Particular requirements for the safety, including essential performance, of automatic cycling non-invasive blood pressure monitoring equipment
<b>ISO 81060-2: 2013</b>	Non-invasive sphygmomanometers - part 2: clinical validation of automated measurement type
<b>IEC 60601-2-49: 2011</b>	Medical electrical equipment - Part 2-49: Particular requirements for the safety of multifunction patient monitoring equipment
<b>ISO 80601-2-61:2011</b>	Medical electrical equipment - part 2-61: particular requirements for basic safety and essential performance of pulse oximeter equipment
<b>EN 1789: 2007+A1:2010</b>	Medical Vehicles and Their Equipment - Road Ambulances
<b>EN 62366-1: 2015</b>	Medical devices - Application of usability engineering to medical devices
<b>EN 62304: 2006</b>	Medical device software - Software lifecycle processes
<b>ISO 80601-2-55:2011</b>	Medical electrical equipment - part 2-55: particular requirements for the basic safety and essential performance of respiratory gas monitors