

## Declaration of Conformity

**LEGAL MANUFACTURER:** Siemens Healthcare Diagnostics Inc.  
511 Benedict Avenue  
Tarrytown, New York 10591-5097  
USA

**PLACE OF MANUFACTURE:** Siemens Healthcare Diagnostics Products GmbH  
Emil-von-Behring-Str. 76  
35041 Marburg  
Germany

**EU AUTHORIZED REPRESENTATIVE** Siemens Healthcare Diagnostics Ltd.  
Sir William Siemens SQ.  
Frimley, Camberley, UK GU16 8QD

**PRODUCT:** Xprecia Systems™ PT Controls

**PRODUCT CATEGORY:** See Attachment 1

**CLASSIFICATION:** Self-Declaration

**CONFORMITY ASSESSMENT ROUTE:** Annex III Applied

**STANDARDS APPLIED:**

EN ISO 13485:2003 – Medical devices – Quality Management Systems – Requirements for Regulatory Purposes

EN ISO 14971:2012 – Medical devices – Application of risk management to medical devices

EN ISO 18113-1:2011 – In vitro diagnostic medical devices – Information supplied by the manufacturer (labelling) – Part 1: Terms, definitions and general requirements

EN ISO 18113-2:2011 – In vitro diagnostic medical devices – Information supplied by the manufacturer (labelling) – Part 2: In vitro diagnostic reagents for professional use

EN ISO 18113-3:2011 – In vitro diagnostic medical devices – Information supplied by the manufacturer (labelling) – Part 3: In vitro diagnostic instruments for professional use

BS EN 980:2008 – Symbols for use in the labelling of medical devices

EN 13612:2002 – Performance evaluation of in vitro diagnostic medical devices

 2017-01-17

## Declaration of Conformity

### STANDARDS APPLIED:

**EN 13640:2002** – Stability testing of in vitro diagnostic reagents

**ISO 15198:2004** – Clinical laboratory medicine — In vitro diagnostic medical devices — Validation of user quality control procedures by the manufacturer

**ISO 17511:2003** – In Vitro Diagnostic Medical Devices- Measurement of Quantities in Biological Samples- Metrological Traceability of Values assigned to Calibrators and Control Materials.

**ISO 17593:2007** – Clinical laboratory testing and in vitro medical devices – Requirements for in vitro monitoring systems for self-testing of oral anticoagulant therapy (Chapter 6 to be used for verification testing)

**WHO Technical Report Series 889 - Annex 3** – Guidelines for Thromboplastins and Plasmas used to control Oral anticoagulant therapy

**ISO 5725-2:1994** – Accuracy (trueness and precision) of measurement methods and results -- Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method

**Reach Article 33 & 67** – The system shall not contain Substances of Very High Concern (SVHC) in excess of limits set by Reach Articles 33 & 67

**IEC 62321, Ed.1:2008** – Procedures for the determination of levels of six regulated substances (Lead, Mercury, Cadmium, Hexavalent Chromium, Polybrominated Biphenyls, Polybrominated Diphenyl Ethers) in electrotechnical products

**(EC) 1907/2006 – Regulation (EC) 1907/2006** of the European Parliament and the Council concerning the Registration, Evaluation, Authorization and Restriction of Chemicals (REACH)

**EN 61010-2-101:2002** – Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment

**CAN/CSA C22.2 No. 61010-1:2009** – Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

 2017-01-17

Priyank Patel

Date

Regulatory Technical Specialist-POC

DOC # Xprecia System/s Controls

Rev 5.0

## Declaration of Conformity

### STANDARDS APPLIED:

**CAN/CSA C22.2 No. 61010-2-101:2009** – Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment

**UL 61010-1-2008** – Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

**EN 61326-2-6:2006** – Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. In vitro diagnostic (IVD) medical equipment

**IEC/EN 61010-1:2001** – 2nd Edition - Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements

**IEC/EN 61010-1:2010 – 3rd Edition** - Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements

**IEC 61010-2-101 Ed. 1** – Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment

**EN 62304:2006** – Medical device software – Software life-cycle processes

**EN 62366:2008** – Medical devices – Application of usability engineering to medical devices

**ISTA Procedure 3A** - Packaged-Products for Parcel Delivery System Shipments 70kg (150 lb) or Less (standard, small, flat or elongated)

**2002/96/EC** – Council Directive relating to the waste of electrical and electronic equipment (WEEE)

**AS 60417.1:2004** – Graphical symbols for use on equipment

**ASTM D3363-05** – Standard Test Method for Film Hardness by Pencil Test

**IEC 60068-2-64:1993** – Environmental testing – Part 2: Test methods – Test Fh: Vibration, broadband random (digital control) and guidance

**IEC 60529:2001** – Degrees of protection by enclosures (IP code)

 2017-01-17

## Declaration of Conformity

We herewith declare that the below-mentioned product(s) meet the provisions of the Council Directive 98/79/EC for in vitro diagnostic medical devices and therefore has fulfilled all requirements for applying the CE mark to the medical device(s). The Manufacturer retains all supporting documentation.

### Attachment 1

SMN	REF	Product Code	Description
10873436	10873436	10873436	Xprecia™ Systems PT Controls
10873633	10873633	10873633	Xprecia™ System PT Controls
End of List			

 2017-01-17