

Compliance

We, Edan Instruments, Inc. With address: No.15 Jinhui Rd., Jinsha Community, Kengzi Subdistrict, Pingshan District, Shenzhen, 518122, P.R.China, do confirm that offered model of monitor IM50 do comply with the following technical specification:

ECG displayed parameters YES

The pulse YES

SpO2 YES

Photoplethysogram YES

Non-invasive blood pressure YES

Temperature YES

ECG mode "Collecting the ECG signal through 3 "cable I, II, III YES

Heart rate 15 - 300 bpm YES

Detection of arrhythmia YES

Detecting pacemaker YES

Breathing monitoring 0 - 150 rpm YES

Shock protection defibrillation YES

SpO2 Module Range 1 - 100% YES

Accuracy at 70 - 100% $\leq$ 2% YES

Pulse rate monitoring 20 - 300 bpm YES

Plethysmogram monitoring YES

NIBP module Range 10 - 290 mmHg YES

Manual, automatic measurement mode, Start button YES

Automatic mode 5, 10, 15, 30, 60, 120 min. YES

NIBP manual on / off button YES

Oscillometric measurement method YES

Overpressure protection at adult patient regimen "300 mmHg YES

Temperature Mode Range 0 - 50.0 ° C YES

Resolution 0.1 ° C YES

Number of sensors 1 unit. YES

Pre-programmed work protocols YES

User-set work protocols YES

Internal memory YES

Trending events YES

Data archiving YES

The visual alarm is audible YES

physiological alarms YES

Technical alarms YES


Mute / unmute button YES

Name: Marcus.Jiang

Position: National Manager of CIS & Russia

深圳市理邦精密仪器股份有限公司
EDAN INSTRUMENTS, INC.

Sign: _____



Date: 24th, May, 2021