

tongue depressor wooden



ref	SLS-001*	version	sterile	size	150 x 17 mm	transport packaging	20 x 100 pcs



intermediate packaging
100 pcs / box

ref	SLN-001**	version	non-sterile	size	150 x 17 mm	transport packaging	50 x 100 pcs

features

- made of birch wood

indication

- diagnostic examination of the oral cavity



electronic thermometer



vitahealth

ref DMT-437

features

- waterproof
- range 32.0°C–42.9°C (90.0°F–109.9°F)
- accuracy ±0.1°C (±0.2°F)
- has a soft, flexible measuring tip
- battery life: approximately 200 hours

indication

- for measuring body temperature in the mouth, armpit and rectal area



packaging unit
1 pc. / plastic cover and paper box



intermediate packaging
10 pcs



bulk packaging
30 x 10 pcs

safeLANCE pressure-activated safety lancet

safeLANCE



ref	■ NA-18	size	23 G / 1,8 mm
	■ NA-24	size	21 G / 2,4 mm

features

- capillary blood sampling from fingertip or ear flap for testing, for example, glucose level



intermediate packaging
100 pcs



bulk packaging
20 x 100 pcs

easyCARE nitrile

type of gloves	examination, nitrile , non-sterile
outer surface	microtexture + additional texture on fingertips
inner surface	chlorinated
cuff	straight with rolled edge
AQL	1.0
classification	class I medical device personal protective equipment category III
length	245 mm
unit packaging/ bulk packaging	100 pcs / 10 x 100 pcs 200 pcs / 10 x 200 pcs.
reference number	RNBXX10001 (XX - size I 100 pcs.) RNBXX20001 (XX - size I 200 pcs.)



sizes



protection against
blood-borne pathogens



protection against
cytostatics



acceptable quality level
EN 455-1
EN ISO 374-2



easyCARE nitrile



easyCARE nitrile

easyCARE nitrile long

type of gloves	examination, nitrile , non-sterile, with extended cuff
outer surface	microtexture + additional texture on fingertips
inner surface	chlorinated
cuff	straight with rolled edge
AQL	1.0
classification	class I medical device personal protective equipment category III
length	300 mm
unit packaging/ bulk packaging	100 pcs / 10 x 100 pcs
reference number	RNBXX10001 (XX - size I 100 pcs.)



sizes



protection against
blood-borne pathogens



protection against
cytostatics



acceptable quality level
EN 455-1
EN ISO 374-2



easyCARE nitrile long



easyCARE nitrile long

EC Certificate

Production Quality Assurance Directive 93/42/EEC on Medical Devices, Annex V

Registration No.: DD 1023663-1

Manufacturer: ZARYS International Group
Spółka z ograniczoną odpowiedzialnością,
spółka komandytowa
ul. Pod Borem 18
41-808 Zabrze
Poland

Products:

- Sterile and non-sterile cutting gauze
- Non-sterile dressing gauze
- Sterile and non-sterile gauze swabs (with or without X-ray thread)
- Sterile and non-sterile gauze lap sponges (with X-ray thread/ with X-ray chip)
- Sterile and non-sterile gauze balls (with or without X-ray thread)
- Sterile and non-sterile gauze rolls (with or without X-ray thread)
- Sterile and non-sterile non-woven swabs (with or without X-ray thread)
- Sterile paraffin gauze dressings
- Sterile three-way stopcocks
- Sterile transfusion sets for single use
- Sterile infusion sets for single use
- Sterile extension tubes for infusion pump
- Sterile endotracheal tubes
- Sterile tracheostomy tubes

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

Report No.: 84951712-170

Effective date: 2021-05-14

Expiry date: 2024-05-26

Issue date: 2021-05-14



Daniel Świątko
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

EC Certificate

Production Quality Assurance Directive 93/42/EEC on Medical Devices, Annex V

Registration No.: DD 1023663-1

Manufacturer: ZARYS International Group
Spółka z ograniczoną odpowiedzialnością,
spółka komandytowa
ul. Pod Borem 18
41-808 Zabrze
Poland


- Sterile breathing circuits
- Sterile catheter mounts
- Non-sterile anaesthetic masks
- Sterile laryngeal masks
- Sterile oxygen masks
- Sterile Venturi masks
- Sterile non-rebreath masks
- Sterile nebulizer masks
- Sterile nasal oxygen cannulas
- Sterile nebulizers
- Sterile oxygen tubing
- Sterile suction catheters
- Sterile abdominal drains
- Sterile feeding tubes
- Sterile stomach and duodenal tubes
- Sterile urology catheters
- Sterile surgical suction sets
- Sterile surgical suction cannulas
- Sterile syringes for single use
- Sterile insulin syringes
- Sterile tuberculin syringes
- Sterile hypodermic needles

Report No.: 84951712-170

Effective date: 2021-05-14

Expiry date: 2024-05-26

Issue date: 2021-05-14



Daniel Swiatko
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

EC Certificate

Production Quality Assurance Directive 93/42/EEC on Medical Devices, Annex V

Registration No.: DD 1023663-1

Manufacturer: ZARYS International Group
Spółka z ograniczoną odpowiedzialnością,
spółka komandytowa
ul. Pod Borem 18
41-808 Zabrze
Poland

- Sterile insulin pen needles
- Sterile blood lancets
- Sterile IV cannulas
- Sterile needle free valves
- Sterile surgical gloves
- Sterile procedure kits

For the following medical devices the scope covers only the aspects of the manufacture concerned with securing and maintaining sterile conditions:

- Elastic bandages
- Adhesive cannula fixation dressings
- Adhesive wound dressings
- Eye pads
- Incise films
- Transparent film dressings
- Foam dressings
- Alginate dressings
- Absorbent wound dressings
- Surgical gowns
- Surgical drapes
- Sets of surgical drapes

Report No.: 84951712-170

Effective date: 2021-05-14

Expiry date: 2024-05-26

Issue date: 2021-05-14



Daniel Swiatko
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

EC Certificate

Production Quality Assurance Directive 93/42/EEC on Medical Devices, Annex V

Registration No.: DD 1023663-1

Manufacturer: ZARYS International Group
Spółka z ograniczoną odpowiedzialnością,
spółka komandytowa
ul. Pod Borem 18
41-808 Zabrze
Poland

- Fluid collection pouches
- Nelaton catheters
- Vaginal speculums
- Cervical brushes
- Urine bags
- Enema bags
- Tongue depressors
- Oropharyngeal airways
- Intubation stylets
- Endotracheal tube holders
- Suction tubes
- Withdrawal cannulas
- Cannula stoppers
- Umbilical cord clamps

Replaces EC Certificate, Registration No.: DD 60139535 0001

Report No.: 84951712-170

Effective date: 2021-05-14

Expiry date: 2024-05-26

Issue date: 2021-05-14



Daniel Świątko
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

EC Certificate

Production Quality Assurance Directive 93/42/EEC on Medical Devices, Annex V

Registration No.: DD 1023663-1

Manufacturer: ZARYS International Group
Spółka z ograniczoną odpowiedzialnością,
spółka komandytowa
ul. Pod Borem 18
41-808 Zabrze
Poland

The scope of certification includes the following manufacturing sites:

No.	Location	Product groups manufactured
/01	ZARYS International Group Spółka z o.o. sp.k. ul. Pod Borem 18 41-808 Zabrze Poland	Activity: Final inspection and release.
/02	ZARYS International Group Spółka z o.o. sp.k. ul. Guido Henckela Donnersmarcka 1 41-808 Zabrze Poland	Activity: Final inspection and release.

Report No.: 84951712-170

Effective date: 2021-05-14

Expiry date: 2024-05-26

Issue date: 2021-05-14



Daniel Świątko
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

TÜV Rheinland LGA Products GmbH • 51105 Köln

ZARYS International Group spółka z ograniczoną odpowiedzialnością spółka komandytowa
ul. Pod Borem 18,
41-808 Zabrze,
Poland

Contact

Tel. +49 911 655-5225
Mail: medical-products@de.tuv.com

Date October 22, 2025

Notified Body Confirmation Letter

Reference. : ZARYS_PLA0_HZ_2024-05-10 replaced by
ZARYS_PLA0_HZ_2024-05-24 replaced by
ZARYS_PLA0_HZ_2025-02-07 replaced by
ZARYS_PLA0_HZ_2025-02-12 replaced by
ZARYS_PLA0_HZ_2025-10-15
/ 84965323

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that **TÜV Rheinland LGA Products GmbH**, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number **0197** on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

ZARYS International Group spółka z ograniczoną odpowiedzialnością spółka komandytowa
ul. Pod Borem 18,
41-808 Zabrze,
Poland
SRN Number: PL-MF-000000410

The devices covered by the formal application and the written agreement mentioned above are identified in the tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after May 26, 2021 but before March 20, 2023 without having been withdrawn, this letter also confirms that the manufacturer either signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity

TÜV Rheinland
LGA Products GmbH

Am Grauen Stein
51105 Köln
Germany

Headquarter

Tillystraße 2
90431 Nuremberg

Phone. +49 911 655 5225
Fax +49 911 655 5226
service@de.tuv.com
www.tuv.com/safety

Board of Management

Dipl.-Ing.
Thomas Weigand, Spokesman

Dipl.-Kfm.
Dr. Jörg Schlösser

Nuremberg HRB 26013
VAT No.: DE 811835490

Chairman of the
Supervisory Board

Dr.-Ing. Michael Fübi

assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by March 20, 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- May 26, 2026 for Class III custom-made implantable devices
- December 31, 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- December 31, 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- December 31, 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body

 Elektronicznie podpisany
przez Malgorzata Blazniak
Data: 2025.10.22 13:48:00
+02'00'

AUDIT_CERT_REVIEW
Certification body

Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
GAZA lux S Cutting gauze, sterile Basic UDI-DI: 59079968M02010101-ERR	Class IIa	GAZA lux S Cutting gauze, sterile	DD 1023663-1 NB 0197
GAZA lux S Cutting gauze, sterile Basic UDI-DI: 59079968M02010101-SSM	Class IIa	GAZA lux S Cutting gauze, sterile	DD 1023663-1 NB 0197
GAZA lux Cutting gauze, non-sterile Basic UDI-DI: 59079968M02010101-NSB	Class IIa	GAZA lux Cutting gauze, non-sterile	DD 1023663-1 NB 0197
GAZA lux Dressing gauze, non-sterile	Class IIa	GAZA lux Dressing gauze, non-sterile	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968M020107DG			
KOMPRI lux S Gauze swabs without X-ray thread, sterile Basic UDI-DI: 59079968M02010201-ES4	Class IIa	KOMPRI lux S Gauze swabs without X-ray thread, sterile	DD 1023663-1 NB 0197
KOMPRI lux S Gauze swabs without X-ray thread, sterile Basic UDI-DI: 59079968M02010201-SSY	Class IIa	KOMPRI lux S Gauze swabs without X-ray thread, sterile	DD 1023663-1 NB 0197
KOMPRI lux S Gauze swabs with X-ray thread, sterile Basic UDI-DI: 59079968M02010202-ES9	Class IIa	KOMPRI lux S Gauze swabs with X-ray thread, sterile	DD 1023663-1 NB 0197
KOMPRI lux S Gauze swabs with X-ray thread, sterile Basic UDI-DI: 59079968M02010202-ST5	Class IIa	KOMPRI lux S Gauze swabs with X-ray thread, sterile	DD 1023663-1 NB 0197
KOMPRI lux Gauze swabs without X-ray thread, non-sterile Basic UDI-DI: 59079968M02010201-NSN	Class IIa	KOMPRI lux Gauze swabs without X-ray thread, non-sterile	DD 1023663-1 NB 0197
KOMPRI lux Gauze swabs with X-ray thread, non-sterile Basic UDI-DI: 59079968M02010202-NST	Class IIa	KOMPRI lux Gauze swabs with X-ray thread, non-sterile	DD 1023663-1 NB 0197
SERVI lux S Gauze lap sponges with X-ray thread, sterile Basic UDI-DI: 59079968M02010302-ESL	Class IIa	SERVI lux S Gauze lap sponges with X-ray thread, sterile	DD 1023663-1 NB 0197
SERVI lux S Gauze lap sponges with X-ray thread, sterile Basic UDI-DI: 59079968M02010302-STG	Class IIa	SERVI lux S Gauze lap sponges with X-ray thread, sterile	DD 1023663-1 NB 0197
SERVI lux S	Class IIa	SERVI lux S Gauze lap sponges with X-ray	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>Gauze lap sponges with X-ray chip, pre-washed, sterile</p> <p>Basic UDI-DI: 59079968M02010302-PE86</p>		chip, pre-washed, sterile	
<p>SERVI lux S Gauze lap sponges with X-ray chip, pre-washed, sterile</p> <p>Basic UDI-DI: 59079968M02010302-PS92</p>	Class IIa	SERVI lux S Gauze lap sponges with X-ray chip, pre-washed, sterile	DD 1023663-1 NB 0197
<p>SERVI lux Gauze lap sponges with X-ray thread, non-sterile</p> <p>Basic UDI-DI: 59079968M02010302-NT6</p>	Class IIa	SERVI lux Gauze lap sponges with X-ray thread, non-sterile	DD 1023663-1 NB 0197
<p>SERVI lux Gauze lap sponges with X-ray chip, pre-washed, non-sterile</p> <p>Basic UDI-DI: 59079968M02010302-PN8Q</p>	Class IIa	SERVI lux Gauze lap sponges with X-ray chip, pre-washed, non-sterile	DD 1023663-1 NB 0197
<p>TUPFER lux S Gauze balls without X-ray thread, sterile</p> <p>Basic UDI-DI: 59079968M02010501-ET5</p>	Class IIa	TUPFER lux S Gauze balls without X-ray thread, sterile	DD 1023663-1 NB 0197
<p>TUPFER lux S Gauze balls without X-ray thread, sterile</p> <p>Basic UDI-DI: 59079968M02010501-STZ</p>	Class IIa	TUPFER lux S Gauze balls without X-ray thread, sterile	DD 1023663-1 NB 0197
<p>TUPFER lux S Gauze balls with X-ray thread, sterile</p> <p>Basic UDI-DI: 59079968M02010502-ETA</p>	Class IIa	TUPFER lux S Gauze balls with X-ray thread, sterile	DD 1023663-1 NB 0197
<p>TUPFER lux S Gauze balls with X-ray thread, sterile</p> <p>Basic UDI-DI: 59079968M02010502-SU6</p>	Class IIa	TUPFER lux S Gauze balls with X-ray thread, sterile	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
TUPFER lux Gauze balls without X-ray thread, non-sterile Basic UDI-DI: 59079968M02010501-NTP	Class IIa	TUPFER lux Gauze balls without X-ray thread, non-sterile	DD 1023663-1 NB 0197
TUPFER lux Gauze balls with X-ray thread, non-sterile Basic UDI-DI: 59079968M02010502-NTU	Class IIa	TUPFER lux Gauze balls with X-ray thread, non-sterile	DD 1023663-1 NB 0197
SETON lux S Gauze rolls without X-ray thread, sterile Basic UDI-DI: 59079968M02010701-SUP	Class IIa	SETON lux S Gauze rolls without X-ray thread, sterile	DD 1023663-1 NB 0197
SETON lux S Gauze rolls with X-ray thread, sterile Basic UDI-DI: 59079968M02010702-ETY	Class IIa	SETON lux S Gauze rolls with X-ray thread, sterile	DD 1023663-1 NB 0197
SETON lux S Gauze rolls with X-ray thread, sterile Basic UDI-DI: 59079968M02010702-SUU	Class IIa	SETON lux S Gauze rolls with X-ray thread, sterile	DD 1023663-1 NB 0197
SETON lux Gauze rolls without X-ray thread, non-sterile Basic UDI-DI: 59079968M02010701-NUD	Class IIa	SETON lux Gauze rolls without X-ray thread, non-sterile	DD 1023663-1 NB 0197
SETON lux Gauze rolls with X-ray thread, non-sterile Basic UDI-DI: 59079968M02010702-NUJ	Class IIa	SETON lux Gauze rolls with X-ray thread, non-sterile	DD 1023663-1 NB 0197
NONVI lux S Non-woven swab, sterile Basic UDI-DI: 59079968M02020101-ESA	Class IIa	NONVI lux S Non-woven swabs, sterile	DD 1023663-1 NB 0197
NONVI lux S Non-woven swab, sterile Basic UDI-DI: 59079968M02020101-ST6	Class IIa	NONVI lux S Non-woven swabs, sterile	DD 1023663-1 NB 0197
NONVI lux S	Class IIa	NONVI lux S	DD 1023663-1

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Non-woven swabs with X-ray thread, sterile Basic UDI-DI: 59079968M02020102-ESF		Non-woven swabs with X-ray thread, sterile	NB 0197
NONVI lux S Non-woven swabs with X-ray thread, sterile Basic UDI-DI: 59079968M02020102-STB	Class IIa	NONVI lux S Non-woven swabs with X-ray thread, sterile	DD 1023663-1 NB 0197
NONVI lux Non-woven swabs, non-sterile Basic UDI-DI: 59079968M02020101-NSU	Class IIa	NONVI lux Non-woven swabs, non-sterile	DD 1023663-1 NB 0197
NONVI lux Non-woven swabs with X-ray thread, non-sterile Basic UDI-DI: 59079968M02020102-NSZ	Class IIa	NONVI lux Non-woven swabs with X-ray thread, non-sterile	DD 1023663-1 NB 0197
paraffiNET Paraffin gauze dressing, sterile Basic UDI-DI: 59079968M020302DG	Class IIa	paraffiNET Paraffin gauze dressing, sterile	DD 1023663-1 NB 0197
SANVIflon I.V. cannula SANVIflon S I.V. cannula without injection port SANVIflon premium I.V. cannula SANVIflon premium S I.V. cannula without injection port SANVIflon safe Safety I.V. cannula Basic UDI-DI: 59079968C0101017F	Class IIa	SANVIflon I.V. cannula SANVIflon S I.V. cannula without injection port SANVIflon premium I.V. cannula SANVIflon premium S I.V. cannula without injection port SANVIflon safe Safety I.V. cannula	DD 1023663-1 NB 0197
OXYGEN TUBING Basic UDI-DI: 59079968R03010204LA	Class IIa	OXYGEN TUBING	DD 1023663-1 NB 0197
NEBULIZER with T F/M adaptor, mouthpiece, corrugated tube and tubing	Class IIa	NEBULIZER with T F/M adaptor, mouthpiece, corrugated tube and tubing	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968R030103-FMV5			
NEBULIZER mask with tubing Basic UDI-DI: 59079968R030103-MMN	Class IIa	NEBULIZER with mask and tubing	DD 1023663-1 NB 0197
OXYGEN MASK with tubing Basic UDI-DI: 59079968R03010201L4	Class IIa	OXYGEN MASK with tubing	DD 1023663-1 NB 0197
NON-REBREATHER MASK with tubing Basic UDI-DI: 59079968R03010206LE	Class IIa	NON-REBREATHER MASK with tubing	DD 1023663-1 NB 0197
VENTURI MASK with adjustable diluter and tubing Basic UDI-DI: 59079968R03010202-AXK	Class IIa	VENTURI MASK with adjustable diluter and tubing	DD 1023663-1 NB 0197
NASAL OXYGEN CANNULA for adults NASAL OXYGEN CANNULA for children NASAL OXYGEN CANNULA for infants Basic UDI-DI: 59079968R03010203L8	Class IIa	NASAL OXYGEN CANNULA for adults NASAL OXYGEN CANNULA for children NASAL OXYGEN CANNULA for infants	DD 1023663-1 NB 0197
SUCTION CATHETER SUCTION CATHETER with frozen surface, phthalate-free SUCTION CATHETER IDEAL SUCTION CATHETER with vacuum control SUCTION CATHETER with vacuum control IDEAL Basic UDI-DI: 59079968R0501QP	Class IIa	SUCTION CATHETER SUCTION CATHETER with frozen surface, phthalate-free SUCTION CATHETER IDEAL SUCTION CATHETER with vacuum control SUCTION CATHETER with vacuum control IDEAL	DD 1023663-1 NB 0197
Two-way Foley catheter with rubber valve (silicone-coated latex) Basic UDI-DI: 59079968U010201-LRXH	Class IIa	TWO-WAY FOLEY CATHETER with rubber valve	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Two-way Foley catheter with plastic valve (silicone-coated latex) Basic UDI-DI: 59079968U010201-LPXD	Class IIa	TWO-WAY FOLEY CATHETER with plastic valve	DD 1023663-1 NB 0197
Two-way Foley catheter with plastic valve (100% silicone, X-ray contrast) Basic UDI-DI: 59079968U010201-SP6	Class IIa	TWO-WAY FOLEY CATHETER with plastic valve	DD 1023663-1 NB 0197
Three-way Foley catheter with plastic valve (silicone-coated latex) Basic UDI-DI: 59079968U010201-3LUV	Class IIa	THREE-WAY FOLEY CATHETER with plastic valve	DD 1023663-1 NB 0197
Three-way Foley catheter with plastic valve (100% silicone, X-ray contrast) Basic UDI-DI: 59079968U010201-3SVB	Class IIa	THREE-WAY FOLEY CATHETER with plastic valve	DD 1023663-1 NB 0197
Two-way Foley catheter with plastic valve, Tiemann tip (silicone-coated latex) Basic UDI-DI: 59079968U0102R6	Class IIa	TWO-WAY FOLEY CATHETER with plastic valve, Tiemann tip	DD 1023663-1 NB 0197
TIEMANN CATHETER Basic UDI-DI: 59079968U010106HB	Class IIa	TIEMANN CATHETER	DD 1023663-1 NB 0197
PEZZER CATHETER Basic UDI-DI: 59079968U010107HD	Class IIa	PEZZER CATHETER	DD 1023663-1 NB 0197
FEEDING TUBE Basic UDI-DI: 59079968G02020101BU	Class IIa	FEEDING TUBE	DD 1023663-1 NB 0197
STOMACH TUBE DUODENAL TUBE Basic UDI-DI: 59079968G020201A3	Class IIa	STOMACH TUBE DUODENAL TUBE	DD 1023663-1 NB 0197
SUCTION CANNULA with suction control SUCTION CANNULA without suction control	Class IIa	SUCTION CANNULA with suction control	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968A06010184		SUCTION CANNULA without suction control	
SUCTION CANNULA with suction control, ball tip SUCTION CANNULA without suction control, ball tip Basic UDI-DI: 59079968A060101-BA2	Class IIa	SUCTION CANNULA with suction control, ball tip SUCTION CANNULA without suction control, ball tip	DD 1023663-1 NB 0197
SURGICAL SUCTION SET with suction control, Yankauer tip SURGICAL SUCTION SET without suction control, Yankauer tip Basic UDI-DI: 59079968A060101039F	Class IIa	SURGICAL SUCTION SET with suction control, Yankauer tip SURGICAL SUCTION SET without suction control, Yankauer tip	DD 1023663-1 NB 0197
SURGICAL SUCTION SET with suction control, funnel-funnel cut-to-fit, Yankauer tip SURGICAL SUCTION SET without suction control, funnel-funnel cut-to-fit, Yankauer tip Basic UDI-DI: 59079968A06010103-FFUC	Class IIa	SURGICAL SUCTION SET with suction control, funnel-funnel cut-to-fit, Yankauer tip SURGICAL SUCTION SET without suction control, funnel-funnel cut-to-fit, Yankauer tip	DD 1023663-1 NB 0197
SURGICAL SUCTION SET with suction control, funnel-funnel cut-to-fit, ball tip SURGICAL SUCTION SET without suction control, funnel-funnel cut-to-fit, ball tip Basic UDI-DI: 59079968A06010103-FFB6J	Class IIa	SURGICAL SUCTION SET with suction control, funnel-funnel cut-to-fit, ball tip SURGICAL SUCTION SET without suction control, funnel-funnel cut-to-fit, ball tip	DD 1023663-1 NB 0197
SUCTION TUBE funnel-funnel Basic UDI-DI: 59079968A060304-FFG4	Class IIa	SUCTION TUBE funnel-funnel	DD 1023663-1 NB 0197
SUCTION TUBE funnel-funnel cut-to-fit	Class IIa	SUCTION TUBE funnel-funnel cut-to-fit	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968A060304-FCFW			
SUCTION TUBE funnel-Kapkon	Class IIa	SUCTION TUBE funnel-Kapkon	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A060304-FKGE			
easyWAY Three-way stopcock	Class IIa	easyWAY Three-way stopcock	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A0703KA			
easyWAY L Three-way stopcock with extension	Class IIa	easyWAY L Three-way stopcock with extension	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A0703-LA4			
easyFLOW LINE Extension tube for infusion pump, phthalate-free	Class IIa	easyFLOW LINE Extension tube for infusion pump, phthalate-free	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A03020178			
easyFLOW LINE AMBER Extension tube for infusion pump, amber, phthalate-free	Class IIa	easyFLOW LINE AMBER Extension tube for infusion pump, amber, phthalate-free	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A030201-A8Q			
easyFLOW IS Infusion set easyFLOW IS ECO Infusion set	Class IIa	easyFLOW IS Infusion set easyFLOW IS ECO Infusion set	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A03010103-PHT6H			
easyFLOW IS Infusion set, phthalate-free easyFLOW IS ECO Infusion set, phthalate-free easyFLOW IS PREMIUM Infusion set, phthalate-free	Class IIa	easyFLOW IS Infusion set, phthalate-free easyFLOW IS ECO Infusion set, phthalate-free easyFLOW IS PREMIUM Infusion set, phthalate-free	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A030101037U			
easyFLOW IS SAFE Safety infusion set, phthalate-free	Class IIa	easyFLOW IS SAFE Safety	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>easyFLOW IS SAFE PREMIUM Safety infusion set, phthalate-free</p> <p>Basic UDI-DI: 59079968A03010103-SG2</p>		<p>infusion set, phthalate-free easyFLOW IS SAFE PREMIUM Safety infusion set, phthalate-free</p>	
<p>easyFLOW IS REG Infusion set with precision flow rate regulator, phthalate-free</p> <p>Basic UDI-DI: 59079968A03010103-RFY</p>	Class IIa	<p>easyFLOW IS REG Infusion set with precision flow rate regulator, phthalate-free</p>	DD 1023663-1 NB 0197
<p>easyFLOW IS AMBER Infusion set, amber, phthalate-free</p> <p>Basic UDI-DI: 59079968A03010103-AEW</p>	Class IIa	<p>easyFLOW IS AMBER Infusion set, amber, phthalate-free</p>	DD 1023663-1 NB 0197
<p>ENDOTRACHEAL TUBE UNCUFFED</p> <p>Basic UDI-DI: 59079968R010301FQ</p>	Class IIa	<p>ENDOTRACHEAL TUBE UNCUFFED</p>	DD 1023663-1 NB 0197
<p>ENDOTRACHEAL TUBE CUFFED</p> <p>Basic UDI-DI: 59079968R010302FS</p>	Class IIa	<p>ENDOTRACHEAL TUBE CUFFED</p>	DD 1023663-1 NB 0197
<p>REINFORCED ENDOTRACHEAL TUBE CUFFED WITH PRELOADED STYLET</p> <p>Basic UDI-DI: 59079968R010302-RMF</p>	Class IIa	<p>REINFORCED ENDOTRACHEAL TUBE CUFFED WITH PRELOADED STYLET</p>	DD 1023663-1 NB 0197
<p>BREATHING CIRCUIT FOR CHILDREN with bag, expandable, additional arm BREATHING CIRCUIT FOR ADULTS with bag, expandable, additional arm</p> <p>Basic UDI-DI: 59079968R0201-BGG</p>	Class IIa	<p>BREATHING CIRCUIT FOR CHILDREN with bag, expandable, additional arm BREATHING CIRCUIT FOR ADULTS with bag, expandable, additional arm</p>	DD 1023663-1 NB 0197
<p>BREATHING CIRCUIT FOR CHILDREN BREATHING CIRCUIT FOR ADULTS</p>	Class IIa	<p>BREATHING CIRCUIT FOR CHILDREN BREATHING CIRCUIT FOR ADULTS</p>	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968R0201Q8			
CATHETER MOUNT with double swivel elbow connector, smooth-bore Basic UDI-DI: 59079968R020202-SMP	Class IIa	CATHETER MOUNT with double swivel elbow connector, smooth-bore	DD 1023663-1 NB 0197
CATHETER MOUNT with double swivel elbow connector, expandable Basic UDI-DI: 59079968R020202-ELT	Class IIa	CATHETER MOUNT with double swivel elbow connector, expandable	DD 1023663-1 NB 0197
CATHETER MOUNT with double swivel elbow connector, corrugated Basic UDI-DI: 59079968R020202-CLP	Class IIa	CATHETER MOUNT with double swivel elbow connector, corrugated	DD 1023663-1 NB 0197
CATHETER MOUNT with straight connector, smooth-bore Basic UDI-DI: 59079968R020201-SMJ	Class IIa	CATHETER MOUNT with straight connector, smooth-bore	DD 1023663-1 NB 0197
CATHETER MOUNT with straight connector, corrugated Basic UDI-DI: 59079968R020201-CLJ	Class IIa	CATHETER MOUNT with straight connector, corrugated	DD 1023663-1 NB 0197
CATHETER MOUNT with straight connector, expandable Basic UDI-DI: 59079968R020201-ELN	Class IIa	CATHETER MOUNT with straight connector, expandable	DD 1023663-1 NB 0197
CATHETER MOUNT with elbow connector, smooth-bore Basic UDI-DI: 59079968R0202-SHP	Class IIa	CATHETER MOUNT with elbow connector, smooth-bore	DD 1023663-1 NB 0197
CATHETER MOUNT with elbow connector, corrugated Basic UDI-DI: 59079968R0202-CGP	Class IIa	CATHETER MOUNT with elbow connector, corrugated	DD 1023663-1 NB 0197
CATHETER MOUNT with elbow connector, expandable	Class IIa	CATHETER MOUNT with elbow	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968R0202-EGT		connector, expandable	
TRACHEOSTOMY TUBE cuffed	Class IIa	TRACHEOSTOMY TUBE cuffed	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968R010502G4			
TRACHEOSTOMY TUBE uncuffed	Class IIa	TRACHEOSTOMY TUBE uncuffed	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968R010501G2			
LARYNGEAL MASK, PVC, disposable	Class IIa	LARYNGEAL MASK, PVC, disposable	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968R0102-PH6			
LARYNGEAL MASK, silicone, disposable	Class IIa	LARYNGEAL MASK, silicone, disposable	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968R0102-SHC			
AIR CUSHION ANAESTHETIC MASK	Class IIa	AIR CUSHION ANAESTHETIC MASK	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968R030101-CLQ			
ANAESTHETIC MASK with open seal	Class IIa	ANAESTHETIC MASK with open seal	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968R030101-OMG			
duoNEX Single use syringe, 2-part	Class IIa	duoNEX Single use syringe, 2-part	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A0201020101DK			
dicoNEX Single use syringe, 3-part (luer)	Class IIa	dicoNEX Single use syringe, 3-part (luer)	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A0201020102DM			
Apteczka ABC Strzykawka 3-częściowa	Class IIa	Apteczka ABC Strzykawka 3-częściowa	DD 1023663-1 NB 0197
Basic UDI-DI: 59079968A0201020102DM			
dicoNEX Single use syringe, 3-part (luer lock)	Class IIa	dicoNEX Single use syringe, 3-part (luer lock)	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968A0201020201DQ			
dicoNEX Single use amber syringe, 3-part (luer lock) Basic UDI-DI: 59079968A0201020201-AVY	Class IIa	dicoNEX Single use amber syringe, 3-part (luer lock)	DD 1023663-1 NB 0197
dicoNEX Single use catheter syringe, 3-part Basic UDI-DI: 59079968A020102037G	Class IIa	dicoNEX Single use catheter syringe, 3-part	DD 1023663-1 NB 0197
dicoNEX MN Single use syringe, 3-piece with mounted needle (luer) dicoNEX SN Single use syringe, 3-piece, with needle alongside the syringe (luer) Basic UDI-DI: 59079968A0201020102-IWA	Class IIa	dicoNEX MN Single use syringe, 3-piece with mounted needle (luer) dicoNEX SN Single use syringe, 3-piece, with needle alongside the syringe (luer)	DD 1023663-1 NB 0197
dicoNEX MN Single use syringe, 3-piece with mounted needle (luer lock) dicoNEX SN Single use syringe, 3-piece, with needle alongside the syringe (luer lock) Basic UDI-DI: 59079968A0201020201-IWG	Class IIa	dicoNEX MN Single use syringe, 3-piece with mounted needle (luer lock) dicoNEX SN Single use syringe, 3-piece, with needle alongside the syringe (luer lock)	DD 1023663-1 NB 0197
dicoNEX MN Single use amber syringe, 3-piece with mounted needle (luer lock) dicoNEX SN Single use amber syringe, 3-piece, with needle alongside the syringe (luer lock) Basic UDI-DI: 59079968A0201020201-IA9D	Class IIa	dicoNEX MN Single use amber syringe, 3-piece with mounted needle (luer lock) dicoNEX SN Single use amber syringe, 3-piece, with needle alongside the syringe (luer lock)	DD 1023663-1 NB 0197
dicoSULIN Insulin syringe	Class IIa	dicoSULIN Insulin syringe	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968A02010672			
dicoTUBER Tuberculin syringe Basic UDI-DI: 59079968A02010978	Class IIa	dicoTUBER Tuberculin syringe	DD 1023663-1 NB 0197
dispoFINE Injection needle Basic UDI-DI: 59079968A0101010102CK	Class IIa	dispoFINE Injection needle	DD 1023663-1 NB 0197
dispoGUARD Safety injection needle Basic UDI-DI: 59079968A0101010101CH	Class IIa	dispoGUARD Safety injection needle	DD 1023663-1 NB 0197
dispoSULIN Insulin pen needle Basic UDI-DI: 59079968A01010101026Q	Class IIa	dispoSULIN Insulin pen needle	DD 1023663-1 NB 0197
easyFLOW TS Transfusion set Basic UDI-DI: 59079968A03010102- PHT66	Class IIa	easyFLOW TS Transfusion set	DD 1023663-1 NB 0197
easyFLOW TS Transfusion set, phthalate-free easyFLOW TS PREMIUM Transfusion set, phthalate-free Basic UDI-DI: 59079968A030101027S	Class IIa	easyFLOW TS Transfusion set, phthalate-free easyFLOW TS PREMIUM Transfusion set, phthalate-free	DD 1023663-1 NB 0197
NEEDLE FREE VALVE blue Basic UDI-DI: 59079968A0705KE	Class IIa	NEEDLE FREE VALVE blue	DD 1023663-1 NB 0197
NEEDLE FREE VALVE transparent Basic UDI-DI: 59079968A07050295	Class IIa	NEEDLE FREE VALVE transparent	DD 1023663-1 NB 0197
NEEDLE FREE VALVE transparent with extension line, single NEEDLE FREE VALVE transparent with extension line, double	Class IIa	NEEDLE FREE VALVE transparent with extension line, single NEEDLE FREE VALVE transparent	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>NEEDLE FREE VALVE transparent with extension line, triple NEEDLE FREE VALVE transparent with extension line, quadruple</p> <p>Basic UDI-DI: 59079968A070502-LCQ</p>		<p>with extension line, double NEEDLE FREE VALVE transparent with extension line, triple NEEDLE FREE VALVE transparent with extension line, quadruple</p>	
<p>safeCARE Surgical gloves, latex, powdered, sterile</p> <p>Basic UDI-DI: 59079968T01010101-RYM</p>	Class IIa	safeCARE Surgical gloves, latex, powdered, sterile	DD 1023663-1 NB 0197
<p>safeCARE PF Surgical gloves, latex, powder free, sterile</p> <p>Basic UDI-DI: 59079968T01010102-RYS</p>	Class IIa	safeCARE PF Surgical gloves, latex, powder free, sterile	DD 1023663-1 NB 0197
<p>safeCARE basic Surgical gloves latex, powdered, sterile</p> <p>Basic UDI-DI: 59079968T01010101-RYM</p>	Class IIa	safeCARE basic Surgical gloves latex, powdered, sterile	DD 1023663-1 NB 0197
<p>safeCARE basic PF Surgical gloves latex, powder-free, sterile</p> <p>Basic UDI-DI: 59079968T01010102-RYS</p>	Class IIa	safeCARE basic PF Surgical gloves latex, powder-free, sterile	DD 1023663-1 NB 0197
<p>safeCARE premium Surgical gloves latex, powder-free, sterile safeCARE UG Surgical gloves latex, powder-free, sterile safeCARE micro Surgical gloves latex, powder-free, sterile safeCARE ortho Surgical gloves latex, powder-free, sterile safeCARE dual Surgical gloves latex, powder-free, sterile</p> <p>Basic UDI-DI: 59079968T01010102-RYS</p>	Class IIa	safeCARE premium Surgical gloves latex, powder-free, sterile safeCARE UG Surgical gloves latex, powder-free, sterile safeCARE micro Surgical gloves latex, powder-free, sterile safeCARE ortho Surgical gloves latex, powder-free, sterile safeCARE dual Surgical gloves latex, powder-free, sterile	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
		latex, powder-free, sterile	
safeCARE synthetic Surgical gloves neoprene, powder-free, sterile safeCARE synthetic UG Surgical gloves neoprene, powder-free, sterile Basic UDI-DI: 59079968T010102-NRWL	Class IIa	safeCARE synthetic Surgical gloves neoprene, powder-free, sterile safeCARE synthetic UG Surgical gloves neoprene, powder-free, sterile	DD 1023663-1 NB 0197
safeCARE fusion Surgical gloves polyisoprene, powder-free, sterile Basic UDI-DI: 59079968T010102-PRWS	Class IIa	safeCARE fusion Surgical gloves polyisoprene, powder-free, sterile	DD 1023663-1 NB 0197
safeCARE virtuo Surgical gloves flexylon, powder-free, sterile safeCARE virtuo UG Surgical gloves flexylon, powder-free, sterile safeCARE pro protect Surgical gloves flexylon, powder-free, sterile Basic UDI-DI: 59079968T010102-FRVU	Class IIa	safeCARE virtuo Surgical gloves flexylon, powder-free, sterile safeCARE virtuo UG Surgical gloves flexylon, powder-free, sterile safeCARE pro protect Surgical gloves flexylon, powder-free, sterile	DD 1023663-1 NB 0197
safeLANCE Pressure-activated safety lancet Basic UDI-DI: 59079968V0104RM	Class IIa	safeLANCE Pressure-activated safety lancet	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-SETA	Class IIa	deltaset Urinary bladder catheterization kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-SHTG	Class IIa	deltaset Urinary bladder catheterization kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-CERS	Class IIa	deltaset Dialysis kit	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-CHRY	Class IIa	deltaset Dialysis kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-WETN	Class IIa	deltaset Dressing change kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-WHTU	Class IIa	deltaset Dressing change kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-RET7	Class IIa	deltaset Suture application kit deltaset Suture removal kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-RHTD	Class IIa	deltaset Suture application kit deltaset Suture removal kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-IESC	Class IIa	deltaset Anesthesia kit deltaset Puncture kit	DD 1023663-1 NB 0197
deltaset Procedure kit O Basic UDI-DI: 59079968V0599-IHSJ	Class IIa	deltaset Anesthesia kit deltaset Puncture kit	DD 1023663-1 NB 0197
elastopor STERIL Non-woven dressing with absorbent pad, self-adhesive, sterile elastoKIDS STERIL Non-woven dressing with absorbent pad, self-adhesive, sterile Basic UDI-DI: 59079968M040101-WJW	Class I devices placed on the market in sterile condition	elastopor STERIL Non-woven dressing with absorbent pad, self-adhesive, sterile elastoKIDS STERIL Non-woven dressing with absorbent pad, self-adhesive, sterile	DD 1023663-1 NB 0197
elastoDERM PAD Foil dressing, with absorbent pad, self-adhesive, sterile	Class I devices placed on the market in sterile condition	elastoDERM PAD Foil dressing, with absorbent pad,	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968M040101-FHU		self-adhesive, sterile	
elastoSTRIP Wound closure strips, sterile Basic UDI-DI: 59079968M040499FL	Class I devices placed on the market in sterile condition	elastoSTRIP Wound closure strips, sterile	DD 1023663-1 NB 0197
UMBILICAL CORD CLAMP, sterile Basic UDI-DI: 59079968V0202RN	Class I devices placed on the market in sterile condition	UMBILICAL CORD CLAMP, sterile	DD 1023663-1 NB 0197
ALPHAtex Procedure gown NORMAL, sterile Basic UDI-DI: 59079968T0205R6	Class I devices placed on the market in sterile condition	ALPHAtex Surgical gown NORMAL, sterile	DD 1023663-1 NB 0197
ALPHAtex Procedure gown NORMAL-P Basic UDI-DI: 59079968T0205R6	Class I devices placed on the market in sterile condition	ALPHAtex Surgical gown NORMAL-P, sterile	HD 1023663-1 NB 0197
ALPHAtex Surgical gown STANDARD, sterile ALPHAtex Surgical gown COMFORT sterile Basic UDI-DI: 59079968T020401HA	Class I devices placed on the market in sterile condition	ALPHAtex Surgical gown STANDARD, sterile ALPHAtex Surgical gown COMFORT sterile	DD 1023663-1 NB 0197
ALPHAtex Surgical gown CLASSIC-P ALPHAtex Surgical gown STANDARD-P ALPHAtex Surgical gown COMFORT-P Basic UDI-DI: 59079968T020401HA	Class I devices placed on the market in sterile condition	ALPHAtex Surgical gown CLASSIC-P, sterile ALPHAtex Surgical gown STANDARD-P, sterile ALPHAtex Surgical gown COMFORT-P sterile	HD 1023663-1 NB 0197
ALPHAtex Surgical gown STANDARD PLUS, with impermeable parts, sterile ALPHAtex Surgical gown COMFORT PLUS, with impermeable parts, sterile ALPHAtex Surgical gown EXTRA SAFE, with impermeable parts, sterile	Class I devices placed on the market in sterile condition	ALPHAtex Surgical gown STANDARD PLUS, with impermeable parts, sterile ALPHAtex Surgical gown COMFORT PLUS, with impermeable parts, sterile ALPHAtex Surgical gown EXTRA SAFE, with	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968T020402HC		impermeable parts, sterile	
ALPHAtex Surgical gown STANDARD PLUS-P, with impermeable parts ALPHAtex Surgical gown COMFORT PLUS-P, with impermeable parts ALPHAtex Surgical gown EXTRA SAFE-P, with impermeable parts Basic UDI-DI: 59079968T020402HC	Class I devices placed on the market in sterile condition	ALPHAtex Surgical gown STANDARD PLUS-P, with impermeable parts, sterile ALPHAtex Surgical gown COMFORT PLUS-P, with impermeable parts, sterile ALPHAtex Surgical gown EXTRA SAFE-P, with impermeable parts, sterile	HD 1023663-1 NB 0197
ALPHAtex Surgical drape, sterile ALPHAtex 2-layer surgical drape, with cellulose layer, sterile ALPHAtex 2-layer surgical drape, sterile ALPHAtex 2-layer surgical drape with adhesive edge, sterile ALPHAtex 2-layer surgical drape with central fenestration, sterile ALPHAtex 2-layer surgical drape with central adhesive fenestration, sterile ALPHAtex 2-layer surgical drape with central, adjustable adhesive fenestration, sterile ALPHAtex 2-layer epidural surgical drape with adhesive fenestration, sterile ALPHAtex 3-layer surgical drape, sterile ALPHAtex 3-layer surgical drape with adhesive edge, sterile ALPHAtex 3-layer surgical drape with	Class I devices placed on the market in sterile condition	ALPHAtex Surgical drape, sterile ALPHAtex 2-layer surgical drape, with cellulose layer, sterile ALPHAtex 2-layer surgical drape, sterile ALPHAtex 2-layer surgical drape with adhesive edge, sterile ALPHAtex 2-layer surgical drape with central fenestration, sterile ALPHAtex 2-layer surgical drape with central adhesive fenestration, sterile ALPHAtex 2-layer surgical drape with central, adjustable adhesive fenestration, sterile ALPHAtex 2-layer epidural surgical drape with adhesive fenestration, sterile ALPHAtex 3-layer surgical drape, sterile ALPHAtex 3-layer surgical drape with adhesive fenestration, sterile ALPHAtex 3-layer surgical drape, sterile	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>central fenestration, sterile ALPHAtex 3-layer surgical drape with central adhesive fenestration, sterile ALPHAtex Reinforced surgical drape with adhesive fenestration, sterile</p> <p>Basic UDI-DI: 59079968T0201QW</p>		<p>ALPHAtex 3-layer surgical drape with adhesive edge, sterile ALPHAtex 3-layer surgical drape with central fenestration, sterile ALPHAtex 3-layer surgical drape with central adhesive fenestration, sterile ALPHAtex Reinforced surgical drape with adhesive fenestration, sterile</p>	
<p>ALPHAtex Surgical drape ALPHAtex 2-layer surgical drape, with cellulose layer ALPHAtex 2-layer surgical drape ALPHAtex 2-layer surgical drape with adhesive edge ALPHAtex 2-layer surgical drape with central fenestration ALPHAtex 2-layer surgical drape with central adhesive fenestration ALPHAtex 3-layer surgical drape ALPHAtex 3-layer surgical drape with adhesive edge ALPHAtex 3-layer surgical drape with central fenestration ALPHAtex 3-layer surgical drape with central adhesive fenestration</p> <p>Basic UDI-DI: 59079968T0201QW</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>ALPHAtex Surgical drape, sterile ALPHAtex 2-layer surgical drape, with cellulose layer, sterile ALPHAtex 2-layer surgical drape, sterile ALPHAtex 2-layer surgical drape with adhesive edge, sterile ALPHAtex 2-layer surgical drape with central fenestration, sterile ALPHAtex 2-layer epidural surgical drape with adhesive fenestration, sterile ALPHAtex 3-layer surgical drape, sterile ALPHAtex 3-layer surgical drape with adhesive edge, sterile ALPHAtex 3-layer surgical drape with central fenestration, sterile ALPHAtex 3-layer surgical drape with</p>	<p>HD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
		central adhesive fenestration, sterile	
ALPHAtex Instrument table cover, sterile Basic UDI-DI: 59079968T030101-INJ	Class I devices placed on the market in sterile condition	ALPHAtex Instrument table cover, sterile	DD 1023663-1 NB 0197
ALPHAtex Reinforced Mayo stand cover, sterile ALPHAtex Reinforced Mayo stand cover, red, sterile Basic UDI-DI: 59079968T030101-MNS	Class I devices placed on the market in sterile condition	ALPHAtex Reinforced Mayo stand cover, sterile ALPHAtex Reinforced Mayo stand cover, red, sterile	DD 1023663-1 NB 0197
ALPHAtex Armboard cover, sterile ALPHAtex Surgical pocket for syringes, sterile Basic UDI-DI: 59079968T030101-NNU	Class I devices placed on the market in sterile condition	ALPHAtex Armboard cover, sterile ALPHAtex Surgical pocket for syringes, sterile	DD 1023663-1 NB 0197
ALPHAtex Camera cables cover, sterile ALPHAtex Circular banded cover for medical devices, sterile ALPHAtex Square banded cover for medical devices, sterile ALPHAtex C-arm cover set, sterile ALPHAtex Lamp handle cover, sterile ALPHAtex Ultrasound cover kits, sterile Basic UDI-DI: 59079968T030101-FNC	Class I devices placed on the market in sterile condition	ALPHAtex Camera cables cover, sterile ALPHAtex Circular banded cover for medical devices, sterile ALPHAtex Square banded cover for medical devices, sterile ALPHAtex C-arm cover set, sterile ALPHAtex Lamp handle cover, sterile ALPHAtex Ultrasound cover kits, sterile	DD 1023663-1 NB 0197
ALPHAtex Absorbent drape, sterile ALPHAtex Absorbent drape for newborn, sterile Basic UDI-DI: 59079968T020199-SRU	Class I devices placed on the market in sterile condition	ALPHAtex Absorbent drape, sterile ALPHAtex Absorbent drape for newborn, sterile	DD 1023663-1 NB 0197
ALPHAtex Absorbent drape, sterile	Class I devices placed on the	ALPHAtex Absorbent drape, sterile	HD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
ALPHAtex Absorbent drape for newborn, sterile Basic UDI-DI: 59079968T020199-SRU	market in sterile condition	ALPHAtex Absorbent drape for newborn, sterile	
ALPHAtex Limb cover, sterile ALPHAtex Head Turban drape, sterile ALPHAtex Under buttocks drape, sterile Basic UDI-DI: 59079968T020102GV	Class I devices placed on the market in sterile condition	ALPHAtex Limb cover, sterile ALPHAtex Head Turban drape, sterile ALPHAtex Under buttocks drape, sterile	DD 1023663-1 NB 0197
ALPHAtex Adhesive pouch, one-chamber, sterile ALPHAtex Adhesive pouch, two-chamber, sterile ALPHAtex Adhesive pouch, three-chamber, sterile ALPHAtex Arthroscopy fluid collection pouch, cone-shaped, with outlet tip, sterile ALPHAtex Fluid collection pouch, cone-shaped, with outlet tip, sterile Basic UDI-DI: 59079968T020199-PRN	Class I devices placed on the market in sterile condition	ALPHAtex Adhesive pouch, one-chamber, sterile ALPHAtex Adhesive pouch, two-chamber, sterile ALPHAtex Arthroscopy fluid collection pouch, cone-shaped, with outlet tip, sterile ALPHAtex Fluid collection pouch, cone-shaped, with outlet tip, sterile	DD 1023663-1 NB 0197
ALPHAtex Non-woven surgical tape, adhesive, sterile ALPHAtex Velcro surgical tape, sterile Basic UDI-DI: 59079968T020199-TRW	Class I devices placed on the market in sterile condition	ALPHAtex Non-woven surgical tape, adhesive, sterile ALPHAtex Velcro surgical tape, sterile	DD 1023663-1 NB 0197
ALPHAtex Abdominal drape, sterile ALPHAtex Abdo-Perineal drape, sterile ALPHAtex Angiography drape, sterile ALPHAtex Cardiology drape, sterile ALPHAtex Cardiac drape, sterile	Class I devices placed on the market in sterile condition	ALPHAtex Abdominal drape, sterile ALPHAtex Abdo-Perineal drape, sterile ALPHAtex Angiography drape, sterile	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>ALPHAtex C-section drape, sterile ALPHAtex Delivery drape, sterile ALPHAtex Extremity drape, sterile ALPHAtex Gynaecology drape, sterile ALPHAtex Laparoscopy drape, sterile ALPHAtex Ophthalmic drape, sterile ALPHAtex Orthopaedic drape, sterile ALPHAtex Shoulder drape, sterile ALPHAtex Vertical isolation drape, sterile</p> <p>Basic UDI-DI: 59079968T0202QY</p>		<p>ALPHAtex Cardiology drape, sterile ALPHAtex Cardiac drape, sterile ALPHAtex C-section drape, sterile ALPHAtex Delivery drape, sterile ALPHAtex Extremity drape, sterile ALPHAtex Gynaecology drape, sterile ALPHAtex Laparoscopy drape, sterile ALPHAtex Ophthalmic drape, sterile ALPHAtex Orthopaedic drape, sterile ALPHAtex Shoulder drape, sterile ALPHAtex Vertical isolation drape, sterile</p>	
<p>ALPHAtex Abdominal drape, sterile ALPHAtex Abdo-Perineal drape, sterile ALPHAtex Angiography drape, sterile ALPHAtex Cardiology drape, sterile ALPHAtex Cardiac drape, sterile ALPHAtex C-section drape, sterile ALPHAtex Delivery drape, sterile ALPHAtex Extremity drape, sterile ALPHAtex Gynaecology drape, sterile ALPHAtex Laparoscopy drape, sterile ALPHAtex Ophthalmic drape, sterile</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>ALPHAtex Abdominal drape (from 1 to 100) ALPHAtex Abdo-Perineal drape (from 1 to 100) ALPHAtex Angiography drape (from 1 to 100) ALPHAtex Cardiology drape (from 1 to 100) ALPHAtex Cardiac drape (from 1 to 100) ALPHAtex C-section drape (from 1 to 100) ALPHAtex Delivery drape (from 1 to 100)</p>	<p>HD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>ALPHAtex Orthopaedic drape, sterile ALPHAtex Shoulder drape, sterile ALPHAtex Vertical isolation drape, sterile</p> <p>Basic UDI-DI: 59079968T0202QY</p>		<p>ALPHAtex Extremity drape (from 1 to 100) ALPHAtex Gynaecology drape (from 1 to 100) ALPHAtex Laparoscopy drape (from 1 to 100) ALPHAtex Ophthalmic drape (from 1 to 100) ALPHAtex Orthopaedic drape (from 1 to 100) ALPHAtex Shoulder drape (from 1 to 100) ALPHAtex Vertical isolation drape (from 1 to 100)</p>	
<p>ALPHAtex Abdominal set, sterile ALPHAtex Abdo-Perineal set, sterile ALPHAtex Ablation set, sterile ALPHAtex Angiography set, sterile ALPHAtex Arthroscopy set, sterile ALPHAtex Basic set, sterile ALPHAtex Cardiology set, sterile ALPHAtex Cardiac set, sterile ALPHAtex Craniotomy set, sterile ALPHAtex C-section set, sterile ALPHAtex Cystoscopy set, sterile ALPHAtex Delivery set, sterile ALPHAtex Dental set, sterile ALPHAtex Dynamic hip screw set, sterile ALPHAtex Extremity set, sterile</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>ALPHAtex Abdominal set, sterile ALPHAtex Abdo-Perineal set, sterile ALPHAtex Ablation set, sterile ALPHAtex Angiography set, sterile ALPHAtex Arthroscopy set, sterile ALPHAtex Cardiology set, sterile ALPHAtex Cardiac set, sterile ALPHAtex Craniotomy set, sterile ALPHAtex Cardiac set, sterile ALPHAtex Craniotomy set, sterile ALPHAtex C-section set, sterile ALPHAtex Cystoscopy set, sterile ALPHAtex Delivery set, sterile</p>	<p>DD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>ALPHAtex Gynaecology set, sterile ALPHAtex Hip set, sterile ALPHAtex Laparoscopy set, sterile ALPHAtex Laryngology set, sterile ALPHAtex Ophthalmic set, sterile ALPHAtex Otolaryngology set, sterile ALPHAtex Pediatric set, sterile ALPHAtex Percutaneous lithotripsy set, sterile ALPHAtex Shoulder set, sterile ALPHAtex Spine set, sterile ALPHAtex Thyroid set, sterile ALPHAtex TUR set, sterile ALPHAtex Universal set, sterile ALPHAtex Uro/gynaecology set, sterile ALPHAtex Varicose vein set, sterile ALPHAtex Vertical isolation set, sterile</p> <p>Basic UDI-DI: 59079968T0202QY</p>		<p>ALPHAtex Dental set, sterile ALPHAtex Dynamic hip screw set, sterile ALPHAtex Extremity set, sterile ALPHAtex Gynaecology set, sterile ALPHAtex Hip set, sterile ALPHAtex Laparoscopy set, sterile ALPHAtex Laryngology set, sterile ALPHAtex Ophthalmic set, sterile ALPHAtex Otolaryngology set, sterile ALPHAtex Pediatric set, sterile ALPHAtex Percutaneous lithotripsy set, sterile ALPHAtex Shoulder set, sterile ALPHAtex Spine set, sterile ALPHAtex Thyroid set, sterile ALPHAtex TUR set, sterile ALPHAtex Universal set, sterile ALPHAtex Uro/gynaecology set, sterile ALPHAtex Varicose vein set, sterile ALPHAtex Vertical isolation set, sterile</p>	

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>ALPHAtex Abdominal set, sterile ALPHAtex Abdo-Perineal set, sterile ALPHAtex Ablation set, sterile ALPHAtex Angiography set, sterile ALPHAtex Arthroscopy set, sterile ALPHAtex Basic set, sterile ALPHAtex Cardiology set, sterile ALPHAtex Cardiac set, sterile ALPHAtex Craniotomy set, sterile ALPHAtex C-section set, sterile ALPHAtex Cystoscopy set, sterile ALPHAtex Delivery set, sterile ALPHAtex Dental set, sterile ALPHAtex Dynamic hip screw set, sterile ALPHAtex Extremity set, sterile ALPHAtex Gynaecology set, sterile ALPHAtex Hip set, sterile ALPHAtex Laparoscopy set, sterile ALPHAtex Laryngology set, sterile ALPHAtex Ophthalmic set, sterile ALPHAtex Otolaryngology set, sterile ALPHAtex Pediatric set, sterile ALPHAtex Percutaneous lithotripsy set, sterile ALPHAtex Shoulder set, sterile ALPHAtex Spine set, sterile ALPHAtex Thyroid set, sterile</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>"ALPHAtex Abdominal set (from 1 to 200) ALPHAtex Abdo-Perineal set (from 1 to 200) ALPHAtex Ablation set (from 1 to 200) ALPHAtex Angiography set (from 1 to 200) ALPHAtex Arthroscopy set (from 1 to 200) ALPHAtex Basic set (from 1 to 200) ALPHAtex Cardiology set (from 1 to 200) ALPHAtex Cardiac set (from 1 to 200) ALPHAtex Craniotomy set (from 1 to 200) ALPHAtex C-section set (from 1 to 200) ALPHAtex Cardiac set (from 1 to 200) ALPHAtex Craniotomy set (from 1 to 200) ALPHAtex C-section set (from 1 to 200) ALPHAtex Cystoscopy set (from 1 to 200) ALPHAtex Delivery set (from 1 to 200) ALPHAtex Dental set (from 1 to 200) ALPHAtex Dynamic hip screw set (from 1 to 200) ALPHAtex Extremity set (from 1 to 200) ALPHAtex Gynaecology set (from 1 to 200) ALPHAtex Hip set (from 1 to 200) ALPHAtex Laparoscopy set (from 1 to 200) ALPHAtex Laryngology set (from 1 to 200)</p>	<p>HD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>ALPHAtex TUR set, sterile ALPHAtex Universal set, sterile ALPHAtex Uro/gynaecology set, sterile ALPHAtex Varicose vein set, sterile ALPHAtex Vertical isolation set, sterile</p> <p>Basic UDI-DI: 59079968T0202QY</p>		<p>ALPHAtex Ophthalmic set (from 1 to 200) ALPHAtex Otolaryngology set (from 1 to 200) ALPHAtex Pediatric set (from 1 to 200) ALPHAtex Percutaneous lithotripsy set (from 1 to 200) ALPHAtex Shoulder set (from 1 to 200) ALPHAtex Spine set (from 1 to 200) ALPHAtex Thyroid set (from 1 to 200) ALPHAtex TUR set (from 1 to 200) ALPHAtex Universal set (from 1 to 200) ALPHAtex Uro/gynaecology set (from 1 to 200) ALPHAtex Varicose vein set (from 1 to 200) ALPHAtex Vertical isolation set (from 1 to 200)"</p>	
<p>elastoLUMENAL S Eye pad, superabsorbent, multi-layer, sterile</p> <p>Basic UDI-DI: 59079968M040301-SKC</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>elastoLUMENAL S Eye pad, superabsorbent, multi-layer, sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>elastopor EYE Eye dressing, non-woven, with absorbent pad, self-adhesive, sterile elastoKIDS EYE Eye dressing, non-woven, with absorbent pad, self-adhesive, sterile</p> <p>Basic UDI-DI: 59079968M0403NX</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>lastopor EYE Eye dressing, non-woven, with absorbent pad, self-adhesive, sterile elastoKIDS EYE Eye dressing, non-woven, with absorbent pad,</p>	<p>DD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
		self-adhesive, sterile	
COMBI STOPPER Basic UDI-DI: 59079968C01018085	Class I devices placed on the market in sterile condition	COMBI STOPPER	DD 1023663-1 NB 0197
LUER LOCK STOPPER Basic UDI-DI: 59079968C01018085	Class I devices placed on the market in sterile condition	LUER LOCK STOPPER	DD 1023663-1 NB 0197
elastopor STERIL D Non-woven dressing with absorbent pad, with incision and O-hole, self-adhesive, sterile elastoKIDS STERIL D Non-woven dressing, with absorbent pad, with incision and X-hole, self-adhesive, sterile Basic UDI-DI: 59079968M04010201-DTG	Class I devices placed on the market in sterile condition	elastopor STERIL D Non-woven dressing with absorbent pad, with incision and O-hole, self-adhesive, sterile elastoKIDS STERIL D Non-woven dressing, with absorbent pad, with incision and X-hole, self-adhesive, sterile	DD 1023663-1 NB 0197
NONVI lux S Non-woven swab, with O-incision, sterile NONVI lux S Non-woven swab, with Y-incision, sterile Basic UDI-DI: 59079968M04010201-NU4	Class I devices placed on the market in sterile condition	NONVI lux S Non-woven swab, with O-incision, sterile NONVI lux S Non-woven swab, with Y-incision, sterile	DD 1023663-1 NB 0197
elastopor IV IV cannula dressing, non-woven, selfadhesive, sterile elastoKIDS IV IV cannula dressing, non-woven, selfadhesive, sterile Basic UDI-DI: 59079968M04010201H2	Class I devices placed on the market in sterile condition	elastopor IV IV cannula dressing, non-woven, selfadhesive, sterile elastoKIDS IV IV cannula dressing, non-woven, selfadhesive, sterile	DD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>elastoDERM F-IV IV cannula dressing, foil, with frame, with U-incision, self-adhesive, sterile elastoDERM IV IV cannula dressing, foil, selfadhesive, sterile elastoDERM Foil dressing, self-adhesive, sterile elastoDERM F Foil dressing, with frame, selfadhesive, sterile elastoDERM C Foil dressing with a pocket to secure catheter, sterile</p> <p>Basic UDI-DI: 59079968M04010202H4</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>elastoDERM F-IV IV cannula dressing, foil, with frame, with U-incision, self-adhesive, sterile elastoDERM IV IV cannula dressing, foil, selfadhesive, sterile elastoDERM Foil dressing, self-adhesive, sterile elastoDERM F Foil dressing, with frame, selfadhesive, sterile elastoDERM C Foil dressing with a pocket to secure catheter, sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>MULTIabsorb S ABD pad, non-woven and cellulose, sterile</p> <p>Basic UDI-DI: 59079968M040201-SJZ</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>MULTIabsorb S ABD pad, non-woven and cellulose, sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>VAGINAL SPECULUM</p> <p>Basic UDI-DI: 59079968U089006MJ</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>VAGINAL SPECULUM</p>	<p>DD 1023663-1 NB 0197</p>
<p>URINE BAG</p> <p>Basic UDI-DI: 59079968A0603038J</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>URINE BAG</p>	<p>DD 1023663-1 NB 0197</p>
<p>URINE BAG with sample port, sterile URINE BAG with sample port 2W, sterile</p> <p>Basic UDI-DI: 59079968A060303-PBW</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>URINE BAG with sample port, sterile URINE BAG with sample port 2W, sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>SAMPLES TAKING URINE BAG for boys, with sponge SAMPLES TAKING URINE BAG for boys, without sponge SAMPLES TAKING URINE BAG for girls, with sponge</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>SAMPLES TAKING URINE BAG for boys, with sponge SAMPLES TAKING URINE BAG for boys, without sponge</p>	<p>DD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>SAMPLES TAKING URINE BAG for girls, without sponge Apteczka ABC Woreczek do pobierania próbek moczu dla dla chłopców z gąbką Apteczka ABC Woreczek do pobierania próbek moczu dla dziewczynek z gąbką</p> <p>Basic UDI-DI: 59079968A06030301AB</p>		<p>SAMPLES TAKING URINE BAG for girls, with sponge</p> <p>SAMPLES TAKING URINE BAG for girls, without sponge</p> <p>Apteczka ABC Woreczek do pobierania próbek moczu dla dla chłopców z gąbką</p> <p>Apteczka ABC Woreczek do pobierania próbek moczu dla dziewczynek z gąbką</p>	
<p>ENEMA BAG sterile</p> <p>Basic UDI-DI: 59079968G020301-SDY</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>ENEMA BAG sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>WOODEN TONGUE DEPRESSOR Sterile</p> <p>Basic UDI-DI: 59079968V9001-SNM</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>WOODEN TONGUE DEPRESSOR sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>NELATON CATHETER NELATON CATHETER transparent</p> <p>Basic UDI-DI: 59079968U010105H9</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>NELATON CATHETER NELATON CATHETER transparent</p>	<p>DD 1023663-1 NB 0197</p>
<p>GUEDEL AIRWAY</p> <p>Basic UDI-DI: 59079968R010102FG</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>GUEDEL AIRWAY</p>	<p>DD 1023663-1 NB 0197</p>
<p>ENDOTRACHEAL TUBE HOLDER, vertical fixation ENDOTRACHEAL TUBE HOLDER, horizontal fixation</p> <p>Basic UDI-DI: 59079968R010380-SNX</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>ENDOTRACHEAL TUBE HOLDER, vertical fixation</p> <p>ENDOTRACHEAL TUBE HOLDER, horizontal fixation</p>	<p>DD 1023663-1 NB 0197</p>
<p>INTUBATION STYLET TRACHEAL TUBE INTRODUCER CURVED</p> <p>Basic UDI-DI: 59079968R010380-PNR</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>INTUBATION STYLET TRACHEAL TUBE INTRODUCER CURVED</p>	<p>DD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
<p>dicoSPIKE Withdrawal cannula with bacteria filter dicoSPIKE Withdrawal cannula with bacteria and particle filter dicoSPIKE Chemo Withdrawal cannula with bacteria and particle filter</p> <p>Basic UDI-DI: 59079968A0704KC</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>dicoSPIKE Withdrawal cannula with bacteria filter dicoSPIKE Withdrawal cannula with bacteria and particle filter dicoSPIKE Chemo Withdrawal cannula with bacteria and particle filter</p>	<p>DD 1023663-1 NB 0197</p>
<p>elastoBAND BASIC S Knitted supporting bandage, sterile</p> <p>Basic UDI-DI: 59079968M030301-SJT</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>elastoBAND BASIC S Knitted supporting bandage, sterile</p>	<p>HD 1023663-1 NB 0197</p>
<p>elastoBAND FLEX S Elastic bandage, sterile</p> <p>Basic UDI-DI: 59079968M030402-SKB</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>elastoBAND FLEX S Elastic bandage, sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>elastoFILM Incise film, self-adhesive, sterile elastoFILM M Incise film, self-adhesive, sterile</p> <p>Basic UDI-DI: 59079968T020101GT</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>elastoFILM Incise film, self-adhesive, sterile elastoFILM M Incise film, self-adhesive, sterile</p>	<p>DD 1023663-1 NB 0197</p>
<p>CERVICAL BRUSH standard CERVICAL BRUSH special</p> <p>Basic UDI-DI: 59079968U089002MA</p>	<p>Class I devices placed on the market in sterile condition</p>	<p>CERVICAL BRUSH standard CERVICAL BRUSH special</p>	<p>DD 1023663-1 NB 0197</p>
<p>omegapack Surgical set B</p> <p>Basic UDI-DI: 59079968V0599-EP2</p>	<p>Class IIb excluding Class IIb implantable non-WET</p>	<p>omegapack Surgical set B Orthopedic surgery set B Universal set B C-section set B Cardiac surgery set B Neurosurgical set B</p>	<p>HD 1023663-1 NB 0197</p>
<p>omegapack Surgical set B</p>	<p>Class IIb excluding Class</p>	<p>omegapack Surgical set B</p>	<p>HD 1023663-1 NB 0197</p>

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Basic UDI-DI: 59079968V0599-KPE	IIb implantable non-WET	Orthopedic surgery set B Universal set B C-section set B Cardiac surgery set B Neurosurgical set B	
omegapack Surgical set Basic UDI-DI: 59079968V0599-ANS	Class IIa	omegapack Surgical set Angiography set C-section set Laparoscopy set Gynecological surgery set Cardiac surgery set Neurosurgical set Orthopedic surgery set Otolaryngologic surgery set Urologic surgery set Delivery set Dressing set Universal set	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-WQ6	Class IIa	deltaset Central venous access kit Neonatal kit Universal kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-IPA	Class IIa	deltaset Universal kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-CNW	Class IIa	deltaset Dialysis kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-NPL	Class IIa	deltaset Universal kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-OPN	Class IIa	deltaset Universal kit	HD 1023663-1 NB 0197

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
deltaset Procedure kit Basic UDI-DI: 59079968V0599-FP4	Class IIa	deltaset Urinary bladder catheterization kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-RPU	Class IIa	deltaset Sewing kit Suture removal kit Dressing change kit Operating field disinfection kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-SPW	Class IIa	deltaset Anesthesia kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-NIT3	Class I devices placed on the market in sterile condition	deltaset Universal kit I	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-UITQ	Class I devices placed on the market in sterile condition	deltaset Universal kit I	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-BIRX	Class I devices placed on the market in sterile condition	deltaset Dialysis kit	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-MISY	Class I devices placed on the market in sterile condition	deltaset Universal kit I	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-DIS5	Class I devices placed on the market in sterile condition	deltaset Operating field disinfection kit Operating field disinfection kit I	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-ZIU7	Class I devices placed on the market in sterile condition	deltaset Suture removal kit I	HD 1023663-1 NB 0197
deltaset Procedure kit Basic UDI-DI: 59079968V0599-PIT9	Class I devices placed on the market in sterile condition	deltaset Protective kit I Hygiene kit I Neonatal kit I	HD 1023663-1 NB 0197

Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
none			

Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action
2024-05-15	ZARYS_CL607_2024-05-15	Initial issue
2024-06-04	ZARYS_CL607_2024-06-04	Update of the device list, minor correction.
2025-02-12	ZARYS_CL607_2025-02-12	Minor corrections of the devices names.
2025-02-14	ZARYS_CL607_2025-02-14	Update of the device list, minor correction.
2025-10-22	ZARYS_CL607_2025-10-22	Update of the device list, minor correction.

Certificate

Quality Management System EN ISO 13485:2016 + AC:2018 + A11:2021 ISO 13485:2016

Registration No.: SX 1023663-1

Certificate Holder: ZARYS International Group
Spółka z ograniczoną odpowiedzialnością,
spółka komandytowa
ul. Pod Borem 18
41-808 Zabrze
Poland

Scope: Design and development, production and distribution of sterile surgical sets, procedure kits, surgical drapes, sets of surgical drapes, surgical gowns and knitted bandages.
Production and distribution of sterile and non-sterile disposable medical devices and non-sterile reusable medical devices.
Distribution of in-vitro medical devices.

The Certification Body of TÜV Rheinland LGA Products GmbH certifies that the organization has established and applies a quality management system for medical devices.

Proof has been furnished that the requirements specified in the abovementioned standard are fulfilled. The quality management system is subject to yearly surveillance.

Report No.: 84984397-20

Effective date: 2025-12-11

Expiry date: 2026-06-08

Issue date: 2025-12-11

Replaces certificate SX 1023663-1 issued 2023-06-06

This certificate can be validated on <https://www.certipedia.com>


Rafał Byczkowski
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

© TÜV, TÜEV and TÜV are registered trademarks. Utilisation and application requires prior approval.

Certificate

Quality Management System EN ISO 13485:2016 + AC:2018 + A11:2021 ISO 13485:2016

Registration No.: SX 1023663-1
Certificate Holder: ZARYS International Group
Spółka z ograniczoną odpowiedzialnością,
spółka komandytowa
ul. Pod Borem 18
41-808 Zabrze
Poland

The scope of certification also covers the following sites:

No.	Facility	Scope
/01	ZARYS International Group Spółka z o.o. sp.k. ul. Pod Borem 18 41-808 Zabrze Poland	Design and development, production and distribution of sterile surgical sets, procedure kits, surgical drapes, sets of surgical drapes, surgical gowns and knitted bandages. Production and distribution of sterile and non-sterile disposable medical devices and non-sterile reusable medical devices. Distribution of in-vitro medical devices.
/02	ZARYS International Group Spółka z o.o. sp.k. ul. Guido Henckela Donnersmarcka 1 41-807 Zabrze Poland	Storage, quality control, release and distribution of medical devices.
/03	ZARYS International Group Spółka z o.o. Produkcja sp.k. ul. Pod Borem 18 41-808 Zabrze Poland	Production of sterile: surgical sets, procedure kits, surgical drapes, sets of surgical drapes, surgical gowns and knitted bandages. Production of non-sterile disposable medical devices.
/04	ZARYS International Group Spółka z o.o. sp.k. ul. Ziemska 44 41-803 Zabrze Poland	Administration. Design and development and distribution of medical devices.

This certificate can be validated on <https://www.certipedia.com>



Certificate No.
NC-3444

CERTIFICATE

Issued for:

ZARYS International Group
spółka z ograniczoną odpowiedzialnością spółka komandytowa

ul. Pod Borem 18
41-808 Zabrze

Management Systems Certification Bureau of Polski Rejestr Statków S.A., al. gen. Józefa Hallera 126, 80-416 Gdańsk, certifies that the Integrated Management System including the Quality Management System and Environmental Management System of the above Organization has been assessed and found to be in accordance with the requirements of:

ISO 9001:2015
ISO 14001:2015

Scope of certification:

**PRODUCTION OF: STERILE MEDICAL DEVICES FOR SINGLE USE,
NON-STERILE MEDICAL DEVICES FOR SINGLE AND REUSABLE USE**

Place of business:

ul. Ziemska 44
41-803 Zabrze, Polska

ul. Guido Henckela Donnersmarcka 1
41-807 Zabrze, Polska

Scope of certification:

**PRODUCTION OF: STERILE MEDICAL DEVICES FOR SINGLE USE,
NON-STERILE MEDICAL DEVICES FOR SINGLE AND REUSABLE USE;
DISTRIBUTION OF: MEDICAL DEVICES, IN-VITRO DIAGNOSTIC DEVICES**

The Certificate is valid until:

16.11.2028

Gdańsk, 17.11.2025



AC 014



Certification Division Director
Przemysław Gałka