

C E R T I F I C A T E
of Conformity



Registration No.: AK 50361192 0001

Report No.: 50054337 001

Holder: Hangzhou Sejoy Electronics & Instruments
Co., Ltd.
Building 2, No. 202, Zhenzhong Road
West Lake Economy & Technology Zone
Hangzhou 310030
P.R. China

Product: Blood Sugar Test Device
(Blood glucose monitoring system)

Identification: Blood Glucose Meter : BG-202; BG-102
Blood Glucose test strips: BS-102; BS-101

Remark: The scope of this evaluation covers only
clause 6.3 of EN ISO 15197:2015 (ISO 15197:2013).
All other requirements and aspects are excluded. This
certificate is based on a review of which the results
provided in Report No. NJZQ160524002(2016-05-24).

Tested acc. to: ISO 15197:2013
EN ISO 15197:2015

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen
is in conformity with the assessment requirement mentioned above. This certificate does not imply
assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of
conformity.

Date: 19.10.2016



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg