

REUSABLE JARS are equipped with screw cover with handle for easy grip, autoclavable silicone o-ring for tight seal, overflow valve system integrated in the cover, clear graduated scale in ml with 100ml or 200ml intervals, CA-MI branding. Suitable for central vacuum systems and CA-MI suction units. Available in two types of polycarbonate, Makrolon® and Apec®.

## REUSABLE JARS in MAKROLON® – FOR STEAM STERILIZATION UP TO 121 °C



	Jar with: cover overflow valve o-ring	Jar	Cover with: overflow valve o-ring	O-ring	Suitable for:
400 ml	RE 210301	RE 210305	RE 210302	RE 210304	EMIVAC
1000 ml	RE 210001/02	RE 210003	RE 210352/01	RE 210354	ASPIRET ASKIR series (all except C30 series) Central Vacuum Plants
2000 ml	RE 210351/01	RE 210353	RE 210352/01	RE 210354	ASPIRET - ASKIR Series HOSPIVAC Series Central Vacuum Plants
4000 ml	RE 210006	RE 210007	RE 210008	RE 210306	HOSPIVAC Series Vuoto centralizzato
5000 ml	RE 210010	RE 210013	RE 210012	RE 210307	HOSPIVAC Series Central Vacuum Plants

## REUSABLE JARS in APEC® 1745 – FOR STEAM STERILIZATION UP TO 143 °C



	Jar with: cover overflow valve o-ring	Jar	Cover with: overflow valve o-ring	O-ring	Suitable for:
1000 ml	RE 210009	RE 210002	RE 210352/02	RE 210354	ASPIRET ASKIR series (all except C30 series) Central Vacuum Plants
2000 ml	RE 210351/05	RE 210353/01	RE 210352/02	RE 210354	ASPIRET - ASKIR Series HOSPIVAC Series Central Vacuum Plants
5000 ml	RE 210010/01	RE 210013/01	RE 210012/01	RE 210307	HOSPIVAC Series Vuoto centralizzato

FLOVAC® DISPOSABLE LINERS equipped with polyethylene disposable liner, hydrophobic filter, antibacterial filter, overflow system, reusable container with clear graduated scale in ml with 50ml intervals. The gelling kit is a powder inside liners with germicidal function turning the sucked liquid into a semisolid mass, preventing cross-contamination risks of staff in charge for hygiene and waste disposal.

## FLOVAC® DISPOSABLE LINERS & CONTAINERS

	Liner with: cover filter		Liner with: cover filter gelling kit		Reusable Container	Suitable for
1000 ml	31848		31858		31843**	ASPIRET ASKIR series (all except C30 series)
1000 ml	31845		31854		31840	Central Vacuum Plants
2000 ml	31846		31855		31841	ASPIRET - ASKIR Series HOSPIVAC Series Central Vacuum Plants
3000 ml	31847		31856		31842	HOSPIVAC Series Central Vacuum Plants

\*\* Requires additional round spacer (SP.0220) when ordered for ASKIR Series

ASPIRET  
ASKIR 20  
ASKIR 30  
ASKIR 230-12V BR  
ASKIR 30 12V




ASKIR 36 BR  
ASKIR 36 LI-ION  
ASKIR 118  
ASKIR 118 BASIC

EMIVAC

ASKIR C30  
ASKIR C30 BR

HOSPIVAC 350  
HOSPIVAC 400  
HOSPIVAC BR

### SET of silicone TUBES, FILTERS and CONICAL CONNECTORS

	Tube Ø 6x10mm Conical connector	RE 210355		RE 210355/01		
	Tube Ø 6x10mm Conical connector Antibacterial filter	SP 0036		SP 0043		
	Tube Ø 8 x 14 mm Conical connector		RE 210355/03		RE 210355/03	RE 210355/03
	Tube Ø 8 x 14 mm Conical connector Antibacterial filter		SP 0036/02		SP 0036/02	SP 0032/01 (for 350 and BR) SP 0032 (for Hospivac 400)
	FLOVAC® liners Tube Ø 6x10mm Conical connector	SP 0158/01				
	FLOVAC® liners Tube Ø 8x14mm Conical connector		SP 0160/01		SP 0160/01	SP 0160/01

Roll of silicone tube Ø 6x10 mm

Length 1m = SP 0045/02 - Length 10m = SP 0045/03 - Length 50m = SP 0045/04

Roll of silicone tube Ø 8x14 mm

Length 1m = SP 0045/05 - Length 10m = SP 0045/06 - Length 50m = SP 0045/07

### MALE CONNECTORS

	Ø 8-9-10 mm (pack of 5's)	SP 0223	SP 0223		SP 0223	SP 0223
--	---------------------------	---------	---------	--	---------	---------

### CONICAL CONNECTORS

	Ø 8-9-10 mm	RE 210410		RE 210410		
	Ø 10-11-12 mm		RE 210420		RE 210420	RE 210420

### FILTERS (Antibacterial and Hydrophobic)

	Ø 64 with 8 mm connector	SP 0046		SP 0046		
	Ø 64 with 11mm connector		SP 0121		SP 0121	SP 0121 (350 and BR only)
	Ø 90 with 11mm connector					SP 0047 (for 400 only)

### ASPIRATION PROBES

	CH20	RE 210400 (10 pcs)	RE 210400 (10 pcs)		RE 210400 (10 pcs)	
--	------	-----------------------	-----------------------	--	-----------------------	--


### YANKAUER CANNULAS

	Yankauer Handle Flat Tip with Hole	2044403	2044403	2044403	2044403	2044403
	Yankauer Handle Crown Tip with Hole	2044401	2044401	2044401	2044401	2044401
	Yankauer Tube L= 180	204413018	204413018	204413018	204413018	204413018

### CATHETER CONTAINER

	Tube of polycarbonate Ø 54 mm by 400 mm length. Fully autoclavable (121°C - 15 min)					000032
--	---	--	--	--	--	--------

### SILICONE FETAL VACUUM CUPS

	Length 210 mm, Ø 50 mm, size XS				VC-95100	VC-95100
	Length 210 mm, Ø 60 mm, size S				VC-95200	VC-95200
	Length 210 mm, Ø 70 mm, size M				VC-95300	VC-95300



Benannt durch/Designated by  
Zentralstelle der Länder  
für Gesundheitsschutz  
bei Arzneimitteln und  
Medizinprodukten  
www.zlg.de  
ZLG-BS-244.10.08



Product Service

# EC Certificate

Production Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex V

(Devices in Class IIa, IIb or III)

**No. G2 063105 0047 Rev. 01**

## Manufacturer:

**CA-MI S.R.L.**

Via Ugo La Malfa, 13  
Frazione Pilastro  
43013 Langhirano (PR)  
ITALY

## Product Category(ies):

**Aerosol Therapy Equipment, Kits for Aerosol Therapy,  
Thermal Water Inhaler, Suction Unit, Surgical Suction  
Equipment, Breast Pump, Kit Accessory for Electric  
Breast Pump, Blood Pressure Monitor, Electronic  
Thermometer, Infrared Thermometer, Tens Device,  
Pulse Oximeter**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex V. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class IIb and III devices an additional Annex III certificate is mandatory. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:G2 063105 0047 Rev. 01](http://www.tuvsud.com/ps-cert?q=cert:G2 063105 0047 Rev. 01)

## Report No.:

ITA1626749

## Valid from:

2021-02-09

## Valid until:

2024-05-26

## Date,

2021-02-09

Christoph Dicks  
Head of Certification/Notified Body



# Certificate

No. Q5 063105 0045 Rev. 02

## Holder of Certificate:



## CA-MI S.R.L.

Via Ugo La Malfa, 13  
Frazione Pilastro  
43013 Langhirano (PR)  
ITALY

## Certification Mark:



## Scope of Certificate:

Design and development, production, sale and after-sales technical assistance of medical equipments for surgery (electrical and manual suction pumps), electric breast pumps, medical devices for breathing (aerosol therapy equipments and thermal water inhaler), medical devices for stimulation (tens) and related accessories. Placing on the market under its own name of devices for monitoring of vital physiological parameters (pulse oximeters, thermometers, blood pressure monitors, sphygmomanometer and stethoscopes), incentive spirometers, devices for phlebology (graduated compression medical stockings) and anti-decubitus mattress.  
Distribution of active and non-active non implantable medical devices.

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:Q5 063105 0045 Rev. 02](http://www.tuvsud.com/ps-cert?q=cert:Q5 063105 0045 Rev. 02)

Report No.: ITA1620011

Valid from: 2021-03-25

Valid until: 2022-07-31

Date, 2021-03-25



Christoph Dicks

Head of Certification/Notified Body

# Certificate

No. Q5 063105 0045 Rev. 02

**Applied Standard(s):** EN ISO 13485:2016  
Medical devices - Quality management systems -  
Requirements for regulatory purposes  
(ISO 13485:2016)  
DIN EN ISO 13485:2016

**Facility(ies):** CA-MI S.R.L.  
Via Ugo La Malfa, 13, Frazione Pilastro, 43013 Langhirano (PR), ITALY

**Design and development, production, sale and after-sales technical assistance of medical equipments for surgery (electrical and manual suction pumps), electric breast pumps, medical devices for breathing (aerosol therapy equipments and thermal water inhaler), medical devices for stimulation (tens) and related accessories. Placing on the market under its own name of devices for monitoring of vital physiological parameters (pulse oximeters, thermometers, blood pressure monitors, sphygmomanometer and stethoscopes), incentive spirometers, devices for phlebology (graduated compression medical stockings) and anti-decubitus mattress. Distribution of active and non-active non implantable medical devices.**

CA-MI S.r.l.  
Via Strada per Parma 34, Frazione Pilastro, 43013 Langhirano (PR), ITALY

**Warehouse of active and non-active non implantable medical devices and components used in production.**

CA-MI S.r.l.  
Via Ugo La Malfa 27, Frazione Pilastro, 43013 Langhirano (PR), ITALY

**Production of medical devices for surgery (electrical and manual suction pumps), warehouse of active and non-active non implantable medical devices and components used in production.**

/