



GLOVES

MEDICAL EXAMINATION GLOVE

NITRILE GLOVES · LATEX GLOVES · VINYL GLOVES



LyncMed Group Corporation

Room 1601, Building No.2, Zhubang 2000 Business Building, Balizhuangxili 99,
Chaoyang District, 100022 Beijing, China.

+86 010 8646 8058
contactus@lyncmed.com
www.lyncmed.com



You're protected.

Our gloves will be manufactured through rigorous tests based on the corresponding regulations. We will ensure the demand and protection in accordance with the highest quality standards.

Quality products

01



02

Integrated logistics service system at home and abroad

The whole network online integration process, convenient, fast and efficient

03



NITRILE GLOVES





LATEX GLOVES





VINYL GLOVES



Clear/Blue/Black






Key Features:

- Outstanding chemical resistance, anti-certain pH, solvent, oil and other corrosive substances to provide good chemical protection.
- Good physical properties, good tear resistance, anti-piercing, anti-friction performance.
- Comfortable style, according to ergonomic design glove palm machine bending fingers to wear comfortable, conducive to blood circulation. Does not contain protein, amino compounds and other harmful substances, rarely produce allergies.
- Widely used in electronics factories, medical inspection, food industry, domestic labor, chemicals, aquaculture, glass products and scientific research and other industries.


Serves Worldwide



LyncMed has sales offices across the world including Italy, Spain, Germany, Mexico, American, Portugal and Turkey with additional warehouses in strategically placed locations to support distributions on a global scale.

 **LyncMed Medical Technology (Beijing) Co., Ltd.**

ADD: Room 1601, Building 2, ZhuBang 2000 Business Center, ChaoYang District, BJ, China.
 Phone: +86-010-86469058
 E-mail: Contactus@LyncMed.com

 **LyncMed S.L.**

ADD: Carrer de les Filadores, 2, 08918 Badalona, Barcelona
 Phone: 009615215106391
 E-mail: Contactus@LyncMed.com

 **LyncMed GmbH**

ADD: Charlotten Str. 61, 51149, Köln, Nordrhein-Westfalen, Germany.
 Phone: +86-010-86469058
 E-mail: Contactus@LyncMed.com

 **LyncMed Inc.**

ADD: 2276 Pickwick Pl, Fullerton CA 92833.
 Phone: +86-010-86469058
 E-mail: Contactus@LyncMed.com

 **LyncMed & Akso Technology Co (UK) Ltd**

ADD: 19 Garrick Street London
 Contact: Michael Hyams
 Phone: +44 202 077 1402
 E-mail: MikeHyams@LyncMed.com

 **LyncMed Technology International Ltd.**

ADD: Room 06, 13A/F, South Tower, World Finance Center, Harbour City, 17 Canton Road, Tsim Sha Tsui, Kowloon, HongKong.
 Phone: +86-010-86469058
 E-mail: Contactus@LyncMed.com

 **LyncMed Technology S.R.L.**

ADD: Italy Milano (MI) Via Procaccini Giulio Cesare 32 Cap 20154
 Phone: +39 3778578323
 E-mail: Italy@lyncmed.com

 **LYNCMEDICAL TECHNOLOGY MÉXICO, S. DE R.L. DE C.V.**

ADD: Wework 707 varsovia 36.DF México
 Phone: +86-010-86469058
 E-mail: Contactus@LyncMed.com

 **LyncMed Technology MediKal Ltd Şti**

ADD: MOLLA GURANI MAHALLES TURGUT OZAL MILLET CAD. KARAGÜL İŞ MERKEZİ Apt. NO: 84/123 FATİH/İSTANBUL
 Phone: +86-010-86469058
 E-mail: Contactus@lyncmed.com

In House Production

Personalized Service and 24/7 Online Tracking System



Personalized Service and 24/7 Online Tracking System

Increased Efficiency

Raw Material Supply Chain Management & Control



Raw Material Supply Chain / Management & Control

Short Leadtime

Overseas Distribution Centers



MARITIME, SHIPPING, RAIL TRANSPORT, TRUCK TRANSPORT

Supply, logistics, post-storage, customs declaration, customs clearance, sea, transportation, last mile delivery, settlement, complete online transactions, and localized European landing teams, with offices, sales teams, and Customer service, warehouse, inventory, give you a convenient and fast experience

Risk Control & Management

SGS Inspection and Quality Control



Our factory



production line



MODEL: Sample -Nitrile gloves

PRODUCT INFORMATION	
Material	100% Nitrile
Color	Blue, Violet
Cuff length	Standard
Glove length(mm/inches) min	240mm\9.4inch
Powder content	≤2 mg per glove
External glove surface	Textured Finger
Freedom from holes\Inspection level 1)	1.5 AQL
Tensile strength	≥ 6.0 N
Application Temperature	<40°C

Specification	Size	Length(mm)	Width(mm)
LM-NG-01	S	>240	≤ 80
LM-NG-02	M	>240	95±10
LM-NG-03	L	>240	110±10
LM-NG-04	XL	>240	≥ 110

Medical examination nitrile gloves

Medical examination nitrile gloves are intended to be worn on the hands of healthcare personnel to prevent contamination between patients and examiners. This is a single-use, powderfree, non-sterile device.

FEATURES

1. Excellent mechanical strength provides a high level of hand protection.
2. Textured in finger tips for a secure grip.
3. Skin irritation&sensitization tested.

MANUFACTURING ACCREDITATIONS

ISO 13485 Medical Device Quality Management System

REGULATORY COMPLIANCE

MDR 2017/745
EN455

STORAGE INSTRUCTIONS



Package Components

Exterior Package Design

Inner box size: 220 × 120 × 63 mm

100PCS

Carton size: 330 × 230 × 220 mm

1000PCS

MOQ: 1*40HQ : 4000-4200 cartons

Delivery Time: 10-12 weeks upon the date of confirmed the final artwork and prepayment

Size Ratio: S+M>60%. If L+XL >60%, price will be adjusted.

Payment Term: 50% prepayment, 50% payment against copy of BL.





Test Report

No.: QDHL1806013408MD

Date: JUL.09,2018

Page 1 of 3

LYNCMED MEDICAL TECHNOLOGY (BEIJING) CO.,LTD.
NO.1111, SOUTH HUIHE ROAD, CHAOYANG DISTRICT, BEIJING, CHINA 100020

The following samples were submitted and identified by/on behalf of the client as:

Name of Product/Item : NITRILE EXAMINATION GLOVES
 SGS Ref. No. : SL218022416072TX
 Size : M
 Country of Origin : CHINA
 Test Performed : Selected test(s) as requested by applicant against:EN 420:2003+A1:2009
 Sample Receiving Date : JUN.21,2018
 Testing Period : JUN.21,2018 TO JUL.09,2018
 Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested. For further details, please refer to the following page(s).

Summary of test results:

Chemical tests

Test item	Result
pH Value	Pass

EN 420:2003+A1:2009

Test item	Result	Level of performance
5.1 & 6.1 Sizing	8	-
5.2 & 6.2 Dexterity (mm)	5	5

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Approved by:

Zhou Xinkuan, SK
Lab Manager

Attribution: To check the authenticity of testing inspection reports & certificates, please contact us at telephone: (86-755)93971143, or email: CN-Doccheck@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/ETD/Terms-and-Conditions/Terms-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

QDHL 001262

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Hardies

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101 | (86-532) 86999187 | (86-532) 80991952 | www.sgs.com.cn
 中国·青岛·莱山区株洲路143号通标中心 | 邮编: 266101 | (86-532) 86999187 | (86-532) 80991952 | sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report

No.: QDHL1806013408MD

Date: JUL.09,2018

Page 2 of 3

Test Result

pH Value
EN 1413:1998

Test item

pH value

7.5

Conclusion

Pass

Note:

- Requirement of EN 420:2003+A1:2009: greater than 3.5 and less than 9.5

Clause 5.2 Dexterity

Specimen	Smallest diameter of pin fulfilling test conditions (mm)	Level of performance (#)
1	5	5
2	5	
3	5	
4	5	

Remark: # = The performance level is defined as the smallest diameter of pin that can be picked up.

Level of performance	Smallest diameter of pin fulfilling test condition (mm)
1	11
2	9.5
3	8
4	6.5
5	5

Clause 5.1.2 & 6.1 Sizing

Claimed size	Length of glove (mm)	Glove size
M	242	8

Note: Sizes of hands (reference only)

Hand size	Hand circumference (20 mm from the crotch between thumb and index finger)	Hand length (mm)
6	152	160
7	178	171
8	203	182
9	229	192
10	254	204
11	279	215

Attention: To check the authenticity of testing inspection report & certificate, please contact us at telephone: (86-75)50071443, or email: CH.Dochy@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions_3103 and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

SGS-CS (Shanghai) Technical Service Co., Ltd.
Harding

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101

I (86-532) 68899187

QDHL

001263

I (86-532) 80991952

www.sgs.com.cn

中国·青岛·崂山区株洲路143号通标中心

邮编: 266101

I (86-532) 68899187

I (86-532) 80991952

e sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report

No.: QDHL1806013408MD

Date: JUL.09,2018

Page 3 of 3

Sample Description : One sample of nitrile examination glove in blue

Sample Photo:



SGS authenticate the photo on original report only

*** End of Report ***

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755)85071443, or email: CN.Certificate@sgs.com





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Forms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Forms-and-Conditions/Forms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

QDHL 001254

SGS (China) Technical Services Co., Ltd.
Incorporated in the People's Republic of China

SGS Center, No. 143, Zhushou Road, Laoshan District, Qingdao, China 266101 | (86-532) 68999187 | (86-532) 80991952 | www.sgsgroup.com.cn
中国·青岛·崂山区株洲路143号总部中心 邮编: 266101 | (86-532) 68999187 | (86-532) 80991952 | sgs.china@sgs.com

Member of the SGS Group (SGS SA)



Test Report

Report No.: QDHL2009009945MD

Sample Description: NITRILE EXAMINATION GLOVES
 LYNCMED MEDICAL TECHNOLOGY
 (BEIJING) CO., LTD.

Test Type: SUBMITTED BY CLIENT

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Page 1 of 6






Report No.: QDHL2009009945MD

Test Report

Sample Information	Sample Description	NITRILE EXAMINATION GLOVES	Color	BLUE
	Received sample quantity/ Tested sample quantity	220PCS/ 200PCS	Type/Specifications	EXAMINATION/ L
Lot No.	LMNE 20200816	Lot Quantity	NOT PROVIDED	
Manufacture Date	NOT PROVIDED	Expiration Date	NOT PROVIDED	
Material Appearance	NITRILE	Storage Condition	NOT PROVIDED	
Manufacturer	NOT PROVIDED			
Client Information	Applicant	LYNCMED MEDICAL TECHNOLOGY (BEIJING) CO., LTD.		
	Applicant address	ROOM 1601, BUILDING NO.2,ZHUBANG 2000 BUSINESS BUILDING, BALIZHUANG XIL159, CHAOYANG DISTRICT, 100022 BEIJING, CHINA		

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Page 2 of 6


Report No.: QDHL2009009945MD

Test Information	Sample Receiving Date	SEP.25.2020	Test Period Date	SEP.25.2020 TO OCT.16.2020
	Sample No.	QDHL2009009945MD	Test environment	Meat requirement
	Test items	Water tightness test		
Testing Accordance	EN 455-1:2020 Medical Gloves for Single Use – Part 1: Requirements and Testing for Freedom from Holes Clause 5.1			
Test conclusion	This report only provides the test results and individual judgment, conclusion please see follow pages. Issue date: OCT.16.2020			
Remark	/			

Approver: *Jesson Guo* Auditor: *Jesson Guo* Compiler: *Lillian Diao*

Date: OCT.16.2020 Date: OCT.16.2020 Date: OCT.16.2020

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Page 3 of 6





Report No.: QDHL2009009945MD

Sample Photo



SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Page 4 of 6



Report No.: QDHL2009009945MD

Test Results

Test Items	Unit	Test Method	Requirement	Test Result	Assessment
Water tightness test	/	EN455-1: 2020	Sample quantity: 200 AQL: 1.5 Ac: 7 Re: 5	Found: 0	Pass

End of Report

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Page 5 of 6





statement

- The report is considered invalidated in one or more of the following conditions: no approval signature; no testing seal of SGS; altered; a copy without the red testing seal of SGS.
- Above information and sample(s) was/were submitted and certified by the client. SGS quoted the information with no responsibility as to the accuracy, adequacy and/or completeness.
- Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.
- The test report cannot be reproduced in any way, except in full content, without prior approval in writing by the laboratory.
- Should you have any queries or objection to the test report, please contact us within 15 days after receiving the report.

Address: SGS Center, No.143, Zhuzhou Road, Laoshan District, Qingdao, China.
 Tel: 0532-6899187 Zip: 266101 Fax: 0532-80991952
 E-mail: Emily.Zhang@sgs.com

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Page 6 of 6




SGS


Test Report No.: QDHL1810022710MD Date: OCT.30.2018 Page: 1 of 5

LYNCMED MEDICAL TECHNICAL(BEIJING)CO., LTD
ROOM 119, NO.1111, SOUTH HUJIE ROAD, CHAOYANG DISTRICT, BEIJING

The following sample(s) was/were submitted and identified by the client as:
Sample Description : NITRILE GLOVE
SGS Ref. No. : SHHL1810057370MD
Sample Receiving Date : OCT.16.2018
Testing Period : OCT.16.2018 TO OCT.30.2018
Test Performed : SELECTED TEST(S) AS REQUESTED BY APPLICANT
Test Requested : EN 455-2:2015 MEDICAL GLOVES FOR SINGLE USE – PART 2: REQUIREMENTS AND TESTING FOR PHYSICAL PROPERTIES

Test Result(s) : PLEASE REFER TO THE FOLLOWING PAGE(S)
Conclusion : THE SUBMITTED SAMPLE MET THE TEST REQUIREMENT.

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.
Approved by:

Zhou Xinkuan, SK
Lab Manager



SGS


Test Report No.: QDHL1810022710MD Date: OCT.30.2018 Page: 2 of 5

Test Conducted:
EN 455-2:2015 Medical gloves for single use – part 2: Requirements and testing for physical properties

Number of test sample	: 39 Pieces
The type of gloves	: examination/procedure gloves b)
Manufacturing batch code	: /
Size	: Surgical gloves: / Examination/procedure gloves: M
Defects observed before testing	: No defects

Clause	Test Items	Result	Note
4	Dimensions	—	—
4.2	Length	Pass	# 1
4.3	Width	Pass	# 1
5	Strength	—	—
5.2	Force at break	Pass	# 2
5.3	Force at break after challenge testing	Pass	# 2

Notes : #1 See result 1
#2 See result 2



SGS

Test Report No.: QDHL1810022710MD Date: OCT.30.2018 Page: 3 of 5

Test Result:
1. Dimensions

Sample Quantity: 13pcs

Size	M											
Length(mm)	242	243	231	244	245	240	239	243	243	240	248	238
Width(mm)	96	95	96	95	95	95	95	96	98	95	95	96

Median value: Length (mm): 243
Width (mm): 95


Requirements: see table 1&2

Table 1 Dimensions for surgical gloves

Size	Median length in mm	Median width in mm
5	≥250	67 ± 4
5.5	≥250	72 ± 4
6	≥260	77 ± 5
6.5	≥260	83 ± 5
7	≥270	89 ± 5
7.5	≥270	95 ± 5
8	≥270	102 ± 6
8.5	≥280	108 ± 6
9	≥280	114 ± 6
9.5	≥280	121 ± 6

Table 2 Dimensions for examination/procedure gloves

Size	Median length in mm	Median width in mm
Extra small	—	≤80
Small	—	80 ± 10
Medium	≥240	95 ± 10
Large	—	110 ± 10
Extra Large	—	≥110



SGS

Test Report No.: QDHL1810022710MD Date: OCT.30.2018 Page: 4 of 5

2. Strength

Sample Quantity: 26pcs

Size	M												
Force at break(N)	5.0	5.0	6.6	6.6	5.8	5.2	6.6	6.2	5.2	6.1	6.6	6.4	6.9
Force at break after challenge testing(N)	7.8	8.2	6.0	8.3	9.7	6.9	8.6	6.3	8.5	8.0	8.5	7.3	8.3


Median value:
Force at break during shelf life (N): 6.2
Force at break after challenge testing (N): 8.2

Requirements: see table 3

Table 3 – Median values of force at break

	Force at break in Newton		
	Surgical gloves (c)	Examination/procedure gloves (d)	(e)
Throughout shelf life tested according to 5.2 and within 12 months of manufacture tested according to 5.3	≥ 6.0	≥ 6.0	≥ 3.0

1) Requirements for all surgical gloves.
2) Requirements for all examination gloves except gloves made from thermoplastic elastomer (e.g. polyurethane polyurethane).
3) Requirements for gloves made from thermoplastic elastomer (e.g. polyurethane polyurethane).



SGS

Test Report No.: QDHL1810022710MD Date: OCT.30.2018 Page: 5 of 5

Sample Photo:

Received sample



SGS authenticate the photo on original report only
End of Report



SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 1 of 10

LYNCMED MEDICAL TECHNOLOGY (BEIJING) CO., LTD
ROOM 119, FLOOR 1, GUOTIUSHANGKE BUILDING NO. 1111, SOUTH HUIHE ROAD
CHAOYANG DISTRICT, CHINA

THE TEST REPORT IS TO SUPERSEDE THE TEST REPORT No.: SHHL1811067691SD-01, DATE: MAR. 29, 2019. ORIGINAL REPORT SHALL BE INVALID. THE TEST REPORT UPDATED PROTOCOL.

Sample Description : NITRILE GLOVE
Style No. : M

Sample Receiving Date : NOV. 27, 2018
Testing Period : NOV. 27, 2018 TO APR. 02, 2019
Test Performed : (SELECTED TEST(S) AS REQUESTED BY APPLICANT
(SELECTED TEST(S) AS REQUESTED BY APPLICANT
Test Requested : FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING PAGE(S)
Test Result(s) : FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING PAGE(S)
Conclusion : FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING PAGE(S)

Signed for and on behalf of
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd.

Minky Zheng
Authorized Signatory

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 2 of 10

Test Conducted:

Clause	Test Category	TEST RESULT
4.1	General – sanitization (Maximization method)	PASS (SEE RESULT PAGE)
4.3	Endotoxins	PASS (SEE RESULT PAGE)
4.4	Powder	PASS (SEE RESULT PAGE)
4.5	Protein, Total leachable	NA
4.6	Labeling	NC

REMARK: NA = Not Applicable NC = Not conducted as per client's request

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 3 of 10

Test Result Page:

Attachment 1: Test for Irritation (Animal skin irritation test)

SUMMARY

The animal skin irritation test of the test article, Nitrile Glove, was conducted to assess the potential of the material to produce irritation. The study was conducted based on the requirements of the International Organization for Standardization ISO 10993-10: 2010 Biological Evaluation of Medical Devices - Part 10: Tests for Irritation and Skin Sensitization, ISO 10993-12: 2012: Biological evaluation of medical devices Part 12: Sample preparation and reference materials.

The test article was extracted in 0.9% sodium chloride injection (SC) and cotton seed oil (CSO). Each extract and corresponding reagent control was contacted on animal skin directly. Observations for erythema and edema were conducted at 24, 48 and 72 hours after contact.

Under the conditions of this study, there was no evidence of significant irritation from the test article to rabbits. The response category for the extracts of the test article was **negligible**.

MATERIALS

The test article provided by the sponsor was identified and handled as follows:

Test Article: Nitrile Glove
Sanitization Status: Non-sterile
Storage Conditions: Room temperature
Extraction Vehicle: 0.9% sodium chloride injection (SC)
Cotton seed oil (CSO)

Test Article Preparation: According to the requirement of the sponsor, the test articles were sterilized by ethylene oxide two weeks before the treatment.
Based on the ISO 10993-12:2012, the ratio of 6cm²:1 ml (Surface area of the test sample to volume of extraction

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 4 of 10

vehicle, 950ml of the test article were covered with 150ml extraction vehicle under aseptic conditions for preparing the SC and CSO test extract at 37 °C for 72h respectively. The extracts were used after extraction. The extraction vehicles (without test article) were similarly prepared to serve as the reagent control. All the extract of the test and controls were clear.

Reagent Control:

Condition of extracts:

In addition, according ISO 10993-10 requirement, 10% Sodium Dodecyl Sulfate as a positive control was used previously for another study (2018.12.24–2018.12.28). Complete data is traceable in laboratory records.

METHODS

Test System

Species: Rabbit
Strain: New Zealand White
Source: SHANGHAI SONGLIAN LAB ANIMAL-FIELD
Sex: Male
Body weight range: 2.4 kg – 2.5 kg
Age: Young adult
Number of animals: Six

Animal Management

Husbandry: Conditions conformed to "Laboratory animal Requirements of environment and housing facilities". Diet was provided from Shanghai Pu Lu Teng Biological Technology Co., Ltd.
Foot: Healthy animals were acclimatized to the laboratory conditions for 7 days before the treatment, and then they were individually housed in stainless steel

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 5 of 10

suspended cages identified by a card indicating the Identification No of the test article and test treatment date.

Environmental: The room temperature and humidity was monitored daily. The room temperature range was from 20°C to 26°C. The room humidity range was from 50% to 70%.

Personnel: Associates involved were appropriately qualified and trained.

Selection: Only healthy, previously unused rabbits were selected.

Experimental Procedure:

On the day before the test, the rabbits were closely clipped the fur on the backs of the animals, and both sides of the spinal for application and observation of all test sites, approximately 10 cm x15 cm, A 25 mmx25 mm section of absorbent gauze patch was saturated with freshly prepared the extract, and then was applied to the test sites. The extract of test article and the reagent control were directly applied to the region as illustrated below:

The application sites were covered with a gauze patch and then the application sites were wrapped with a semiocclusive bandage for 24 h. At the end of the contact time, the dressings were removed. A natural lighting was used to visualize the skin reactions. The skin reactions for erythema and edema were described and scored at 1, 24, 48 and 72 hours. The tissue reaction for erythema and edema were graded according to the classification system given below for each site and at each time observed, and the results were recorded.

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 6 of 10

Reaction	Primary Irritation Score
Erythema and eschar formation	
No erythema	0
Very slight erythema (barely perceptible)	1
Well-defined erythema	2
Moderate erythema	3
Severe erythema (well-redness) to eschar formation preventing grading of erythema	4
Oedema formation	
No oedema	0
Very slight oedema (barely perceptible)	1
Well-defined oedema (edges of area well-defined by definite raising)	2
Moderate oedema (raised approximately 1 mm)	3
Severe oedema (raised more than 1mm and extending beyond exposure area)	4

Only the 24, 48 and 72hours observations were used for calculation. For each animal, the score both erythema and oedema at each time point were added together separately for each test article and the negative control. The primary irritation score for an animal was calculated by dividing the sum of all the scores by 6 (2 test sites x 3 time points). All the primary irritation scores of individual animals were added and divided by the number of animals, and then the primary irritation scores for each test article were obtained. A similar calculation was made with the negative control. The primary irritation index was obtained by subtracting the score of the negative control from the test article score and the response categories were given as below:

Mean score	Response category
0 to 0.4	Negligible
0.5 to 1.8	Slight
2 to 4.9	Moderate
5 to 8	Severe

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 7 of 10

RESULTS

All animals appeared clinically normal throughout the study. All sites of the test extract and the reagent control appeared normal following removal of the patches; the score of the test extract and the reagent control all were 0.

The Primary Irritation Index (PII) of the test article was 0.0.

CONCLUSION

Under the conditions of this study, there was no evidence of significant irritation from the test article to rabbits. The response category for the extracts of the test article was **negligible**.

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 8 of 10

PHOTOGRAPH OF THE TEST ARTICLE

SGS

Test Report No.: SHHL1811067691SD-01 Date: APR. 02, 2019 Page: 9 of 10

The following test was performed by SGS other internal laboratory

BS EN 455-3:2015 Medical gloves for single use—Part 3: Requirements and testing for biological evaluation

Number of test sample	6 Pieces
Frapsies of gloves	Prepared/Free gloves, other than surgeon's gloves
Defects observed before testing	No defects
Test Result	Pass

Clause	Test Name	Result	Note
4.4	Powder-free gloves	Pass	#1

Notes : #1 Test according to EN ISO 21171:2006, the average mass of powder per glove was 0.00g.

Remark:

- Since the data and / or information above division line of front page is provided by the applicant, the relevant results or conclusions of this report are only made for these data and / or information. SGS is not responsible for the authenticity, integrity and results of the data and information and / or the validity of the conclusion. Testing results only apply to the sample as received.
- The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.

Certificates – EN374 Test Report

SGS TEST REPORT

Report No.: CH-TX-1042003821 DATE: 25/01/2019
 1042003821
 LYNCEMED MEDICAL TECHNOLOGY (BEIJING) CO., LTD
 ROOM 119, FLOOR 1, GUOTOU SHANGKE BUILDING NO 111, SOUTH HUIHE ROAD
 CHINA
 AIC F919391 SGS-STD STANDARDS TECHNICAL SERVICES (SHANGHAI) CO., LTD.
 CONTACT PERSON:

THE FOLLOWING SAMPLE(S) WAS/WERE SUBMITTED AND IDENTIFIED BY/ON BEHALF OF THE CUSTOMER AS:
 SAMPLE DESCRIPTION: GLOVES
 STYLE NO.: NITRILE GLOVE
 PHOTO APPENDIX:

SAMPLE RECD ON: 22/01/2019 TESTING PERIOD: 22/01/2019 – 25/01/2019

Test Method / Standard	Clause/Test Name	Status / Performance Level
EN 374-2014	Protective gloves against chemicals and micro-organisms: Determination of resistance penetration	Pass
EN 1623-1-2015	Permeation by Liquid chemical under conditions of continuous contact	Level=0
EN 374-4-2013	Resistance to Degradation by Chemicals	Noted results.

Per pro SGS India Private Ltd,
 K. PACHAYAPPAN
 ASST. MANAGER
 Email your Test Report Related Enquiries at Feedback_Sl_T@sgs.com

SGS TEST REPORT

Report No.: CH-TX-1042003821 DATE: 25/01/2019
 1042003821

RESULTS

EN 374-2: 2014 Protective gloves against chemicals and micro-organisms – Part 2: Determination of resistance penetration

Clause	Test Name	Test Results	Performance level	
4.1	Air leak Test (Air Pressure Used: 0.5 kPa)	Specimen # Size M Size M Size M Size M	Leakage No Leakage No Leakage No Leakage No Leakage	Pass
4.2	Water leak test	Specimen # Size M Size M Size M Size M	Leakage No Leakage No Leakage No Leakage No Leakage	Pass

EN 1623-1-2015 Determination of material resistance to permeation by chemicals – Part 1: Permeation by liquid chemical under conditions of continuous contact

Chemical CAS NO	Loop syringe/injection medium	Analytical technique used	Mean thickness (mm)	NBT at NPR L10 µg cm ⁻² min ⁻¹ (minutes)	Performance level according to EN ISO 374-1: 2016 Table 1	Observation
Methanol 67-66-1	Open Nipox/ Nipogen	Continuous measurement With GC/ED	0.08 0.08 0.08	<1 <1 <1	Level=0	Severe swelling

EN ISO 374-4: 2013 – Protective gloves against dangerous chemicals and micro-organisms. Part 1: Terminology and performance requirements for chemical risks. Table 1: Permeation performance levels.

Permeation performance level	Measured breakthrough time (minutes)
0	>10
1	>30
2	>60
3	>120
4	>240
5	>480

Performance levels are based on the lowest individual results achieved per chemical

SGS TEST REPORT

Report No.: CH-TX-1042003821 DATE: 25/01/2019
 1042003821

RESULTS

EN 374-4: 2013 Protective Gloves against Chemicals and Micro Organisms – Determination of resistance to degradation by chemicals

Chemical / CAS NO	Exposure Duration	Percentage change in puncture resistance	Test Results (Glove sample Result (%))	Observation
Methanol 67-66-1	60/65 minutes	38.0	1	Severe swelling
		45.7	2	
		45.2	3	
		Mean	44.3	
		Standard Deviation	5.645	

Certificates – EN1186 Test Report

SGS Test Report

No. SHAHG1628661102 A01 Date: 07 Dec 2018 Page 1 of 4

LYNCEMED MEDICAL TECHNOLOGY (BEIJING) CO., LTD
 ROOM 119, FLOOR 1, GUOTOU SHANGKE BUILDING NO 111, SOUTH HUIHE ROAD CHAOYANG DISTRICT, CHINA.

THIS REPORT IS TO SUPPLEMENT TEST REPORT NO. SHAHG1628661101, DATE 20181206. The following sample(s) was/were submitted and identified on behalf of the client as: NITRILE GLOVE

SGS JOB NO.: SHH1610095400 - SH
 Style No.: S.M.L.XL
 Date of Sample Received: 29 Nov 2018
 Testing Period: 29 Nov 2018 - 08 Dec 2018
 Test Requested: Selected tests as requested by client.
 Test Method: Please refer to next page(s).
 Test Results: Please refer to next page(s).

Result Summary:

Test Requested	Conclusion
French Annex A3 9 November 1994, French Decree 2007-766 with amendments, DGCCRF Methodological Sheet "Fib organic materials - synthetic rubber" - Chemical	PASS
French Annex A3 9 November 1994, French Decree 2007-766 with amendments, DGCCRF Methodological Sheet "Fib organic materials - synthetic rubber" - Chemical	PASS
French Annex A3 9 November 1994, French Decree 2007-766 with amendments, DGCCRF Methodological Sheet "Fib organic materials - synthetic rubber" - Volatile organic matter (VOM)	PASS

Signed for and on behalf of: LyncMed Medical Technology Services (Shanghai) Co., Ltd.
 Approved Signatory:

SGS Test Report

No. SHAHG1628661102 A01 Date: 07 Dec 2018 Page 2 of 4

Test Results:

Specimen No.	SGS Sample ID	Description	Material (determined by the client)	Nitrile rubber
001	SHH16-205611-001	Blue rubber glove (XL)		

Remarks:

- (1) mg/kg – milligram per kilogram
- (2) mg/kg – milligram per square decimeter
- (3) °C – degree Celsius
- (4) % – less than
- (5) NDC – Method Deviation Limit
- (6) N/A – Not Detected / NCL

French Annex A3 9 November 1994, French Decree 2007-766 with amendments, DGCCRF Methodological Sheet "Fib organic materials - synthetic rubber - chemical analysis"

Test Method: With reference to Commission Regulation (EU) No 102011 of 14 January 2011 Annex II and EN 1186-1:2002 for selection of test methods: EN 1186-9:2002 aqueous food simulants by article 8.8.8.8. method: EN 1186-14:2002 - solution test

Simulant/Load	Time	Temperature	Max. Permissible LEI	Result of 1001	Conclusion
10% Ethanol (V/V) aqueous solution	2.0hrs	47°C	10mg/kg	<3.0mg/kg	PASS
20% Ethanol (V/V) aqueous solution	2.0hrs	47°C	10mg/kg	<3.0mg/kg	PASS
50% Ethanol (V/V) aqueous solution	2.0hrs	47°C	10mg/kg	<3.0mg/kg	PASS
95% Ethanol	2.0hrs	47°C	10mg/kg	8.2mg/kg	PASS
Isocouane	0.5hrs	47°C	10mg/kg	<3.0mg/kg	PASS

Notes:

- (1) Analytical tolerance of aqueous simulants is 2 mg/kg or 12mg/kg
- Analytical tolerance of fatty food simulants is 3 mg/kg or 20mg/kg
- Test condition & simulant were specified by client.

French Annex A3 9 November 1994, French Decree 2007-766 with amendments, DGCCRF Methodological Sheet "Fib organic materials - synthetic rubber - chemical analysis"

SGS Test Report

No. SHAHG1628661102 A01 Date: 07 Dec 2018 Page 3 of 4

Test Method: With reference to European pharmacopoeia, 2005 Appendix X. F. Periodic Value Method A.

Test Result	Limit	Unit	MPD	OT
Peroxide Value	0.50	% (w/w)	0.10	0.12
Conclusion				PASS

Notes: (1) * = Absent.

French Annex A3 9 November 1994, French Decree 2007-766 with amendments, DGCCRF Methodological Sheet "Fib organic materials - synthetic rubber - chemical analysis"

SGS Test Report

No. SHAHG1628661102 A01 Date: 07 Dec 2018 Page 4 of 4

Sample photo:

SGS authenticates the photo on original report only
 *** End of Report ***

TÜV SÜD
 ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFICAT



Certificate

No. Q6 099730 0004 Rev. 02

Holder of Certificate: **Lyncmed Medical Technology (Beijing) Co., Ltd.**
 Room 1601, Building No. 2
 Zhubang 2000 Business Building, Balizhuang Xili 99
 Chaoyang District
 100022 Beijing
 PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate: **Distribution, Production and Sales of Dental high speed air turbine handpieces, Dental low speed air turbine handpieces, Ultrasonic Scaler, LED Curing Lights, Root Apex locators, Endo motors, Pulp tester, Portable Dental Unit, Dental Implant Systems, Dental Alginate Mixers, Sterile Surgical Gowns, Sterile Surgical Packs, Sterile Surgical Drapes, Sterile Dressing Pouches, Caps, Face mask, Shoe cover, Bed cover, Protective coverall, Urine bag, Foley Catheter, Oxygen Mask, Nebulizer Mask, Laryngeal Mask, Anaesthetic Mask, Wheelchair, Walker, Crutch.**

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system (excluding subclause 7.3), which meets the requirements of the listed standard(s). All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:Q6_099730_0004_Rev_02

Report No.: BJ20020501
Valid from: 2020-12-20
Valid until: 2023-12-19

Date, 2020-12-11

Christoph Dicks
 Head of Certification/Notified Body

TÜV SÜD
ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT



Certificate

No. Q6 099730 0004 Rev. 02

Applied Standard(s): EN ISO 13485:2016
Medical devices - Quality management systems -
Requirements for regulatory purposes
(ISO 13485:2016)
DIN EN ISO 13485:2016

Facility(ies): Lyncmed Medical Technology (Beijing) Co., Ltd.
Room 1601, Building No. 2, Zhubang 2000 Business Building,
Balizhuang Xili 99, Chaoyang District, 100022 Beijing, PEOPLE'S
REPUBLIC OF CHINA

See scope of certificate



TÜV Rheinland
LGA Products GmbH
Tillystraße 2, 90431 Nürnberg

Doc. 1/1, Rev. 0

**Attachment to
Certificate**

Registration No.: SX 60148110 0001
Report No.: 50180180 006

Organization: Lyncmed Medical Technology
(Beijing) Co., Ltd.
Room 1601, Building No.2
Zhubang 2000 Business Building, Balizhuangxili
Chaoyang District
100022 Beijing
P.R. China

Scope:

Products:

Sterile Latex Surgical Gloves, Non-sterile Examination Latex Gloves, Dental High Speed Air Turbine Hand-pieces, Dental Low Speed Air Turbine Hand-pieces, Patient Examination Gloves, Ultrasonic Scaler, Endo Motors, Apex Locators, Pulp Testers, Masks, Laryngeal Mask Airways, Overalls, Isolation Gowns, Sterile Surgical Gowns, Sterile Surgical Kits, Sterile Surgical Caps, Sterile Face Masks, Sterile Shoe Covers, Alcohol Swabs, Diapers and Adult Diapers, Disposable Under Pads, Medical Tapes, Urine Bags for Single Use, Foley Catheter Silicone, Dental Alginate Mixers, Vinyl Examination Gloves and Sterile Gauze Sponges

Certification Body



Date: 2020-06-22

Jing Zhang



Certificate

The Certification Body of
TÜV Rheinland LGA Products GmbH

hereby certifies that the organization
**Lyncmed Medical Technology
(Beijing) Co., Ltd.**
Room 1601, Building No.2
Zhubang 2000 Business Building, Balizhuangxili 99
Chaoyang District
100022 Beijing
P.R. China

has established and applies a quality management system for medical devices
for the following scope:

Manufacture and Distribution of Medical Devices

(See attachment for Products included)

Proof has been furnished that the requirements specified in

EN ISO 13485:2016

are fulfilled. The quality management system is subject to yearly surveillance.

Effective Date: 2020-06-22
Certificate Registration No.: SX 60148110 0001
An audit was performed Report No.: 50180180 006
This Certificate is valid until: 2021-12-03

Certification Body



Date 2020-06-22



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
Tel.: +49 221 808-1371 Fax: +49 221 808-3935 e-mail:cert-validity@de.tuv.com http://www.tuv.com/safety

Government Approved Qualified Manufacturer Name List

取得国外标准认证或注册的医疗物资生产企业清单

Name List of Medical Devices and Supplies Companies with Certification/Authorization from other Countries

<http://www.ccmhpie.org.cn/>

序	生产企业	统一社会信用代码	国外注册认证情况
No.	Company	Uniform Social Credit Code	Status of Certification / Authorization in Other Countries
一、医用口罩 Medical Face Masks			
1	稳健医疗用品股份有限公司 Winner Medical Co., Ltd.	91440300723009295R	CE
2	湖北省潜江市江赫医用材料有限公司 Hubei Qianjiang Kingphar Medical Material Co.,Ltd	914290051836675600	CE
3	合肥美迪普医疗卫生用品有限公司 Hefei Medpro Healthcare Co., Ltd.	91340100744891942W	CE
4	海安美佳医用敷料有限公司 Haian Medigauze Co.,Ltd.	91320621733774564Q	CE
5	湖北康宁防护用品有限公司 Hubei Kangning Protective Products Co.,Ltd.	914290047959276261	CE
6	仙桃市兴荣防护用品有限公司 Xiantao Xingrong Protective Products Co., Ltd.	91429004753447771A	CE
7	奥美医疗用品股份有限公司 Allmed Medical Products Co.,Ltd.	9142058373914001XH	CE
8	金士达医疗(咸宁)有限公司 Kingstar Medical (Xianning) Co.,Ltd.	91421200080935876P	CE
9	江苏省健尔康医用敷料有限公司 Jiangsu Province Jianerkang Medical Dressing Co., Ltd	91320413714946201R	CE
10	安徽天康医疗科技股份有限公司 Anhui Tiankang Medical Technology Co.,Ltd	9134110071990801XY	CE
11	苏州恒祥进出口有限公司 Suzhou Hengxiang Import &Export Co.,Ltd.	91320508724448775B	CE
12	河南康灸来医疗科技有限公司 Henan Kanjoray Medical Technology Co., Ltd.	91410600MA46CYC12X	CE
13	联医医疗科技(北京)有限公司 LyncMed Medical Technology (Beijing) Co., Ltd.	91110108MA002T3N4R	CE
14	威海威高医疗国际贸易有限公司 Weihai Medical Intemational Co., Ltd.	9137100005624731X3	CE



LyncMed Group Corporation

📍 Room 1601, Building No.2, Zhubang 2000 Business Building, Balizhuangxili 99, Chaoyang District, 100022 Beijing, China.

☎ +86 010 8646 8058

✉ contactus@lyncmed.com

🌐 www.lyncmed.com

LOVES