

## Certificate of Analysis

**Description:** Immunoaffinity column for Deoxynivalenol

**Catalog No.:** BQ0008

**Lot No.:** IAC20250326

Test Item	Requirement	Result	
Appearance	Matrix: white, even distribution, no air bubble, no crevice	<input checked="" type="checkbox"/> Qualified	<input type="checkbox"/> Unqualified
	Protective solution: clear, no impurity	<input checked="" type="checkbox"/> Qualified	<input type="checkbox"/> Unqualified
Upper column plug	Stable, no crevice	<input checked="" type="checkbox"/> Qualified	<input type="checkbox"/> Unqualified
Bottom column cover	Stable, leakage-free, no crevice	<input checked="" type="checkbox"/> Qualified	<input type="checkbox"/> Unqualified
Column capacity	Maximum column capacity spike: 85%-110% (n=6)	<input checked="" type="checkbox"/> Qualified	<input type="checkbox"/> Unqualified
Spike test	85%-110% (n=3)	<input checked="" type="checkbox"/> Qualified	<input type="checkbox"/> Unqualified
Inner batch CV%	≤ 5% (n=6)	<input checked="" type="checkbox"/> Qualified	<input type="checkbox"/> Unqualified
Conclusion	<input checked="" type="checkbox"/> Qualified <input type="checkbox"/> Unqualified		

Approved:



Date: 2025-03-26