

**INDUS MEDICARE PRIVATE LIMITED****"SURYODAYA", 1-10-60/3, BEGUMPET, HYDERABAD – 500 016, TELANGANA, INDIA.****TEL : +91 40 27762603****FAX : +91 40 27761418****Q.A. LOT RELEASE TEST REPORT**

DATE : 27.08.2022

REPORT NO. : IMPL / 22 / 304

CUSTOMER : DITA ESTFARM LLC – MOLDOVA

LOT NO. : 22 06 001

QUANTITY : 1131.33 GROSS

MFD. : 2022 - 06

TYPE OF PRODUCT: NATURAL, MALE LATEX CONDOMS (THICK 53mm)

EXP. : 2027 - 05

BRAND NAME : ANY TIME 12'S

SPECIFICATION: EN ISO 4074 : 2015

| SL. NO | CRITERIA / TEST                                       | INSPECTION LEVEL                           | AQL                               | SPECIFICATIONS                                      | SAMPLE QTY. IN PIECES | NO. OF PIECES FAILED | MEAN                   | MINIMUM                | MAXIMUM                |
|--------|-------------------------------------------------------|--------------------------------------------|-----------------------------------|-----------------------------------------------------|-----------------------|----------------------|------------------------|------------------------|------------------------|
| 1.     | DIMENSIONS<br>LENGTH<br>NOMINAL WIDTH<br>THICKNESS    | 13 PIECES                                  | ACCEPT - 0<br>REJECT - 1          | L = min.190 mm<br>W = 53 ± 2 mm<br>T = Min 0.100 mm | 13                    | 0                    | 196.8<br>52.7<br>0.113 | 195.0<br>52.5<br>0.111 | 199.0<br>53.0<br>0.119 |
| 2.     | BURSTING VOLUME AND PRESSURE BEFORE OVEN CONDITIONING | G - 1                                      | 1.5<br>ACCEPT - 10<br>REJECT - 11 | B.V = min.18.0 ltrs.<br>B.P = min. 1.0 KPa.         | 315                   | 1                    | 45.9<br>3.45           | 17.0<br>1.45           | 60.0<br>4.05           |
| 3.     | FREEDOM FROM HOLES                                    | G - 1<br>(min. 315pieces)                  | 0.25<br>ACCEPT - 2<br>REJECT - 3  | AS PER AQL                                          | 315                   | 0                    | COMPLIES               |                        |                        |
| 4.     | VISIBLE DEFECTS                                       | G - 1<br>(min. 315pieces)                  | 0.4<br>ACCEPT - 3<br>REJECT - 4   | AS PER AQL                                          | 315                   | 0                    | COMPLIES               |                        |                        |
| 5.     | VISIBLY OPEN SEALS                                    | G - 1<br>(min. 315pieces)                  | 0.4<br>ACCEPT - 3<br>REJECT - 4   | AS PER AQL                                          | 315                   | 0                    | COMPLIES               |                        |                        |
| 6.     | PACKAGE INTEGRITY                                     | S - 3                                      | 2.5<br>ACCEPT - 2<br>REJECT - 3   | AS PER AQL                                          | 32                    | 0                    | COMPLIES               |                        |                        |
| 7.     | LUBRICANT PLUS POWDER                                 | 13 PIECES                                  | ACCEPT - 0<br>REJECT - 1          | Min.350 mg                                          | 13                    | 0                    | 428                    | 402                    | 465                    |
| 8.     | PACKAGING & LABELING                                  | 13 Individual packs +<br>13 Consumer packs | 0                                 | AS PER AQL                                          | 26                    | 0                    | COMPLIES               |                        |                        |

REMARKS: PASS

Venkateswarlu  
(S.VENKATESWARLU) 27/08/22  
FINAL LOT TESTING IN-CHARGE

27/08/2022  
(RAJA. R)  
Q.A. HEAD

BV – BURST VOLUME

BP – BURST PRESSURE

100 27.11