

EC CERTIFICATE

for the Quality Assurance System



according the Directive 93/42/EEC, Annex II excluding section (4)

As a Notified Body of the European Union, DEKRA Certification GmbH certifies, that the company
Richard Wolf GmbH

Pforzheimer Straße 32, 75438 Knittlingen, Germany

Certified location:

Pforzheimer Straße 32, 75438 Knittlingen, Germany

applies a quality assurance system according to the Directive 93/42/EEC Annex II for the medical devices listed in the annex. The approval is based on the result of the re-certification audit report no. 50593-Z7-00, the decision dated 2020-04-01 and is only valid in connection with the successful performance of the annual surveillance audits.

This certificate is valid from 2020-04-01 to 2024-05-26

Registration No.: 50593-16-05



Ruth Delbeck-Bayer
DEKRA Certification GmbH Stuttgart; 2020-04-01
Notified Body ID-number: 0124

DEKRA Certification GmbH * Handwerkstraße 15 * D-70565 Stuttgart * www.dekra-certification.de



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-295.10.02

Annex to the EC Certificate No. 50593-16-05

Valid from 2020-04-01 to 2024-05-26

Revision status of the annex: 0 dated 2020-04-01

Devices/device categories included in the certificate:

Class I s:

For the products listed below, review of the Quality Assurance System refers exclusively to aspects of manufacture concerned with securing and maintaining sterile conditions.

- Endoscopic suction valve, single-use, sterile
- Suction system filter, plume particulate
- Suction/irrigation tubing, single use

Class II a:

- Basic endotracheal tube, reusable
- Basic roller pump
- Bone cutting forceps
- Bone graft funnel
- Bronchoscopy tube
- Cannulated surgical drill bit, reusable
- Endoscope assembly adaptor
- Endoscope sheath, reusable
- Endoscopic electrosurgical handpiece/electrode, bipolar, reusable
- Endoscopic electrosurgical handpiece/electrode, monopolar, reusable
- Endoscopic insufflation tubing set, single-use
- Endoscopic insufflation tubing set, sterile, reusable
- Flexible fibreoptic cystourethroscope
- Flexible fibreoptic hysteroscope
- Flexible fibreoptic nasopharyngoscope
- Flexible fibreoptic ureterorenoscope
- Flexible video bronchoscope, reusable
- Flexible video cystoscope, reusable
- Flexible video ureterorenoscope, reusable
- Fluted surgical drill bit, reusable
- General-purpose endoscopic needle, reusable
- General-purpose endoscopic needle, single-use
- Haemorrhoid ligator
- High-pressure medical gas tubing
- Laparoscopic access cannula, reusable
- Laparoscopic multi-instrument access port, reusable
- Laparoscopic multi-instrument access port, single-use
- Laser fibre
- Line-powered surgical power tool system motor
- Medical air low pressure tubing
- Microbial medical gas filter, sterile, single-use
- Operating room audiovisual data/device management system application software
- Orthopaedic bur, reusable
- Orthopaedic bur, single-use
- Resectoscope
- Rigid bronchoscope
- Rigid cystourethroscope
- Rigid endoscope telescope
- Rigid endoscopic grasping forceps, reusable
- Rigid optical hysteroscope
- Rigid intubation laryngoscope, reusable

Annex to the EC Certificate No. 50593-16-05

Valid from 2020-04-01 to 2024-05-26

Revision status of the annex: 0 dated 2020-04-01

Devices/device categories included in the certificate:

- Rigid mediastinoscope
- Rigid nephroscope
- Rigid optical laparoscope
- Rigid ureterorenoscope
- Spinal needle, single-use
- Spring-loaded pneumoperitoneum needle, reusable
- Surgical drill guide, reusable
- Surgical fluid/smoke waste management system suction unit
- Surgical guillotine
- Surgical irrigation tubing set, reusable
- Surgical irrigation tubing set, single-use
- Surgical irrigation/aspiration handpiece, reusable
- Surgical irrigation/aspiration tubing set
- Surgical power tool system control unit, line-powered
- Tissue extraction bag
- Tissue morcellation system
- Tissue morcellation system handpiece, line-powered
- Uterine manipulator cervical cup/transilluminator
- Uterine manipulator, reusable
- Uterine probe

Class II b:

- Electrosurgical system generator
- Endoscopic electrosurgical electrode, bipolar, reusable
- Endoscopic electrosurgical electrode, bipolar, single-use, sterile
- Endoscopic electrosurgical handpiece/electrode, monopolar, reusable
- Endoscopic electrosurgical electrode, monopolar, single-use
- Endoscopic electrosurgical handpiece/electrode, bipolar, reusable
- Endoscopic electrosurgical handpiece/electrode, monopolar, reusable
- Endoscopic electrosurgical handpiece/electrode, monopolar, single-use
- General/multiple surgical diode Laser system
- Hysteroscopic irrigation/insufflation system
- Laparoscopic insufflator
- Laser lithotripsy system
- Operating room audiovisual data/device management system application software
- Piezoelectric lithotripsy system
- Soft-tissue/mesh anchor, non-bioabsorbable
- Ultrasonic lithotripsy system
- Electromechanical orthopaedic extracorporeal shock wave therapy system





Ruth Delbeck-Bayer
DEKRA Certification GmbH, Stuttgart, 2020-04-01
Notified Body ID-number: 0124

DEKRA Certification GmbH * Handwerkstraße 15 * D-70565 Stuttgart * www.dekra-certification.de

CERTIFICATE



EN ISO 13485:2016

DEKRA Certification GmbH hereby certifies that the organization

Richard Wolf GmbH

Scope of certification:

Design and development, production, distribution, installation and service of systems, active medical devices (sterile, non-sterile), non-active medical devices (sterile, non-sterile) for human medicine, in particular for endoscopy and extracorporeal shockwave application.
Design and development, production, and distribution of non-active implants in urology and surgery as well as accessories for processing (cleaning, disinfection, sterilization)

Certified location:

Pforzheimer Straße 32, 75438 Knittlingen, Germany

(further locations see annex)

has established and maintains a quality management system according to the above mentioned standard. The conformity was adduced with audit report no. 50593-R2-00.

Certificate registration no.:	50593-14-02	Certificate valid from:	2021-11-29
Validity of previous certificate:	2021-11-28	Certificate valid to:	2024-11-28



Ruth Delbeck-Bayer
DEKRA Certification GmbH, Stuttgart, 2021-11-29



Annex to the Certificate No. 50593-14-02

Revision status: 0

valid from 2021-11-29 to 2024-11-28

The following locations / companies belong to the certificate above:

	Headquarter	Certified location	Scope of certification
	Richard Wolf GmbH	Pforzheimer Straße 32 75438 Knittlingen Germany	see page 1
	at the following locations / at the companies at the following locations		Scope of certification
1.	Richard Wolf GmbH	Reuchlinstraße 10-11 10553 Berlin Germany	Manufacture of flexible and rigid endoscopes



Ruth Delbeck-Bayer
DEKRA Certification GmbH, Stuttgart, 2021-11-29

OptiFLOW™

HIGHFLOW CONTINUOUS IRRIGATION SURGICAL SHEATH



5.5 MM

FOR TELESCOPE 2.7 MM, 12° AND 30°

SHEATHS

3



	Sheath	Working channel	WL	Product no.	
 Sheath for hysteroscope 16.5 FR distal sheath end, oval	12 Fr	5 Fr	219 mm	8974351	
Handle for hysteroscope				8986.251	
also:					
	Direction of view	Ø	SL	TL	Product no.
Telescope 12° Ø 2.7 mm SL 310 mm	12°	2.7 mm	310 mm	373 mm	8974.412
Telescope 30° Ø 2.7 mm SL 310 mm	30°	2.7 mm	310 mm	370 mm	8974.402

Recommended accessories	Product no.
Perforated basket for standard telescope 2	Dimensions (L x W x H): 481 x 59 x 54 mm 38016
Fiber light cable bndl consisting of: Fiber light cable (80662523), projector-side adapter (8095.07), and endoscope-side adapter (809509)	Ø 2.5 mm, 2.3 m long 806625231

||23|| Subject to modifications. Please note that some products may not be available in all countries. Please always take note of the information that is provided with the product. For more detailed information, please consult with your contact partner or directly with Richard Wolf GmbH.