



# Technical Data

## Soyabean Casein Digest Agar (Tryptone Soya Agar) (Casein Soyabean Digest Agar)

M290

### Intended use

Soyabean Casein Digest Agar is a general purpose medium used for cultivation of a wide variety of microorganisms from clinical and non-clinical samples and for sterility testing in pharmaceutical procedures.

### Composition\*\*

Ingredients	Gms / Litre
Tryptone #	15.000
Soya peptone	5.000
Sodium chloride	5.000
Agar	15.000
Final pH ( at 25°C)	7.3±0.2

\*\*Formula adjusted, standardized to suit performance parameters

# Equivalent to Pancreatic digest of casein

### Directions

Suspend 40 grams in 1000 ml purified/distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. If desired, aseptically add 5% v/v defibrinated blood in previously cooled medium to 45-50°C for cultivation. Mix well and pour into sterile Petri plates.

### Principle And Interpretation

Soyabean Casein Digest Agar is a widely used medium, which supports the growth of wide variety of organisms even that of fastidious ones such as *Neisseria*, *Listeria*, and *Brucella* etc. The medium with addition of blood provides perfectly defined haemolysis zones, while preventing the lysis of erythrocytes due to its sodium chloride content. It has been frequently used in the health industry to produce antigens, toxins etc. Its simple and inhibitor-free composition makes it suitable for the detection of antimicrobial agents in the food and other products. Tryptone Soya Agar is recommended by various pharmacopoeias as sterility testing medium (6, 3).

Tryptone Soya Agar conforms as per USP (6) and is used in microbial limit test and antimicrobial preservative - effective test. Gunn et al (2) used this medium for the growth of fastidious organisms and study of haemolytic reaction after addition of 5%v/v blood. The combination of tryptone and soya peptone makes this media nutritious by providing amino acids and long chain peptides for the growth of microorganisms. Sodium chloride maintains the osmotic balance.

Soyabean Casein Digest Agar does not contains X and V growth factors. It can be conveniently used in determining the requirements of these growth factors by isolates of *Haemophilus* by the addition of X-factor (DD020), V-factor (DD021), and X+V factor discs (DD022) factor to inoculated TSA plates (1).

### Type of specimen

Pharmaceutical samples, Clinical samples- blood and other body fluids

### Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,5). For Pharmaceutical samples follow appropriate techniques for sample collection, handling and processing as per pharmacopoeias. After use, contaminated materials must be sterilized by autoclaving before discarding.

### Warning and Precautions

In Vitro diagnostic Use. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

## Limitations :

1.This medium is general purpose medium and may not support the growth of fastidious organisms.

## Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

## Quality Control

### Appearance

Cream to yellow homogeneous free flowing powder

### Gelling

Firm, comparable with 1.5% Agar gel

### Colour and Clarity of prepared medium

Basal Medium : Light yellow coloured clear to slightly opalescent gel. After addition of 5-7%w/v sterile defibrinated blood : Cherry red coloured opaque gel forms in Petri plates

pH of 4.0% w/v aqueous solution at 25°C .

### pH

7.10-7.50

### Cultural response

Cultural characteristics was observed after an incubation for Bacterial at 30-35°C 18-24 hours and for Fungal at 30-35°C <=5days.

Organism	Inoculum (CFU)	Observed Lot value (CFU)	Recovery	Observed Lot value (CFU) w/blood	Recovery w/ blood	Haemolysis
<i>Bacillus subtilis subsp. spizizenii</i> ATCC 6633 (00003)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	none
<i>Staphylococcus aureus subsp. aureus</i> ATCC 25923 (00034)*	50 -100	35 -100	>=70 %	35 -100	>=70%	beta
<i>Staphylococcus aureus subsp. aureus</i> ATCC 6538 (00032)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	beta
<i>Escherichia coli</i> ATCC 25922 (00013)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	none
<i>Escherichia coli</i> ATCC 8739 (00012)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	none
<i>Escherichia coli</i> ATCC 11775 (00090)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	none
<i>Escherichia coli</i> NCTC 13167 (00179)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	none
<i>Escherichia coli</i> NCTC 9002	50 -100	35 -100	>=70 %	35 -100	>=70 %	none
<i>Pseudomonas aeruginosa</i> ATCC 27853 (00025)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Pseudomonas aeruginosa</i> ATCC 9027 (00026)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Pseudomonas aeruginosa</i> ATCC 10145 (00024)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Salmonella</i> Abony NCTC 6017 (00029)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Micrococcus luteus</i> ATCC 9341	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Streptococcus pneumoniae</i> ATCC 6305	50 -100	35 -100	>=70 %	35 -100	>=70 %	-

<i>Salmonella Typhimurium</i> ATCC 14028 (00031)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Enterococcus faecalis</i> ATCC 29212 (00087)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Candida albicans</i> ATCC 10231 (00054)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
<i>Candida albicans</i> ATCC 2091 (00055)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-
# <i>Aspergillus brasiliensis</i> ATCC 16404 (00053)*	50 -100	25 -70	50-70%			-
<i>Clostridium perfringenes</i> ATCC 13124 (00007)*	50 -100	35 -100	>=70 %	35 -100	>=70 %	-

Key : (#)- Formerly known as *Aspergillus niger* (\*) - Corresponding WDCM numbers

### Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 20-30°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition. Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

### Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (4,5).

### Reference

1. Forbes B. A., Sahm A. S. and Weissfeld D. F., 1998, Bailey and Scotts Diagnostic Microbiology, 10th Ed., Mosby Inc. St. Louis, Mo
2. Gunn B. A., Ohashi D K., Gaydos C. A., Holt E. S., 1977, J. Clin. Microbiol., 5(6) : 650.
3. Indian Pharmacopoeia, 2018, Govt. of India, Ministry of Health and Family Welfare, New Delhi, India.
4. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2<sup>nd</sup> Edition.
5. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
6. The United States Pharmacopoeia, 2019, The United States Pharmacopoeial Convention Inc., Rockville, MD.


Revision : 02/ 2018



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,  
23 Vadhani Industrial Estate,  
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951  
DB Maarn The Netherlands,  
[www.cepartner4u.eu](http://www.cepartner4u.eu)

#### Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.