



X-CUBE 70 Ultrasound System specifications

X-CUBE 70 Ultrasound System



Product Description

X-CUBE 70 is a high-performing diagnostic ultrasound system that provides diagnostic confidence and user convenience with a vast range of convenient features and superior image quality.

System Architecture

- X⁺ Architecture
- X⁺ FIT
- X⁺ Crystal Signature™
- SensiView™ Technology
- MicroFit™ Technology
- Windows10 Embedded

General Specification

Physical Specification

- Width: 580 mm
- Depth: 835 mm
- Height: 1,440 - 1,605 mm
- Weight: approx. 85 kg (System only)

Electrical Specification

- 100 - 120 V, 200 - 240 V, 50/60 Hz
- Power consumption: Max. 700 VA
- Battery (Option): Max. 22 hours (in Sleep mode)

System Design

- 23" flat panel monitor and articulated arm
- Electric motorized control panel adjustment
 - Range of height: 165 mm
 - Left/right swivel: ±30°
- Front and rear handle
- 4 active transducer ports
- 1 parking transducer port
- 1 pencil transducer port
- Transducer holders
- Gel warmer (3 temperature levels)
 - Low: 30° C
 - Medium: 34° C
 - High: 38° C
- 4 swivel wheels
- Locking mechanism
 - Front wheels: Swivel & Brake Lock
 - Rear wheels: Swivel Lock
- Integrated SSD
- ECG (Option)
- Thermal printer storage
- Built-in stereo speaker
- Built-in cable management
- System boot-up
 - from off: approx. 62 sec
 - from Sleep mode to on: approx. 7 sec

Computing power

- Main processor: Intel core i5-8400H 2.5 GHz
- Main memory: 16GB
- SSD[250GB] and HDD[1TB]

User Interface

Control Panel

- Alphanumeric keyboard with backlit
- Ergonomic key layout
- 8 TGC control levers
- 13 user customizable keys

Touch Screen

- 12.1" high definition LED backlit LCD
- Resolution: 1,280 X 800
- 16.7M color display
- Brightness and contrast adjustable (OSD)
- Adjustable screen angle: Up to 15°

Main Monitor

- 23" high definition LED backlit LCD
- Resolution: 1,920 x 1,080 (16:9 wide)
- 16.7M color display
- Contrast: 1:1,000
- IPS (In Plane Switching) technology
- Brightness and contrast adjustable
- Articulated monitor arm
 - Up/down: 145 mm
 - Swivel: +/- 90°
 - Tilt: +15° / -90°

System Overview

Applications

- Abdomen
- Obstetrics (OB)
- Gynecology (GYN)
- Cardiology
- Small Parts
- Musculoskeletal (MSK)
- Vascular
- Urology
- Emergency Medicine (EM)
- TCD
- Breast
- Appendix

Operating Modes

- 2D mode
- Harmonic mode (HAR)
- M mode
- Color M mode
- Anatomical M mode (AMM)
- Color Flow Doppler (CF) mode
- Power Doppler (PD) mode
- Directional PD mode
- Pulsed Wave Doppler (PWD) mode
- Continuous Wave Doppler (CWD) mode
- High PRF Doppler mode
- Tissue Doppler Imaging (TDI) mode
- 3D/4D mode
- Contrast Enhanced Imaging (CEUS) mode
- Strain Elastography
- Point Shear Wave Elastography (pSWE)
- Micro View Imaging mode
- Panoramic mode

Complex Modes

- Duplex: B+B, B+PW, B+CW, B+CF, B+PD, B+DPDI, B+M, B+3D, B+4D, B+Live HQ, A+B+C+3D, A+B+C+4D, B+Elastography
- Triplex: B+CF+PW, B+CF+CW, B+CF+M

Imaging Presentation

- Dual
- Live Dual
- Quad
- Duplex Image Format
 - 1:1/1:2/1:3/3:1/2:1/Full format
- Image Reverse: Left/Right
- Image Rotation: 0°, 90°, 180°, 270°
- Full screen mode

Transducer Types

- Pencil Doppler
- Linear Array
- Sector Phased Array
- Convex Array
- Endocavity
- Volume Transducer
 - Volume Convex
 - Volume Endocavity

System Standard Features

- 5 Transducer ports
- Full Speckle Reduction Imaging (Full SRI)
- Spatial Compounding Imaging (SCI)
- FTHI (Filtered Tissue Harmonic)
- PTHI (Pulse Inversion Tissue Harmonic)
- Xpeed
- Virtual Convex
- EX-FOV (Extended View)
- Raw Data Processing
- Post Processing
- Measurement Package
Including worksheets/reports for:
 - Abdomen
 - OB
 - GYN
 - Cardiology
 - Vascular
 - Urology
 - Pediatrics
 - Small Parts
 - Breast
 - MSK
 - EM

System Options

- CV M&R
- DICOM 3.0 Connectivity
- DICOM SR (OB)
- DICOM SR (Adult Echo)
- DICOM SR (Vascular)
- DICOM Query
- Auto IMT
- CW
- Anatomical-M
- Panoramic
- Tissue Doppler Image
- Stress Echo
- Cube Strain™
- Live HQ™
- Elastography
- Cube View™
- Needle Vision™ Plus
- Auto NT
- X+ 3D4D

- X+ Volume Advance
- Cube Note
- AutoEF
- STIC (Ship-hold)
- CEUS
- Silhouette View
- STL 3D Printing
- X+ MicroView
- X+ pSWE
- X+ Assistant
- X+ Auto Biometry

Peripheral Options

- External USB DVD-RW
 - Hitachi-LG Data Storage, Inc
GP65NB60
- USB thermal B&W printer
 - Sony Corporation
UP-D897, UP-897MD, UP-D898MD, UP-X898MD,
UP-D25MD, UP-25MD
- USB thermal color printer
 - Sony Corporation
UP-D25MD, UP-25MD
- Wireless LAN
- Tri-pedal footswitch, with programmable functionality
- Side tray
- Endocavity transducer holder

Display Annotation

- Patient ID: Up to 64 characters
- Patient name: First, last and middle
- Gender, age and birth date
- Gestational age from LMP, EDD and GA
- Transducer name
- Institution / hospital name: up to 25 characters
- Date format: 3 types
 - YYYY/MM/DD
 - MM/DD/YYYY
 - DD/MM/YYYY
- Time format: 2 types
 - 12 hours / 24 hours
- Acoustic power
 - Mechanical index (MI)
 - Thermal index (TI): TIS, TIB, TIC
- Transducer name
- Preset

- Transducer orientation
- Depth
- Focus
- TGC line
- Frequency
- Frame rate
- Power
- Mode name
- Gray scale bar
- Arrow
- Indicator
- Text
 - Size: 1 - 10
- Body pattern
 - 201 sets
 - Size: small, medium, large
- Biopsy guideline
- ECG and Respiration line
- Thumbnail
- Stored image info
- Cine gauge
- Measurement result
- Trackball controls and status

- Full SRI
- Edge Enhance
- Reject
- M
 - Colorize
 - Dynamic Range
 - Gain
 - Gray Map
 - Reject
 - Sweep Speed
- PWD
 - Gain
 - Auto Calc.
 - Baseline
 - Sweep Speed
 - Invert
 - Colorize
 - Gray Map
 - Auto Angle
 - Angle Correct
 - Reject
 - Dynamic Range
- CWD
 - Angle Correct
 - Auto Calc.
 - Baseline
 - Colorize
 - Dynamic Range
 - Gray Map
 - Reject
 - Sweep Speed
- CF/PD
 - Baseline
 - Blend / Adaptive Blend
 - Colorize
 - Color Map
 - Invert (CF / DPDI)
 - Threshold
- PD
 - Baseline
 - Blend / Adaptive Blend
 - Colorize
 - Color Map
 - Invert (CF / DPDI)
 - Threshold

System Parameters

System Presets

- User Programmable Preset Capability
- Factory Default Preset Data
- Languages: English, German, Spanish, French, Italian, Portuguese, Russian, Chinese
- Programmable Annotation Library
- Body Patterns

Controls Available on Freeze (Post-Processing)

- Read Zoom
- 2D
 - Zoom
 - Reverse
 - Gain
 - Dynamic Range
 - Up/Down
 - Colorize
 - Gray Map

Controls Available while Live (Pre-Processing)

- Write Zoom
- 2D
 - Angle Steer
 - Colorize
 - Dynamic Range
 - Edge Enhance
 - Frequency
 - Full SRI
 - Gray Map
 - Gain
 - Harmonic
 - Pulse inversion harmonic
 - Invert
 - Line Density
 - Persist
 - Power
 - Reject
 - Spatial Compound
- M
 - Anatomical M
 - AMM Angle
 - Colorize
 - Dynamic Range
 - Gray Map
- PWD
 - Angle
 - Angle Steer
 - Auto Angle
 - Baseline
 - Colorize
 - Dynamic Range
 - Frequency
 - Gray Map
 - Gain
 - Line Density
 - Power
 - Reject
 - Scale (PRF)
 - SV Size
 - Sweep Speed
 - Sound
 - Wall Filter
- CWD
 - Angle

- Baseline
- Dynamic Range
- Gray Map
- Gain
- Scale (PRF)
- Frequency
- Power
- Reject
- Sound
- Sweep Speed
- Wall Filter
- CF/PD
 - Angle Steer
 - Baseline
 - Color Map
 - Ensemble
 - Frequency
 - Flow State
 - Gain
 - Line Density
 - Power
 - Scale (PRF)
 - Smooth
 - Wall Filter

Image Processing and Presentation

- System Frequency Range: 1 - 20 MHz
- System processing channel: 7,645,867
- ADC: 14 bits
- Displayed Imaging Depth: 0 - 40 cm
- Max. Frame Rate (Transducer dependent)
 - 2D: 509 (Hz/FPS)
 - Color: 270 (Hz/FPS)
 - PW: 299 (Hz/FPS)
 - Volume Convex: 17 (Hz/VPS)
 - Volume Endocavity: 100 (Hz/VPS)
- Gray Scale: 256 levels
- Dynamic Range: Max. 250 dB
- Max. Scan line: 1,024
- Adjustable Field of View (FOV)
- Image Reverse: Right/ Left
- Image Rotation: 0°, 90°, 180°, 270°
- Focus
 - Focal number: Max. 8

- Focal position: Max. 19
- Zoom
 - Write Zoom: Max. 11.9x (Transducer dependent)
 - Read Zoom: Max. 10x
- CF (Transducer dependent)
 - 300 Hz - 13.1 KHz
 - 2 cm/s - 423.9 cm/s
- PW (Transducer dependent)
 - 300 Hz - 25 kHz
 - 2.3cm/s - 1000 cm/s
- CW
 - 1.6 KHz - 78.1 KHz
 - 2.5 cm/s - 30.8 m/s

- DICOM Print
- DICOM Storage commitment
- DICOM Media
- DICOM Worklist
- DICOM MPPS
- DICOM Q&R
- DICOM Structured Report (OB-GYN)
- DICOM Structured Report (Echocardiography)
- DICOM Structured Report (Vascular)
- Cube View™
- Cube Note
- Network Storage
- Wireless LAN (Option)

CINE Memory

- Cine memory: 716 MB, 16,384 frames, 312 volumes
- Prospective and retrospective
- Cine review, save and edit
- Run, stop, frame move
- Play speed: 50% 100%, 200%, 300%, 400%, 500%, 600%, 700%, 800%, 900%, 100%

Image Archive

- Clipboard: Displays thumbnail images of the acquired data for the current exam
- Previewing Clipboard Images: An enlarged preview of the image
- Recalling Images from the Clipboard
- E-View™: Archived images from the past patient exams appear as well as images stored for the current exam
- 916GB of the 1TB internal hard drive reserved for patient data
- Image store
 - Up to 916,000 images (with JPG compression)
- Image format
 - BMP, JPEG, DICOM, AVI, MPEG
- Export / Backup / Restore
- DVR
 - VCR: VCR/DVR, USB storage
 - VCR Format: NTSC or PAL, MPEG or AVI
 - VCR Maximum Time (min.): 1 to 20 (Only available for USB storage.)

Connectivity

- Verification
- DICOM Storage

Scanning Parameters

2D Mode

- Angle Steer (Transducer dependent): 7 steps (-15 - 15°)
- Colorize: 0 - 6
- Dynamic Range: 30 - 250 dB
- Edge Enhance: 0 - 10
- Frequency
- Full SRI: 0 - 5
- Gray Map: 0 - 8
- Gain: 0 - 100
- Harmonic
- Pulse inversion harmonic
- Line Density: 0 - 5
- Persist: 0 - 4
- Power: 1 - 100 %
- Reject: 0 - 6
- Spatial Compound: 0 - 3

M Mode

- Anatomical M: On / Off
- AMM Angle
- Colorize: 0 - 6
- Dynamic Range: 30 - 150 dB
- Gray Map: 0 - 6
- Gain: 0 - 100
- Reject: 0 - 10
- Power: 1 - 100 %
- Sweep Speed: 0 - 7

PWD Mode

- Angle: -89° - 89°
- Angle Steer (Transducer dependent): 7 steps (-20 - 20°)
- Auto Angle
- Baseline: -8 - 8
- Colorize: 0 - 44
- Dynamic Range: 30 - 120 dB
- Frequency
- Gray Map: 0 - 13
- Gain: 0 - 100
- Power: 1 - 100 %
- Reject: 0 - 10
- Scale (PRF): 0.3 - 25.7 KHz (Transducer dependent)
- SV Size: 5 - 11 mm
- Sweep Speed: 0 - 7
- Sound: 0 - 100%
- Wall Filter: 0 - 7
- Invert

CWD Mode

- Angle: -89° - 89°
- Baseline: -8 - 8
- Dynamic Range: 30 - 120 dB
- Gray Map: 0 - 9
- Colorize: 0 - 44
- Gain: 0 - 100
- Scale (PRF): 0.3 - 78.1 KHz (Transducer dependent)
- Frequency
- Power: 1 - 100 %
- Reject: 0 - 10
- Sound: 0 - 100 %
- Sweep Speed: 0 - 7
- Wall Filter: 0 - 7
- Invert

CF/PD Mode

- Angle Steer (Transducer dependent): 7 steps (-20 - 20°)
- Baseline: -100 - 100
- Color Map: 0 - 26 (CF), 0 - 8 (PD), 0 - 9 (DPDI)
- Ensemble: 6 - 16
- Frequency
- Flow State: Low, Mid, High
- Gain: 0 - 100
- Line Density: 0 - 3

- Power: 1 - 100 %
- Scale (PRF): 0.3 - 12.1 KHz (Transducer dependent)
- Smooth: 0 - 10
- Wall Filter: 0 - 7

3D/4D Mode (X+ 3D4D)

- Quality: Low, Mid, H1, H2, Max
- Volume Angle
 - Volume Convex: 15 - 80 (5 steps)
 - Volume Endocavity: 10 - 95 (5 steps)
- Curved ROI
- Render Direction: Up, Down, Lt, Rt, Front, Back
- Brightness: 1 - 100
- Contrast: 1 - 100
- Threshold: 0 - 255
- Opacity: 0 - 255
- Smooth: 0 - 20
- 3D SRI: 0 - 5
- 2D SRI: 0 - 5
- ColorMap
 - Gray map: 0 - 8
 - 2D Colorize: 0 - 6
 - 3D Color Map: 1 - 18

3D/4D Aquisition Mode

- STIC (Ship-hold)
 - Acq. Time: 7 - 15 sec
 - Volume Angle: 15, 20, 25, 30, 35, 40, 45, 50

3D/4D Visualization Mode

- MPR
 - Display format
 - Ref. Plane
 - 3D Orientation
 - Edit ROI
 - Render Setup: Surface, Depth, Max, Min, XRay, Light, Light2
 - Alpha Mode: 1, 2
 - Cine
 - Cine Calc.
 - Initial
 - Easy Cut
 - Clear Face
 - 3D Print
- Silhouette: On / Off

- Sectional Planes
 - Multi Plane
 - Cubic View: Mode1, Mode2
 - AnySlice
- Multi Slice
 - Display format: 1*1 / 1*2 / 2*2 / 3*3 / 4*4
 - Ref. plane
 - Active view
 - Guide view
 - Pitch: 1 - 10 mm
 - Prev. / Next
 - Oblique Angle: -90 - 90°
 - Thinkness: 0 - 20 mm
 - Slice Thickness: 2 mm, 5 mm, 8 mm, 12 mm, 20 mm
- AnySlice
 - Display format
 - Reference line (Ref. Line)
 - Delete / Delete All
 - Thinkness: 0 - 20 mm
 - Slice Thickness: 2 mm, 5 mm, 8 mm, 12 mm, 20 mm
- VOCAL
 - Segment Mode: Sphere, Manual, Auto
 - Structure: Hypo, Hyper
 - Segment Angle: 10° / 15° / 30°
 - Frame
 - Sphere Size
 - Render Type: Wire, Surface, Volume
 - Shell Type: Off / In / Out / In-Out
 - Rotate X / Rotate Y / Rotate Z
 - 3D Color Map: 1 - 18
 - Shell Thickness: 1.0 - 20.0 mm

3D/4D Live HQ™

- Live HQ Setup
- Light Direction
- Move Light
- Live HQ color map: 1 - 12
- HQ Scatter: 0 - 10
- Front Scale: 0 - 20
- Back Scale: -20 - 25

3D/4D Silhouette

- Silhouette1, Silhouette2
- Mix: 0 - 100
- Level: -100 - 100

- Light Direction
- Move Light
- Color Map: 1 - 12
- Value: 0 - 100
- Transparency: 0 - 100

Panoramic Imaging

- Colorize: 0 - 6
- Dynamic Range: 30 - 250 dB
- Edge Enhance: 0 - 4
- Frequency
- Full SRI: 0 - 5
- Gray Map: 0 - 6
- Gain: 0 - 100
- Harmonic
- Pulse inversion harmonics
- Line Density: 0 - 5
- Persist: 0 - 4
- Power: 1 - 100%
- Reject: 0 - 6

Strain Elastography

- Live Dual: On / Off
- Invert: On / Off
- Persist: 0 - 5
- Opacity: 10 - 100
- Color Map: 0 - 5

point Shear Wave Elastography

- pSWE Units: kPa, m/s
- 12 samples (S1 - S12)

Needle Vision™ Plus

- L/R Flip
- Angle: Shallow, Medium, Steep

Contrast Enhanced Ultrasound Imaging (CEUS)

- Live Dual: On / Off
- Side Control: CEUS, Tissue
- Flash: On / Off
- Timer: On / Off
- Line Density: 0 - 5
- Edge Enhance: 0 - 10
- Sepia: On / Off

- Rotation: 0°, 180°
- Dynamic Range: 30 - 250 dB
- Sepia Map: 0 - 4
- Output Power: -20 - 0 dB
- Flash Frame: 1 - 100
- Frequency
- Persist: 0 - 4
- Reject: 0 - 10

Transducers

SC1-7H

- Transducer type: Convex
- Frequency: 1.0 - 7.0 MHz
- Field of view: 60°
- Elements: 192
- Radius of curvature: 60 mm
- Footprint: 72.4 X 16.8 mm
- Application: Abdomen, EM, Gynecology, Obstetrics, Pediatric, Urology
- SC1-6 Biopsy Starter Kit
- X⁺ Crystal Signature

L3-12X

- Transducer type: Linear
- Frequency: 3.0 - 12.0 MHz
- Field of view: 51.2 mm
- Elements: 256
- Footprint: 58.2 X 10.4 mm
- Application: Pediatric, EM, MSK, Vascular, Small Parts
- L3-12X Biopsy Starter Kit

L3-15H

- Transducer type: Linear
- Frequency: 3.0 - 15.0 MHz
- Elements: 192
- Footprint: 45.2 X 8.3 mm
- Application: EM, MSK, Vascular, Small Parts
- L3-12 Biopsy Starter Kit

L10-25H

- Transducer type: Linear
- Frequency: 10.0 - 25.0 MHz

- Field of view: 25.0 mm
- Elements: 192
- Footprint: 32.8 X 6.8 mm
- Application: EM, MSK, Vascular, Small Parts
- Biopsy is not available
- Seamless window

SL3-19H

- Transducer type: Linear
- Frequency: 3.0 - 19.0 MHz
- Field of view: 38.4 mm
- Elements: 192
- Footprint: 43.3 X 8.3 mm
- Application: Abdomen, Pediatric, Gynecology, Obstetrics, EM, MSK, Vascular, Small Parts, TCD
- L3-12 Biopsy Starter Kit
- X⁺ Crystal Signature
- Seamless window

MP1-5X

- Transducer type: Sector Phased
- Frequency: 1.0 - 5.0 MHz
- Field of view: 90°
- Elements: 80
- Footprint: 23.0 X 15.4 mm
- Application: Abdomen, Pediatric, Cardiac, EM, TCD
- Biopsy is not available
- X⁺ Crystal Signature

EC2-11H/EV2-11H

- Transducer type: Endocavity
- Frequency: 2.0 - 11.0 MHz
- Field of view: 200° (Max. 230°)
- Elements: 192
- Footprint: 19.0 X 17.4 mm
- Application: Gynecology, Obstetrics, Urology, EM
- EV2-11H Reusable Starter Kit
- X⁺ Crystal Signature

SVC1-8H

- Transducer type: Volume Convex
- Frequency: 1.0 - 8.0 MHz
- Field of view: 79°
- Radius of curvature: 39 mm

- Elements: 192
- Footprint: 58.64 X 38.58 mm
- Application: Abdomen, Gynecology, Obstetrics, Pediatric, Urology, EM
- Biopsy is not available
- Seamless window

VE3-10H

- Transducer type: Volume Endocavity
- Frequency: 3.0 - 10.0 MHz
- Field of view: 152°
- Radius of curvature: 10 mm
- Elements: 192
- Footprint: 25.1 X 25.1 mm
- Application: Gynecology, Obstetrics, Urology, EM
- VE3-10H Disposable Starter Kit
- VE3-10H Reusable Starter Kit

CW2.0

- Transducer type: Pencil Doppler
- Frequency: 2.0 MHz
- Footprint: 17.2 X 17.2 mm
- Application: Cardiac
- Biopsy is not available

Measurement and Calculation

Basic Measurement

- Distance
- Area Circumference
- %Stenosis
- Volume
- Angle
- A/B ratio
- Volume Flow Area
- Time
- Slope
- HR
- Velocity
- Acceleration
- S/D Ratio
- PI
- RI
- Trace

- Auto Trace
- VTI

Abdomen Measurement

- Aorta
- Renal
- Renal A
- Renal V
- Hepatic A
- IVC
- Hepatic V
- Mid HV
- MPV
- Portal V
- Splenic A
- Splenic V
- SMA
- SMV
- IMA
- Iliac A
- Iliac V
- Vessel
- HR

Breast Measurement

- Mass 1 - 10
- Duct dilatation
- Breast Flow
- HR

Cardiology Measurement

- Ao/LA
- LA Vol (Diam)
- Aorta Diam
- Pul. Diam
- Vena Cava
- RV
- Simpson BP
- Simpson SP
- Modified Simpson
- Area Length
- Teichholz
- LV Mass T-E
- LV Mass A-L
- LA Vol (A-L)

- LA Vol (Simpson)
- RA Vol (A-L)
- RA Vol (Simpson)
- TV
- PISA
- MV
- AV
- PV
- Qp/Qs
- Pul. Vein
- TDI
- HR

Gynecology Measurement

- Uterus
- Ovary
- Follicle
- Fibroid
- Pelvic Floor
- Early Gestation
- Ovarian
- Uterine
- FHR
- Ovarian A
- Uterine A
- Vessel
- Aorta
- Umbilical A
- Placenta
- Des. Aorta
- HR

Obstetrics Measurement

- Early Gestation
- Fetal Biometry
- Fetal Long Bones
- Fetal Cranium
- Fetal Others
- Maternal Others
- Generic
- AFI
- CRAR
- 4 Chamber view
- Aortic Arch
- LVOT

- Outflow tract
- RVOT
- Thorax
- Venous
- FHR
- Tricuspid Valve
- Mitral Valve
- Main Pulmonary A
- Aorta
- Ductus Art
- Umbilical A
- Umbilical V
- Ductus Ven
- Pulmonary Valve
- Aortic Valve
- Pulmonary Vein
- Carotid
- MCA
- SMA
- Celiac A
- Uterine A
- IVC
- Automated nuchal transparency measurement
- Automated fetal biometry measurement
 - BPD, HC, AC, FL, Humerus

Emergency Medicine Measurement

- Aorta
- Renal
- Bladder
- %Stenosis
- A/B ratio
- SMA
- IMA
- IVC
- Uterine A
- Ovarian A
- Vessel
- HR

MSK Measurement

- Volume
- HIP (BA) / Hip (AB)
- Vessel
- HR

Pediatric Measurement

- Aorta
- Renal
- Renal A
- CHA
- IVC
- Hepatic V
- Mid HV
- MPV
- Portal V
- Splenic A
- Splenic V
- SMA
- SMV
- IMA
- Iliac A
- Iliac V
- Vessel
- Hip
- HR

Renal Measurement

- Renal L
- Renal
- Aorta
- Renal A
- Portal V
- Vessel
- HR

Small Parts Measurement

- Thyroid
- Isthmus
- Nodule
- Lymph Node
- Parathyroid
- Thyroid Flow
- Testicle
- Testicle Flow
- HR

Urology Measurement

- Renal
- Bladder

- Testicle
- Prostate
- Residual
- Vessel
- Bladder
- HR

Vascular Measurement

- CCA
- ICA
- ECA
- Bulb
- Vert. A
- Subclav. A
- Axill. A
- Basilar A
- Brach. A
- DBA
- Basilic A
- Radial A
- Ulnar A
- SPA
- Innom. A
- IJA
- Subclav. V
- Ceph. V
- Brach. V
- DBV
- Basilic V
- Radial V
- Ulnar V
- Innom. V
- CIA
- EIA
- IIA
- CFA
- DFA
- SFA
- Pop. A
- PTA
- ATA
- Peron. A
- DPA
- Prof. A
- CIV

- EIV
- IIV
- CFV
- DFV
- SFV
- GSV
- LSV
- Pop. V
- PTV
- ATV
- Peron. V
- DPA
- Prof. A
- MCA
- ACA
- PCA P1
- PCA P2
- Vert. A
- Basilar A
- PComA
- AComA
- Renal A
- Renal V
- Segm. A
- Interlob. A
- Arcuate A
- ABD. Aorta
- SMA
- IMA
- Hepatic A
- Portal V
- Splenic A
- Straight Gft
- Loop Gft
- Cimino Gft
- Comp Gft
- HR
- Automated intima media thickness measurement

Inputs and Outputs (I/O)

- USB port (6 ports)
 - 4 USB 3.0 ports
 - 2 USB 2.0 ports
- Audio out port (Left)

- Audio out port (Right)
- HDMI port
- Ethernet port
- S-Video output port
- VGA port
- CVBS output port

Safety Compliance

System Classifications

- IEC/EN 60601-1:6.8.1
 - Type of protection against electrical shock: Class I
 - Degree of protection against electrical shock (Patient connection): Type BF equipment
 - Degree of protection against harmful ingress of water: Ordinary equipment and all of applied parts (IPX7, IPX8) except ECG meet ingress protection level according to IEC 60529.
 - Degree of safety of application in the presence of a flammable anesthetic material with air or with oxygen or nitrous oxide: Equipment not suitable for use in the presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide.
 - Mode of operation: Continuous operation

Standards Compliance

- MDD 93/42/EEC
- EN ISO 10993-1:2009 (ISO 10993-1:2009)
- EN ISO 10993-10:2013 (ISO 10993-10:2010)
- EN ISO 10993-5:2009 (ISO 10993-5:2009)
- EN ISO 13485:2016 (ISO 13485:2016)
- EN ISO 14155:2011+AC:2011 (ISO 14155:2011)
- EN ISO 14971:2012 (ISO 14971:2007)
- EN ISO 15223-1:2016 (ISO 15223-1:2016)
- EN 1041:2008+A1:2013
- EN 55011:2016+A1:2017 (CISPR 11:2015/AMD1:2016)
- EN 55032:2015 (CISPR 32:2015)
- EN 60601-1:2006+A1:2013 (IEC 60601-1:2005/AMD1:2012)
- EN 60601-1-2:2015 (IEC 60601-1-2:2014)
- EN 60601-1-6:2010+A1:2015 (IEC 60601-1-6:2010/AMD1:2013)
- EN 60601-2-37:2008+A11:2011+A1:2015 (IEC 60601-2-37:2007/AMD1:2015)
- EN 61000-3-2:2014 (IEC 61000-3-2:2018)
- EN 61000-3-3:2013 (IEC 61000-3-3:2013)

- EN 61000-4-11:2004 (IEC 61000-4-11:2004)
- EN 61000-4-2:2009 (IEC 61000-4-2:2008)
- EN 61000-4-3:2006+A1:2008+A2:2010 (IEC 61000-4-3:2006/AMD2:2010)
- EN 61000-4-4:2012 (IEC 61000-4-4:2012)
- EN 61000-4-5:2014 (IEC 61000-4-5:2014)
- EN 61000-4-6:2014 (IEC 61000-4-6:2013)
- EN 61000-4-8:2010 (IEC 61000-4-8:2009)
- EN 62304:2006+A1:2015 (IEC 62304:2006/AMD1:2015)
- EN 62366-1:2015 (IEC 62366-1:2015)
- MEDDEV 2.12/2 Rev.2
- MEDDEV 2.7.1 Rev.4
- NEMA UD2:2004 (R2009)
- Medical Devices Regulations (SOR/98-282)
- 의료기기법, 의료기기법 시행령, 의료기기법 시행규칙

Not all features or specifications described in this document may be available in all transducers and/or modes.

© ALPINION MEDICAL SYSTEMS – All rights reserved.

ALPINION MEDICAL SYSTEMS reserves the right to make changes in specifications and features shown hereby, or discontinue the product described at any time without notice or obligation. Contact you ALPINION representative or the most current information.