

Declaration of Conformity

We: madur Polska Sp. z o.o.

(Supplier's name)

ul. Sadowa 37 – 95-100 Zgierz – Poland

(Address)

declare under our sole responsibility that the product:

GA-12 portable gas analyser

GA-12plus portable gas analyser

GA-21plus portable gas analyser

GA-40Tplus portable gas analyser

GA-60 portable gas analyser

(Name and type)

to which this declaration relates is in conformity with the following normative documents:

EN50379 – Specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances

EN50379-1 – General requirements and test methods

EN50379-2 – Performance requirements for apparatus used in statutory inspections and assessment

EN50379-3 – Performance requirements for apparatus used in non-statutory servicing of gas fired heating appliances

EN50270 – Electromagnetic compatibility – Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen


US EPA CTM 034 method – Determination of Oxygen, Carbon Monoxide and Oxides of Nitrogen from Stationary Sources For Periodic Monitoring (Portable Electrochemical Analyzer Procedure)

US EPA CTM 030 method – Determination of Nitrogen Oxides, Carbon Monoxide, and Oxygen Emissions from Natural Gas-Fired Engines, Boilers and Process Heaters Using Portable Analyzers

(Number, date of issue, title of normative documents)

Zgierz – Poland, 2014-07-28

(Place and date of issue)



madur
POLSKA
Sp. z o.o.
MENEDŻER SPRZEDAŻY
mgr Inż. Jacek Kleśta

(Signature of authorised person)