



泰博科技股份有限公司

TaiDoc Technology Corp.

新北市24888五股區五工二路127號B1-7樓

B1-7F, No.127, Wugong 2nd Rd., Wugu Dist.,

New Taipei City 24888, Taiwan

Tel : +886-2-6625-8188

Fax : +886-2-6625-0288

www.taيدoc.com

## EC Declaration of Conformity

We, TaiDoc Technology Corporation

B1-7F, No.127, Wugong 2nd Road, Wugu Dist, 24888 New Taipei City, TAIWAN

declare under our sole responsibility that the product

Product Name : Blood Glucose Test Strip  
Product model : TD-4302  
Classification : 98/79/EC (IVDD), Annex II, List B  
Conformity Assessment Route : 98/79/EC (IVD), Annex IV excluding section 4 & 6  
EC Certificate Number : V1 052126 0042 Rev.01  
European Representative : MedNet GmbH  
Borkstraße 10, 48163 Münster , Germany  
Notified Body (CE0123) : TÜV SÜD Product Service GmbH  
Ridlerstraße 65, 80339 München, Germany  
GMDN code : 62537

to which this declaration relates is in conformity with the following standard(s) or other normative document(s) :

|                     |   |
|---------------------|---|
| ISO 13485:2016      | Medical devices. Quality management systems. Requirements for regulatory purposes   |
| EN ISO 14971:2012   | Medical devices. Application of risk management to medical devices  |
| EN ISO 15197:2015   | In vitro diagnostic test systems. Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus          |
| EN 15223-1:2016     | Medical devices. Symbols to be used with medical device labels, labelling and information to be supplied. General requirements              |
| EN ISO 18113-1:2011 | In vitro diagnostic medical devices. Information supplied by the manufacturer (labelling). Terms, definitions and general requirements      |
| EN ISO 18113-4:2011 | In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic reagents for self-testing    |
| EN ISO 18113-5:2011 | In vitro diagnostic medical Devices. Information supplied by the manufacturer (labelling). In vitro diagnostic instruments for self-testing |
| EN ISO 23640:2015   | In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents  |



泰博科技股份有限公司  
TaiDoc Technology Corp.

28 龍市 24886 五股區五工三路 22 號 61-7 樓  
C/F 2F, No. 137, Wugong 2nd Rd., Wugu Dist.,  
New Taipei City 24886, Taiwan

Tel : +886-2-6625-8188  
Fax : +886-2-6625-8288

www.tai.doc.com

|                      |   |
|----------------------|---|
| EN 1041:2008+A1:2013 | Information supplied by the manufacturer of medical devices                   |
| EN 13532:2002        | General requirements for in vitro diagnostic medical devices for self-testing |
| EN 13612:2002        | Performance evaluation of in vitro diagnostic medical devices                 |

2019.3.11.

Date of Issue

*Jim Jan*

**Jim Jan**

Management Representative