

GAUZE BANDAGE STERILE (14cmx7mt) ANALYSIS REPORT

DATE:05.02.2022

BANDAGE

MANUFACTURER NAME: BİRİNCİ STERİLİZASYON SAĞLIK HİZM. TİC. SAN. VE LTD. ŞTİ
MANUFACTURER ADD.: Battalgazi Mahallesi. Gecesevası Sokak. No: 2/J4 Sultanbeyli / ISTANBUL

APPLIED STANDARD:TS EN ISO 14079

INSPECTION AND EXPERIMENTS	REQUESTED	EXPERIMENT RESULTS
MATERIAL FEATURE	100% COTTON	APPROPRIATE
APPEARANCE FEATURE	NON-DEFECTIVE	APPROPRIATE
DIMENSION	14CMX7MT	APPROPRIATE
DENSITY	32g/m ²	APPROPRIATE
SINKING TIME	9 s	APPROPRIATE
RUPTURE STRENGTH	5 kgf/5 cm (min)	APPROPRIATE
SYNTHETIC FIBER	0.1% (max)	APPROPRIATE
pH	4.5-8.6	APPROPRIATE
ETHER SOLUBLE SUBSTANCE	0.50% (max)	APPROPRIATE
SURFACE ACTIVE SUBSTANCE	Single row foam (max)	APPROPRIATE
SULFATE ASH	0.75% (max)	APPROPRIATE
WATER SOLUBLE MATTER	0.50% (max)	APPROPRIATE
STARCH AND DEXSTRIN	Should Appear Colorless	APPROPRIATE

APPROVED BY :

LOT: 215623
REF: SGB14700
MFG. DATE: 02/2022
EXP. DATE: 02/2027

BİRİNCİ STERİLİZASYON
SAĞLIK HİZMETLERİ TİC. VE SAN. LTD. ŞTİ.
Battalgazi Mh. Gecesevası Sok. No: 2/J4/1 Sultanbeyli / İST
Sultanbeyli / İST 1781392708
İ.T.O.: 216161-Mersis No: 0178139270200001
Firma Tanımlayıcı No: 2967269314960