

Zibo Eastmed Healthcare Products Co., Ltd

Add: No.118 Huaguang Road, Zhangdian District, Zibo City 255000, Shandong China


TECHNICAL SPECIFICATION OF 20ML SYRINGE WITHOUT NEEDLE & LUER LOCK

REF NO.: DMSL-20

- Materials: PP + Latex + ABS
- Size: 20ml
- Type: 3-parts, luer lock

- Non-pyrogenic
- Latex
- For single use
- ETO sterilized

CE 0197



| No. | Designation | Materials |
|-----|-------------|-----------|
| 1 | Core pin | PP |
| 2 | Barrel | PP |
| 3 | Plunger | PP |



PACKING:

1. Individual blister bag
2. Packing is adapted to the ethylene oxide sterilization. The cover mentions the necessary indications for content identification: designation, item code, manufacture date and expire date and lot number.
3. Packing details: 1pc/blister bag, 25pcs/box

QUALITY:

1. The product quality is guaranteed by inspections at all levels of manufacture, all our products are manufactured in 100,000 grade clean room, the final control is made at the level of packaging.
2. The packaged vaginal speculum is sterilized by ethylene oxide according to a validated method in the physical and microbiological field. The shelf life is 5 years from the manufacture date.
3. The product conforms to MDD 93/42/EEC amended by 2007/47/EC and maintains a quality assurance system according to ISO 13485.

For and on behalf of
ZIBO EASTVED HEALTHCARE PRODUCTS CO., LIMITED
John W
Authorized Signature(s)