



Nuova Aptaca Srl Regione Monforte, 30 - 14053 Canelli (Asti) Italy Tel. (+39) 0141/83.50.75 – Fax (+39) 0141/83.52.92 E-Mail: info@aptaca.com – Website: www.aptaca.com

CERTIFIED COMPANY UNI EN ISO 9001 & UNI CEI EN ISO 13485

SCHEDA TECNICA PRODOTTO **TECHNICAL DATA SHEET** **MICRO TEST TUBES WITH SECURE LOCK CLOSURE**



Micro test tubes with secure lock integrated in the cap to prevent accidental cap opening and evaporation. In polypropylene, graduated, with writing area and flat pierceable cap. They are certified free of human DNA, PCR inhibitors, ATP, DNase, RNase and pyrogen. Sterile. Non cytotoxic, non hemolytic. Suitable for PCR and molecular biology applications. Temperature of use: -80°C / +100°C. Autoclavable open at +121°C for 20 minutes.

Cod.	Vol. ml	Dim. mm	RCF max
1103/SCL	0.5	Ø 7.8x30	30,000 g
1003/SCL	1.5	Ø 10.6x38.6	30,000 g
1303/SCL	2.0	Ø 10.7x40.1	30,000 g
1403/SCL	5.0	Ø 16.6x56.5	21,000 g