

**ITEM 2
FREEZER**

No.	Minimum Technical Requirement	Bidder's Offer		
		Conformity		Remarks
		Yes	No	
1	Model	NS 1400/40		
2	Country of Origin	Czech republic		
3	Name of Manufacture	Frigera s.r.o.		
4	For storage and preservation of frozen blood products	yes		
5	Type: vertical, with two doors	Yes		
6	Internal room volume from 1300 lit to 1400 lit	yes		1400l
7	Temperature: - from minus 25 ° C to minus 30 ° C; - adjustable; - programmable; - accuracy $\pm 1^\circ \text{C}$	yes		minus 40°C
8	Temperature monitoring: - ensuring the integration of thermal data with local SAP Blood Service;	Yes		
9	Electronic control system of temperature.	Yes		
10	Information panel: will allow the visualization of the temperature parameters in daily / weekly time	Yes		
11	provided with alarm in case of exceeding the programmed parameters	Yes		
12	Interior surfaces made of stainless steel resistant to the action of biocidal substances used for sanitary processing	Yes		
13	External surfaces made of stainless steel resistant to the action of biocidal substances used for sanitary processing.	Yes		
14	Mandatory components present: material of stainless steel drawers	Yes		



FRIGERA, s.r.o.
Hájecká 1304/14a, 618 00 Brno
IČ 28519400 DIČ CZ28519400

-04-

15	perforated surface for air circulation	Yes		
16	drawers in number of 12 - 14 units, adjustable	Yes		
17	secured with telescopic system	Yes		
18	Cooling system: Forced air	Yes		
19	Automatic defrost system	Yes		
20	Refrigerant: Ecological, approved by the EC	Yes		
21	EC / MS Declaration of Conformity and / or EC / MS Certificate of Conformity	Yes		
22	Declaration from the manufacturer confirming that the medical device is intended for the storage of frozen blood products.	Yes		
23	Year of production of the device 2021-2022	Yes		
24	Warranty period of not less than 36 months	Yes		



FRIGERA, s.r.o.
Hájecká 1304/14a, 618 00 Brno
IČ 28519400 DIČ CZ28519400

04-