

Test Report

No.: QDHL2008008162MD

Date: AUG.27,2020

Page: 1 of 2

SUZHOU COLOUR-WAY NEW MATERIAL CO., LTD.

NO. 20, ANMIN ROAD, HUANGDAI TOWN, XIANGCHENG DISTRICT,

215152 SUZHOU CITY, JIANGSU RPOVINCE, PEOPLE'S REPUBLIC OF CHINA

The following sample(s) was/were submitted and identified by the client as:

Sample Description : EXAMINATION GLOVES

Material : NATURE LATEX

Lot No. : 20060120

Lot Size : 50000PAIRS

Sample Quantity : 300PCS

Manufacturer : SUZHOU COLOUR-WAY NEW MATERIAL CO., LTD.

Manufacturer Date : 2020.06

Storage Condition : AVOID LIGHT
Sample Receiving Date : AUG.12,2020

Testing Period : AUG.12,2020 TO AUG.27,2020

<u>Test Requested</u> <u>Result</u>

EN 455-1:2000 Medical Gloves for Single Use – Part 1: Requirements and

Testing for Freedom from Holes (Clause 5.1)

Remark: - Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

SGS-CSTC Standards Technical Services (Qingdao)

Co., Ltd.

Jesseen Gens

SS 2.

Jessica Gao Approved Signatory



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions for Electronic Documents, subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD

7389595



Test Report

No.: QDHL2008008162MD

Date: AUG.27,2020

Page: 2 of 2

Test Conducted:

EN 455-1:2000 Medical gloves for single use - Part 1: Requirements and testing for freedom

Number of test sample	:	200 PCS
Sample size	:	M
Number of non-conforming gloves	:	0

Clause	Test Items	Result		
5	Watertightness test for detection of holes	- Co - Co - Co - Co		
5.1	Referee testing	Pass (See note 1)		

. 7		- 25	Sample quantity: 200pcs, AQL:1.5, Ac:7, Re:8, Found:0.
Note	1	1	The sample selecting amount for this clause is deviated to 200 pcs as assessed by SGS.


Sample Photo:

Received Sample



Received Sample





SGS authenticate the photo on original report only

End of Report***



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions for Electronic Documents at http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

389596

www.sgsgroup.com.cn e sgs.china@sgs.com