

EDI™ Novel Coronavirus COVID-19 IgM ELISA Kit

Enzyme Linked Immunosorbent Assay (ELISA) for the qualitative detection of the COVID-19 IgM in serum.

REF KT-1033 IVD CE    

INTENDED USE

This kit is for in vitro diagnostics use only. The kit is detecting novel COVID-19 IgM antibody in human serum. It is for screening or to aid in the diagnosis of COVID-19. Patients with suspected clustering cases require diagnosis or differential diagnosis of novel coronavirus infection. The assay is validated manually, but can be adapted to an automated instrument. The assay is for the qualitative detection only.

INTENDED USER

This kit is for laboratory professional use or healthcare professionals.

SUMMARY OF PHYSIOLOGY

2019 novel coronavirus (COVID-19) is a single-stranded RNA coronavirus². Comparisons of the genetic sequences of this virus have shown similarities to SARS-CoV and bat coronaviruses⁷. In humans, coronaviruses cause respiratory infections³. Coronaviruses are composed of several proteins including the spike (S), envelope (E), membrane (M), and nucleocapsid (N)⁴. Results suggest that the spike protein retains sufficient affinity to the Angiotensin converting enzyme 2 (ACE2) receptor to use it as a mechanism of cell entry⁵. Human to human transmission of coronaviruses is primarily thought to occur among close contacts via respiratory droplets generated by sneezing and coughing¹. IgM is the first immunoglobulin to be produced in response to an antigen and will be primarily detectable during the early onset of the disease⁵.

ASSAY PRINCIPLE

This ELISA kit is designed, developed, and produced for the qualitative measurement of the COVID-19 IgM antibody in serum. This assay utilizes the "IgM capture" method on microplate based enzyme immunoassay technique.

Assay controls and samples are added to the microtiter wells of a microplate that was coated with a anti-human IgM specific antibody. After the first incubation period, the unbound protein matrix is removed with a subsequent washing step. A horseradish peroxidase (HRP) labeled recombinant COVID-19 antigen is added to each well. After an incubation period, an immunocomplex of "Anti-IgM antibody - human COVID-19 IgM antibody - HRP labeled COVID-19 antigen" is formed if there is novel coronavirus IgM antibody present in the tested materials. The unbound tracer antigen is removed by the subsequent washing step. HRP-labeled COVID-19 antigen tracer bound to the well is then incubated with a substrate solution in a timed reaction and then measured in a spectrophotometric microplate reader. The enzymatic activity of the tracer antigen bound to the coronavirus IgM on the wall of the microtiter well is proportional to the amount of the coronavirus IgM antibody level in the tested materials.

REAGENTS: PREPARATION AND STORAGE

This test kit must be stored at 2 – 8°C upon receipt. For the expiration date of the kit refer to the label on the kit box. All components are stable until this expiration date.

1. COVID-19 IgM Microplate (31223)

Microplate coated with anti-human IgM specific antibody.

Qty: 1 x 96 well microplate
Storage: 2 – 8°C
Preparation: Ready to use.

2. COVID-19 IgM Sample Diluent (31224)

A ready-to-use sample dilution buffer.

Qty: 1 x 15 mL
Storage: 2 – 8°C
Preparation: Ready to use.

3. HRP Labeled COVID-19 Antigen (31226)

HRP labeled COVID-19 Antigen in a stabilized protein matrix.

Qty: 1 x 11 mL
Storage: 2 – 8°C
Preparation: Ready to use.

4. ELISA Wash Concentrate (10010)

Surfactant in a phosphate buffered saline with non-azide preservative.

Qty: 1 x 30 mL
Storage: 2 – 25°C
Preparation: 30X Concentrate. The contents must be diluted with 870 mL distilled water and mixed well before use.

5. ELISA HRP Substrate (10020)

Tetramethylbenzidine (TMB) with stabilized hydrogen peroxide.

Qty: 1 x 15 mL
Storage: 2 – 8°C
Preparation: Ready to use.

6. ELISA Stop Solution (10030)

0.5 M sulfuric acid.

Qty: 1 x 15 mL
Storage: 2 – 25°C
Preparation: Ready to use.

7. COVID-19 IgM Negative Control (31228)

Negative control with a bovine serum albumin based matrix with non-azide preservative. Control products do not contain any serum from patients with new type of coronavirus infection.

Qty: 1 x 1 mL
Storage: 2 – 8°C.
Preparation: Ready to use.

8. COVID-19 IgM Positive Control (31229)

Positive control with a bovine serum albumin based matrix with non-azide preservative. Control products do not contain any serum from patients with new type of coronavirus infection.

Qty: 1 x 0.5 mL
Storage: 2 – 8°C.
Preparation: Ready to use.

SAFETY PRECAUTIONS

The reagents are for in-vitro diagnostic use only. Source material which contains reagents of bovine serum albumin was derived in the contiguous 48 United States. It was obtained only from healthy donor animals maintained under veterinary supervision and found free of contagious diseases. Wear gloves while performing this assay and handle these reagents as if they were potentially infectious. Avoid contact with reagents containing hydrogen peroxide, or sulfuric acid. Do not get in eyes, on skin, or on clothing. Do not ingest or inhale fumes. On contact, flush with copious amounts of water for at least 15 minutes. Use Good Laboratory Practices.

MATERIALS REQUIRED BUT NOT PROVIDED

1. Precision single channel pipettes capable of delivering 20 µL, 25 µL, 100 µL, and 1000 µL, etc.
2. Repeating dispenser suitable for delivering 100 µL.
3. Disposable pipette tips suitable for above volume dispensing.
4. Disposable 12 x 75 mm or 13 x 100 glass tubes.
5. Disposable plastic 1000 mL bottle with caps.
6. Aluminum foil.
7. Deionized or distilled water.
8. Plastic microtiter well cover or polyethylene film.
9. ELISA multichannel wash bottle or automatic (semi-automatic) washing system.
10. Spectrophotometric microplate reader capable of reading absorbance at 450 nm.
11. Incubator capable of holding the temperature at 37 °C.

SAMPLE COLLECTION & STORAGE

Only 20 µL of human serum is required for measurement in duplicate. Samples should only be used on the same day. Severe hemolytic samples should not be used.

ASSAY PROCEDURE

1. Reagent Preparation

1. Prior to use, allow all reagents to come to room temperature. Reagents from different kit lot numbers should not be combined or interchanged.
2. ELISA Wash Concentrate (10010) must be diluted to working solution prior to use. Please see REAGENTS section for details.

2. Assay Procedure

1. Place a sufficient number of microwell strips (31223) in a holder to run controls (31228, 31229) and samples in duplicate.

2. Test Configuration

| Row | Strip 1 | Strip 2 | Strip 3 |
|-----|------------------|----------|-----------|
| A | Negative Control | SAMPLE 3 | SAMPLE 7 |
| B | Negative Control | SAMPLE 3 | SAMPLE 7 |
| C | Negative Control | SAMPLE 4 | SAMPLE 8 |
| D | Positive Control | SAMPLE 4 | SAMPLE 8 |
| E | SAMPLE 1 | SAMPLE 5 | SAMPLE 9 |
| F | SAMPLE 1 | SAMPLE 5 | SAMPLE 9 |
| G | SAMPLE 2 | SAMPLE 6 | SAMPLE 10 |
| H | SAMPLE 2 | SAMPLE 6 | SAMPLE 10 |

3. Add **100 µL** of controls (31228, 31229) into the designated microwells.

4. Add **10 µL** of samples into the designated microwells.
5. Add **100 µL** of COVID-19 IgM Sample Diluent (31224) to the microwells with the samples.
Note: Do not add sample diluent to the wells with the controls!
6. Mix gently and cover the plate with one plate sealer and aluminum foil. Incubate at **37 °C for 30 minutes.**
7. Remove the plate sealer. Aspirate the contents of each well. Wash each well **5 times** by dispensing **350 µL** of diluted wash solution (10010) into each well, and then completely aspirate the contents. Alternatively, an automated microplate washer can be used.
8. Add **100 µL** of the HRP-labeled COVID-19 antigen (31226) into the microwells.
9. Mix gently and cover the plate with one plate sealer and aluminum foil. Incubate at **37 °C for 30 minutes.**
10. Remove the plate sealer. Aspirate the contents of each well. Wash each well **5 times** by dispensing **350 µL** of diluted wash solution (10010) into each well, and then completely aspirate the contents. Alternatively, an automated microplate washer can be used.
11. Add **100 µL** of the substrate (10020) into the microwells.
12. Mix gently and cover the plate with aluminum foil. Incubate at **room temperature (20-25 °C) for 20 minutes.**
13. Remove the aluminum foil and add **100 µL** of stop solution (10030) into each of the microwells. Mix by gently tapping the plate.
14. Read the absorbance at **450 nm** within **10 minutes** with a microplate reader.

PROCEDURAL NOTES

1. It is recommended that all samples be assayed in duplicate. The average absorbance reading of each duplicate should be used for data reduction and the calculation of results.
2. Keep light-sensitive reagents in the original bottles and avoid unnecessary exposure to the light.
3. Store any unused antibody-coated strips in the foil Ziploc bag with desiccant to protect from moisture.
4. Careful technique and use of properly calibrated pipetting devices are necessary to ensure reproducibility of the test.
5. Incubation times or temperatures other than those stated in this insert may affect the results.
6. Avoid air bubbles in the microwell as this could result in lower binding efficiency and higher CV% of duplicate reading.
7. All reagents should be mixed gently and thoroughly prior to use. Avoid foaming.

QUALITY CONTROL

To assure the validity of the results each assay must include both negative and positive controls. The average of the negative control absorbance values less than 0.25 and the positive control absorbance value is not less than 0.50. We also recommend that all assays include the laboratory's own controls in addition to those provided with this kit.

INTERPRETION OF RESULTS

1. Calculate the average value of the absorbance of the negative control (xNC).
2. Calculate the cutoffs using the following formulas:
 - Positive cutoff = 1.1 X (xNC + 0.10)
 - Negative cutoff = 0.9 x (xNC + 0.10)
3. Determine the interpretation of the sample by comparing the OD to the following table:

| Interpretation | Interval | Results |
|----------------|--|--|
| Negative | Measured value \leq negative cutoff | The sample does not contain the new coronavirus (COVID-19) IgM- related antibody |
| Positive | Measured value \geq positive cutoff | The sample contains novel coronavirus (COVID-19) an IgM - associated antibodies. |
| Borderline | Negative cutoff < Measured value < Positive cutoff | Retest the sample in conjunction with other clinical tests. |

EXPECTED VALUES

Samples from the clinical testing presented ODs of 0.164 – 0.661 for the positive values and 0.000 – 0.151 for the negative values. These values should not be in lieu of the interpretation of results calculation.

LIMITATIONS OF THE PROCEDURE

- This test is only for qualitative detection. Test results should not be the sole basis for clinical diagnosis and treatment. The confirmation of infection with novel coronavirus (COVID-19) must be combined with the patient's clinical signs in conjunction to other tests.
- In the first week of the onset or after four weeks of the infection novel coronavirus (COVID-19) patients may be negative for IgM. In addition, patients with low immunity or other diseases that affect immune function, failure of important systemic organs, and use of drugs that suppress immune function can also lead to negative results of new coronavirus IgM.
- Bacterial or fungal contamination of serum specimens or reagents, or cross-contamination between reagents may cause erroneous results.
- Water deionized with polyester resins may inactivate the horseradish peroxidase enzyme.

PERFORMANCE CHARACTERISTICS

Limit of Detection

No international standardized units are available for COVID-19. A positive sample was serially diluted and the limit of detection was determined to be not higher than 5 U/mL.

Repeatability

The assay control is tested in 10 replicates with a CV of OD values less than 15%.

Reproducibility

Three lots were tested with the same samples 10 times with a CV less than 20%.

Class Specificity

This assay does not show any cross reaction to IgG.

Cross-Reactivity

Panels were studied with a minimum of five confirmed disease state samples. No interference was observed for the following disease or infectious agents:

- Anti-influenza A
- Anti-influenza B
- Hepatitis C (HCV)
- Antinuclear Antibodies (ANA)
- Respiratory Syncytial (RSV)

CLINICAL TESTING

Serum samples from two cohorts of patients were tested using the IgM ELISA kit at the Jiaying City Center for Disease Control and Prevention and Zhejiang University Hospital. The combined cohort consisted of normal healthy patients with samples collected prior to the COVID-19 outbreak [December 3, 2019] (n = 54) and RT-PCR confirmed positive patients in after the second week of the onset of the disease (n = 20). The results are as follows:

| | Confirmed Positive | Confirmed Negative |
|-----------------|--------------------|--------------------|
| Test Positive | 9 | 0 |
| Test Negative | 10 | 54 |
| Test Borderline | 1 | 0 |

The diagnostic sensitivity is 45%.
The diagnostic specificity is 100%.
The negative predictive value is 83.1%.
The positive predictive value is 100%.

IgM is the first immunoglobulin to be produced in response to an antigen and will be primarily detectable during the early onset of the disease. The National Health Commission of the People's Republic of China states that IgM antibodies begin to show positive after 3-5 days of onset of COVID-19. Serum samples for the clinical test were from patients after two weeks of the onset of the disease. Therefore, low levels of clinical sensitivity for IgM can be attributed to the collection date of the positive cohort where IgM levels are expected to be lower.

WARRANTY

This product is warranted to perform as described in its labeling and literature when used in accordance with all instructions. Epitope Diagnostics, Inc. DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, and in no event shall Epitope Diagnostics, Inc. be liable for consequential damages. Replacement of the product or refund of the purchase price is the exclusive remedy for the purchaser. This warranty gives you specific legal rights and you may have other rights, which vary from state to state.

REFERENCES

- CDC (2020). Transmission of Novel Coronavirus (COVID-19).
- Chenjia Yuan , Shi Jinsong , Qiudong An , Liu Chang , Li Xin , Qiang , Ruanji Shou , mountains . Wuhan 2019 Bioinformatics coronavirus genome analysis [J / OL]. Bioinformatics : 1-10 [2020-02-10] .
- Fehr, A. R., & Perlman, S. (2015). Coronaviruses: An Overview of Their Replication and Pathogenesis. *Coronaviruses Methods in Molecular Biology*, 1–23. doi: 10.1007/978-1-4939-2438-7_1
- Li, F., Li, W., Farzan, M., & Harrison, S. (2005). Structure of SARS coronavirus spike receptor-binding domain complexed with its receptor. doi: 10.2210/pdb2ajf/pdb
- Wu, L.-P., Wang, N.-C., Chang, Y.-H., Tian, X.-Y., Na, D.-Y., Zhang, L.-Y., ... Liang, G.-D. (2007). Duration of Antibody Responses after Severe Acute Respiratory Syndrome. *Emerging Infectious Diseases*, 13(10), 1562–1564. doi: 10.3201/eid1310.070576
- Xu, X., Chen, P., Wang, J., Feng, J., Zhou, H., Li, X., ... Hao, P. (2020). Evolution of the novel coronavirus from the ongoing Wuhan outbreak and modeling of its spike protein for risk of human transmission. *Science China Life Sciences*. doi: 10.1007/s11427-020-1637-5
- Zhou, P., Yang, X.-L., Wang, X.-G., Hu, B., Zhang, L., Zhang, W., ... Shi, Z.-L. (2020). A pneumonia outbreak associated with a new coronavirus of probable bat origin. *Nature*. doi: 10.1038/s41586-020-2012-7

TECHNICAL ASSISTANCE AND CUSTOMER SERVICE












For technical assistance or place an order, please contact Epitope Diagnostics, Inc. at (858) 693-7877 or fax to (858) 693-7678.

This product is manufactured by
Epitope Diagnostics, Inc.
7110 Carroll Road
San Diego, CA 92121, US

Please visit our website at www.epitopediagnostics.com to learn more about our products and services.

| | | |
|----|-----|--|
| EC | REP | MDSS GmbH Schiffgraben 41, 30175 Hannover, Germany |
|----|-----|--|

GLOSSARY OF SYMBOLS (EN 980/ISO 15223)

| | | |
|--|---|--|
|  In Vitro Diagnostic Device |  European Conformity |  Lot Number |
|  Catalog Number |  Read instructions before use |  Number of Tests |
|  Store at |  Use by |  Keep away from heat and direct sun light |
|  Manufacturer |  Authorized Representative in Europe | |