

ITEM 3
Blood mixer with stirring scale

No	Minimum Technical Requirement		
		Conformity	
		Yes	No
1	Model	Hemo Mix 3	
2	Country of Origin		
3	Name of Manufacture	Delcon	
4	For shaking and weighing blood units during collection	Yes	
5	Equipped with a screen that can view the selected parameters	Yes	
6	Provided with safety clips and tube detectors	Yes	
7	Provided with multicolored light indicator, with the possibility of remote viewing	Yes	
8	Provided with audio and visible alarms	Yes	
9	Provided with warning system in case of deviations from the scheduled regimen during blood collection.	Yes	
10	Provided with warning system when the scheduled dose is reached	Yes	
11	Automatic weighing of the empty container at the beginning of the blood collection	Yes	
12	Automatic weighing during the blood collection	Yes	
13	Automatic weighing when the scheduled dose is reached.	Yes	
14	Possibility to adjust the volume during blood collection	Yes	
15	Orbital agitation of the blood unit (container), with programmable movement, pauses and reversal	Yes	
16	Possibility to store data with data export on a common or analog SD card	Yes	
17	Barcode reader	Yes	
18	The unit of measurement of the scheduled dose should be possible to indicate in ml and grams	Yes	
19	Weight measuring capacity to be from 0 to 1000ml and/or grams	Yes	
20	Power grid parameters: 220 V, 50 Hz	Yes	
21	EC / MS Declaration of Conformity and / or EC / MS Certificate of Conformity	Yes	
22	Copies of the service manual, service manual, electronic diagram, technical passports, including the description of the technical parts and components of the device	Yes	
23	Year of production of the device 2021-2022	Yes	
24	Warranty period of not less than 36 months	Yes	

[illegible]