

## **EU DECLARATION OF CONFORMITY**

MANUFACTURER

Warwick SASCo Ltd. Warwick House, Heathcote Way, Warwick. CV34 6TE.

United Kingdom.

SRN: GB-MF-000008346

AUTHORISED REPRESENTATIVE

Medical Device Management Ltd. Block B, The Crescent Building, Northwood, Santry, Dublin 9, D09 C6X8. Ireland.

SRN: IE-AR-000002496

## Declares that the products described hereafter as:

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Instrument tray, reusable	CIT1814, CIT2013, CIT2718, IT1813, IT1913, IT2015, IT2715, IT3025, IT4030, LCT2215, T2015, T3025	Class I	506009813AAA001NF

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Sterilization/ Disinfection container, accessory, reusable	DINT2525, DINT4825, MIT1515, MIT2015, MIT2709, MIT2715, MIT2817, MIT3025, MIT4030, PIT3025, PIT4030, SMT5510	Class I (Accessory)	506009813AAA002NH
Sterilization/ Disinfection container, reusable	DSS1000, DSSW1000, DSS2711, DSSV2711, DSSVW2711, DSSW2711, DSS4215, DBS2718BSRPL1	Class I	506009813AAA002NH
Disinfection container, reusable	DSS8515, DSS8515TP	Class I	506009813AAA002NH

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Kidney dish, reusable	KD150, KD200, KD250, KD300, KDT185, KDY250, VB300	Class I	506009813AAA003NK

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
General-purpose bowl, reusable	GPLB100, GPLB150, SB140, LB100, LB150, LB200, LB250, WB3513, WB3112	Class I	506009813AAA004NM
Washbowl, reusable	WBN3112, WBN3513	Class I	506009813AAA004NM









Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Sterilization container accessory mat, reusable	SP2313, SP2324, SP2719, SP2823, SP4622, SP4724, SR10	Class I (Accessory)	506009813AAA005NP

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Drinking cup, reusable	BFH250	Class I	506009813AAA007NT
Adult drinking/feeding cup, reusable	BMS, BNS, BWS, BRF, GPMC6250, HDMS, NS, TFB, WS	Class I	506009813AAA007NT
Adult drinking/feeding cup accessory, reusable	BC.RUSSKA, BC250, BFC250	Class I (Accessory)	506009813AAA007NT

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Bedpan, reusable	BPN, DPNF, FPN, FPNW, FP, FPW, SP, SPN, VP, VPN	Class I	506009813AAA008NV
Bedpan accessory support, reusable	PBPS, PSPS	Class I (Accessory)	506009813AAA008NV

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Hand-held urinal, reusable	PUF1000, UF1000, UM1000, UMA1000, UML1000	Class I	506009813AAA009NX

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Urinal/non leak valve, reusable	USP1200	Class I	506009813AAA016NU
Non leak valve accessory, reusable	UMT	Class I (Accessory)	506009813AAA016NU







Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Kidney dish, single use	KB200, KB250, KB300	Class I	506009813AAA010NG

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Instrument tray, single use	LCT2215S, ST4330	Class I	506009813AAA011NJ

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Surgical instrument containment holder, single use	QV30	Class I	506009813AAA012NL
Surgical instrument containment holder accessory, single use	QV47	Class I (Accessory)	506009813AAA012NL

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Sponge bowl, single use	SB170, SB200	Class I	506009813AAA013NN
General-purpose bowl, single use	TB350	Class I	506009813AAA013NN

Description/ Intended Use	Product Codes	Class (MDR)	Basic UDI-DI
Jug, reusable	JG500, JG1000, JG2000	Class I	506009813AAA014NQ

Are in conformity with Regulation (EU) 2017/745 and classified in respect to Annex VIII Chapter III Rule 1.

This declaration is supported by the Quality System approval to ISO 13485:2016 issued by SGS United Kingdom Ltd.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

The products described above are manufactured in the United Kingdom.

Signed by:

Paul Yeomans.

Quality and Compliance Manager.

22 July 2024.

Doc. 3.35 Rev. 07/24-0

For and on behalf of Warwick SASCo Ltd.





