

NADAL® C. difficile Toxin A/B Test (test cassette)

REF 582008



DE	Gebrauchsanweisung	2	PT	Instruções de Utilização	17
EN	Instructions for use	5	CZ	Návod k použití	19
FR	Instructions d'utilisation	7	SE	Användarinstruktioner	21
ES	Instrucciones de uso	10	NO	Bruksanvisning	23
IT	Istruzioni per l'uso	12		Symbols	27
PL	Sposób użycia	14		Our Teams	28



1. Intended Use

The NADAL® C. difficile Toxin A/B Test is a rapid chromatographic immunoassay for the qualitative detection of C. difficile Toxin A&B antigens in human stool specimens to aid in the diagnosis of C. difficile infection.

2. Introduction and Clinical Significance

Clostridium difficile is an anaerobic, gram-positive, spore-forming bacterium. The key feature that enables it to persist in patients and physical environments for long periods of time and facilitates its transmission is the ability of C. difficile to form spores. C. difficile is transmitted through the fecal-oral route.

Clostridium difficile is the principal pathogen related to antibiotic associated diarrhea and/or pseudomembranous colitis in hospitalized patients.

Mature colonic bacterial flora in a healthy adult is generally resistant to C. difficile colonization. However, if the normal colonic flora is altered, resistance to colonization is lost. Thus, any factor associated with alteration of the normal enteric flora increases the risk of C. difficile colonization after exposure to antibiotics, especially those with broad-spectrum activity such as penicillins, cephalosporins and clindamycin.

C. difficile can release two high-molecular-weight toxins, toxin A and toxin B. These are responsible for the clinical manifestations, which range from mild, self-limited watery diarrhea to fulminant pseudomembranous colitis, toxic megacolon, and death.

3. Test Principle

The NADAL® C. difficile Toxin A/B Test is a qualitative immunoassay for the detection of C. difficile Toxin A and Toxin B antigens in human stool samples. The membrane is pre-coated with antibodies against Toxin A and antibodies against Toxin B antigens on the corresponding test line regions. During testing, the sample binds to the anti-Toxin A and anti-Toxin B antibodies on the conjugate pad. The mixture moves upward on the membrane by capillary action. In the case of a positive result the specific antibodies present in the test line regions will react with the mixture conjugates and generate one or two coloured test lines. A green coloured line (third line) should always develop in the control line region. It serves as an internal test control and as verification that sufficient volume of specimen has been added and proper membrane wicking has occurred.

4. Reagents and Materials Supplied

- 10 NADAL® C. difficile Toxin A/B test cassettes
- 10 specimen collection vials with buffer
- Package insert

5. Additional Materials Required

- Specimen collection container
- Disposable gloves
- Timer

6. Storage & Stability

Store as packaged in the sealed pouch either refrigerated or at room temperature (2-30°C). The test is stable through to the expiration date printed on the sealed pouch. The test must remain in the sealed pouch until use. Do not freeze.

7. Warnings and Precautions

- For professional *in-vitro* diagnostic use only.
- Do not use the test beyond the expiration date.
- The test should remain in the sealed pouch until use.
- Do not use the test if the pouch is damaged.
- All operations linked to the use of the test must be performed in accordance with *Good Laboratory Practices (GLP)*.
- Wear protective clothing such as laboratory coats, disposable gloves and eye protection when specimens are assayed.
- Do not eat, drink or smoke in the area where the specimens and kit reagents are handled.
- All the specimens should be considered potentially hazardous and handled in the same manner as an infectious agent.
- The test should be discarded in a proper biohazard container after testing.
- The test must be carried out within 2 hours after opening the sealed pouch.


8. Specimen Collection and Preparation

Collect sufficient quantity of feces (1-2 g or 1-2 mL for liquid sample). Stool samples should be collected in clean and dry containers (no preservatives or transport media). The samples can be stored in the refrigerator (2-8°C) for 24-48 hours prior to testing. For longer storage the specimen must be kept frozen at -20°C. In this case, the sample should be completely thawed, and brought to room temperature before testing.

9. Test Procedure

Specimen preparation:

Use a separate specimen collection vial for each sample. Unscrew the cap of the vial and introduce the stick into three different spots of the fecal specimen to collect 125 mg of sample.



Close the vial containing the buffer and stool sample.


Shake the vial in order to ensure good sample dispersion.

For liquid stool samples, aspirate the fecal specimen with a dropper and add 125 µL into the specimen collection vial with buffer.


Test Procedure

Bring the test cassettes, stool samples and buffer to room temperature (15-30°C) prior to testing. Do not open pouches until ready to perform the assay.

1. Remove the NADAL® C. difficile Toxin A/B test cassette from its sealed pouch and use it as soon as possible.
2. Shake the specimen collection vial to ensure good sample dispersion. Break off the tip of the vial.



- Use a separate test cassette for each sample. Dispense 4 drops of the sample into the specimen well (S). Start the timer.
- Read the result at 10 minutes after dispensing the sample.



- An excess of sample may cause wrong results (brown bands appear). Dilute the sample with the buffer and repeat the test.
- Some stool samples can decrease the intensity of the control bands.
- The test must be carried out within 2 hours after opening the sealed pouch.
- If the test result is negative and clinical symptoms persist, additional testing using other clinical methods is recommended. A negative result does not at any time preclude the possibility of *Clostridium difficile* infection.
- The NADAL® C. difficile Toxin A/B Test provides a presumptive diagnosis of infection caused by *Clostridium difficile*. All results must be interpreted together with other clinical information and laboratory findings available to the physician.

10. Result Interpretation

Positive:

Toxin A positive:

In addition to the green line in the control region (C), a red-coloured line will develop in the test region (T1).



Toxin B positive:

In addition to the green line in the control region (C), a red-coloured line will develop in the test region (T2).



Toxin A/B positive:

In addition to the green line in the control region (C), two red-coloured lines develop in the test region (T1) and in test region (T2).



13. Expected values

Clostridium difficile is associated with 95-100% of the cases of pseudomembranous colitis, 60-75% of the cases of antibiotic-associated colitis and 35% of the cases of antibiotic-associated diarrhea.

14. Performance Characteristics

Sensitivity and specificity

Stool samples from patients with diarrhea were studied. The NADAL® C. difficile Toxin A/B Test in comparison with other commercial immunoassay tests showed:

Sensitivity >99%

Specificity >99%

Cross reactivity:

An evaluation was performed to determine the cross reactivity of the NADAL® C. difficile Toxin A/B Test. There is no cross-reactivity with common gastrointestinal microorganisms occasionally present in feces.

<i>Campylobacter</i> spp.	<i>Helicobacter pylori</i>	<i>Salmonella</i> spp.
<i>E. coli</i> O157:H7	<i>Shigella</i> spp.	<i>Yersinia</i> spp.
<i>Listeria monocytogenes</i>	<i>Staphylococcus aureus</i>	<i>Yersinia enterocolitica</i>


15. References

- Wren, M.W.D, et al. "Laboratory diagnosis of *Clostridium difficile* infection. An evaluation of tests for faecal toxin, glutamate dehydrogenase, lactoferrin and toxigenic culture in the diagnostic laboratory". British Journal of Biomedical Science, 66 (1), 2009.
- Vaishnavi, Ch., "Clinical spectrum & pathogenesis of *Clostridium difficile* associated diseases". Indian J. Med. Res. 131, April 2010, pp 487-499.
- Poutanen, S. M. et al. "*Clostridium difficile*-associated diarrhoea in adults", CMAJ, 171(1) July 2004, pp. 51-58.

Rev. 2, 2014-10-24 JB

Negative:

Only one green line develops in the control region C. No test lines develop.



Invalid:

No line develops in the control region (C). This indicates a possible error in the performance of the test. A new test should be performed.



Note: Insufficient specimen volume, incorrect procedural techniques or reagent deterioration are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test. If the problem persists, discontinue using the test kit and contact your distributor.

11. Quality Control


An internal procedural control is included in the test:











A green line developing in the control line region (C) confirms sufficient specimen volume and correct procedural technique.



12. Limitations

- The NADAL® C. difficile Toxin A/B Test will only detect the presence of parasites in the stool specimen (qualitative detection) and should be used for the detection of Toxins A and/or B antigens in feces specimens only. Neither the quantity nor the rate of increase in antigen concentration can be determined by this test.

Symbol	Deutsch	English	Français	Español	Italiano	Polski
	CE Konformitätszeichen	CE marking of conformity	Conforme aux normes européennes	Conformidad europea	Conformità europea	Znak zgodności CE
	Gebrauchsanweisung beachten	Consult instructions for use	Consulter la notice d'utilisation	Consúltense las instrucciones de uso	Consultare le istruzioni per l'uso	Przestrzegać instrukcji obsługi
	<i>in-vitro</i> -Diagnostika	<i>in-vitro</i> diagnostic medical device	Dispositif médical de diagnostic <i>in-vitro</i>	Producto sanitario para diagnóstico <i>in-vitro</i>	Dispositivo medico-diagnostico <i>in-vitro</i>	Tylko do diagnostyki <i>in-vitro</i>
	Temperaturbegrenzung	Temperature limitation	Limites de température	Límite de temperatura	Limiti di temperatura	Temperatura przechowywania
	Chargenbezeichnung	Batch code	Numéro de lot	Código de lote	Codice lotto	Numer serii
	Nicht zur Wiederverwendung	Do not reuse	Ne pas réutiliser	No reutilizar	Non riutilizzare	Tylko do jednorazowego użytku
	Verwendbar bis	Use by	Utiliser jusqu'au	Fecha de caducidad	Utilizzare entro	Data ważności
	Bestellnummer	Catalogue Number	Référence du catalogue	Número de catálogo	Riferimento di Catalogo	Numer katalogowy
	Hersteller	Manufacturer	Fabricant	Fabricante	Fabbricante	Producent
	Ausreichend für <n> Ansätze	Sufficient for <n> tests	Suffisant pour "n" tests	Suficiente para <n> utilizations	Sufficiente per "n" saggi	Wystarczający na <n> Powtórzeń

Symbol	Português	Český	Suomi	Svenskt	Nederlands	Dansk	Norsk
	Conformidade com as normas europeias	CE certifikát	CE-merkitty	CE-märkning	CE-markering	CE-mærkning	CE standardisert
	Consultar as instruções de utilização	Viz návod k použití	Katso käyttöohjetta	Läs bruksanvisningen	Raadpleeg de gebruiksaanwijzing	Se brugsanvisningen	Les bruksanvisning nøye
	Dispositivo médico para diagnóstico <i>in-vitro</i>	Diagnostický zdravotnický prostředek <i>in-vitro</i>	<i>in-vitro</i> - diagnostiikkaan tarkoitettu lääkinnällinen laite	Medicinteknisk produkt avsedd för <i>in-vitro</i> -diagnostik	Medisch hulpmiddel voor <i>in-vitro</i> diagnostiek	Medicinsk udstyr til <i>in-vitro</i> -diagnostik	<i>in-vitro</i> diagnostic medisinsk enhet
	Limites de temperatura	Teplotní omezení	Lämpötilarajat	Temperaturbegränsning	Temperatuurlimiet	Temperaturbegrænsning	Temperaturbegrænsning
	Código do lote	Kód šarže	Eräkoodi	Satsnummer	Code van de partij	Batchkode	Merking
	Não reutilizar	Pro jednorázové použití	Kertäkäyttöinen	Får inte återanvändas	Niet opnieuw gebruiken	Må ikke genbruges	Må ikke brukes om igjen
	Prazo de validade	Spotřebuje do	Käytettävä viimeistään	Används före	Houdbaar tot	Udløbsdato	Tidtaking
	Número de catálogo	Katalogov číslo	Luettelonumero	Listnummer	Catalogus nummer	Best il l ingsnummer	Katalog nummer
	Fabricante	Výrobce	Valmistaja	Tillverkare	Fabrikant	Fabrikant	Produsent
	Suficiente para <n> test	Dostačuje pro <n> testů	Lukumäärä <n> test	Räcker till <n> test	Voldoende voor <n> test	Tilstrækkeligt til <n> test	Tilstrækkelig for <n> tester

Our Teams

Germany:

Regensburg

Tel: +49 941 290 10-0
 Fax: +49 941 290 10-50

Moers

Tel: +49 2841 99820-0
 Fax: +49 2841 99820-1

Austria:

Tel: +49 941 290 10-29
 Free Tel: 0800 291 565
 Fax: +49 290 10-50
 Free Fax: 0800 298 197

UK & Ireland:

Tel: +49 941 290 10-18
 Free Tel – UK: 0808 234 1237
 Free Tel – IRE: 1800 555 080
 Fax: +49 290 10-50

France:

France Tel: 0800 915 240
 France Fax: 0800 909 493

Switzerland

Swiss Tel: 0800 564 720
 Swiss Fax: 0800 837 476

Belgium

Belgium Tel: 0800 718 82
 Belgium Fax: 0800 747 07

Luxembourg

Lux. Tel: 800 211 16
 Lux. Fax: 800 261 79

Spain:

Tel: +49 941 290 10-759
 Free Tel: 900 938 315
 Fax: +49 941 290 10-50
 Free Fax: 900 984 992

Italy:

Tel: +49 941 290 10-34
 Fax: +49 941 290 10-50

Poland:

Tel: +49 941 290 10-44
 Free Tel: 00 800 491 15 95
 Fax: +49 941 290 10-50
 Free Fax: 00 800 491 15 94

Portugal:

Tel: +49 941 290 10-735
 Tel. Verde: 800 849 230
 Fax: +49 941 290 10-50
 Fax Verde: 800 849 229

Netherlands:

Tel: +31 30 75 600
 Free Tel: 0800 0222 890
 Fax: +31 70 30 30 775
 Free Fax: 0800 024 9519

Nordic countries (Finland, Norway, Sweden, Denmark):

Tel: +31 703075 607
 Free Tel: +45 80 88 87 53
 Tax: +31 703030 775

Laboratory Diagnostics Team:

Tel: +49 941 290 10-40
 Fax: +49 941 290 10-50



nal von minden GmbH

Carl-Zeiss-Strasse 12 • 47445 Moers • Germany

www.nal-vonminden.com • info@nal-vonminden.com

Fon: +49 2841 99820-0 • Fax: +49 2841 99820-1