



Certificate of Analysis

Product Name: HCV Rapid Test Cassette(Serum/ Plasma)	
Catalog No.: IHCV-C31	
Product Lot No.: HCV23020001	Batch Quantity: 3500PCS
Mfg Date: 2023.02.02	Expiry Date: 2025.02.01
Strip Size (mm): 60*4	Inspection Date: 2023.04.23
Buffer Lot. /Exp.: 23010086/2025.04.05	Remark: /

QC Items		Specification	Test Result	Conclusion
Physical	Appearance	Intact	Intact	PASS
Functional Performance	Positive Sample	100% Positive	100% Positive	PASS
	Negative Sample	100% Negative	100% Negative	PASS

Conclusion: This batch of product meets the QC specifications.

Arno Wang

QC: _____

Arno Wang

Title: Quality Control Supervisor

Date: _____

2023.04.23

