

ERMED TIP MEDİKAL TİC. A.Ş.	GAUZE BANDAGE NON-STERILE (14cmx7mt) ANALYSIS REPORT	DATE: 05.02.2022 BANDAGE
-----------------------------------	---	---------------------------------

MANUFACTURER NAME: ERMED TIP MEDİKAL TİCARET A.Ş.
MANUFACTURER ADD.: Köseköy Mah. Yılmaz Sok. No:15 Kartepe-Kocaeli/TÜRKİYE

APPLIED STANDARD:TS EN ISO 14079

INSPECTION AND EXPERIMENTS	REQUESTED	EXPERIMENT RESULTS
MATERIAL FEATURE	100% COTTON	APPROPRIATE
APPEARANCE FEATURE	NON-DEFECTIVE	APPROPRIATE
DIMENSION	14CMX7MT	APPROPRIATE
DENSITY	32g/m²	APPROPRIATE
SINKING TIME	9 s	APPROPRIATE
RUPTURE STRENGTH	5 kgf/5 cm (min)	APPROPRIATE
SYNTHETIC FIBER	0.1% (max)	APPROPRIATE
pH	4.5-8.6	APPROPRIATE
ETHER SOLUBLE SUBSTANCE	0.50% (max)	APPROPRIATE
SURFACE ACTIVE SUBSTANCE	Single row foam (max)	APPROPRIATE
SULFATE ASH	0.75% (max)	APPROPRIATE
WATER SOLUBLE MATTER	0.50% (max)	APPROPRIATE
STARCH AND DEXSTRIN	Should Appear Colorless	APPROPRIATE

APPROVED BY :

ERMED TIP MEDİKAL TİCARET A.Ş.
Köseköy Mahallesi Yılmaz Sok. No:15/1
Kartepe / KOCAELİ Tel: 0262 223 44 80
Alemdar V.D. 3640490515
Mersis No: 0364049051500001

LOT: 210620
REF: EASY411418
MFG. DATE: 02/2022
EXP. DATE: 02/2027