

Surface disinfection

Velox Top AF

Spray for surface disinfection and cleaning



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 1276	30 sec.	60 sec.
	EN 13697	30 sec.	60 sec.
	EN 13727	30 sec.	60 sec.
Yeast (<i>C. albicans</i>)	EN 1650	30 sec.	60 sec.
	EN 13624	30 sec.	60 sec.
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.	60 sec.
Enveloped viruses	DVV/RKI	30 sec.	30 sec.
<i>BVDV</i>	DVV/RKI	30 sec.	
<i>Adenovirus</i>	EN 14476	60 sec.	
<i>Rotavirus</i>	EN 14476	30 sec.	

Velox Spray

Spray for surface disinfection



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 13697	30 sec.	60 sec.
	EN 13727	30 sec.	30 sec.
	EN 16615	30 sec.	30 sec.
	EN 17387		5 min.
Yeast (<i>C. albicans</i>)	EN 13624	15 sec.	15 sec.
	EN 13697	30 sec.	
	EN 16615	30 sec.	30 sec.
	EN 17387		5 min.
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.	30 sec.
Mycobacteria (<i>M. avium</i>, <i>M. terrae</i>)	EN 14348	30 sec.	60 sec.
Enveloped viruses	EN 14476	30 sec.	60 sec.
<i>BVDV</i>	EN 14476	30 sec.	30 sec.
<i>Norovirus</i>	EN 14476	30 sec.	30 sec.
<i>Rotavirus</i>	EN 14476	30 sec.	60 sec.

Velox Foam Extra

Foam for cleaning and quick disinfection of sensitive surfaces of non-invasive medical devices



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 13727	60 sec.	60 sec.
	EN 16615	60 sec.	3 min.
	EN 13697	30 sec.	60 sec.
Yeast (<i>C. albicans</i>)	EN 13624	60 sec.	60 sec.
	EN 16615	60 sec.	60 sec.
	EN 13697	30 sec.	
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	5 min.	10 min.
Mycobacteria tuberculosis (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	15 min.	15 min.
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	60 sec.	60 sec.
<i>BVDV</i>	EN 14476	60 sec.	60 sec.
Rotavirus	EN 14476	15 min.	

Velox Wipes

Alcohol wipes for cleaning and disinfecting small surfaces of non-invasive medical equipment



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 1276	30 sec.	60 sec.
	EN 13697	30 sec.	60 sec.
	EN 13727	30 sec.	60 sec.
	EN 16615	30 sec.	30 sec.
Yeast (<i>C. albicans</i>)	EN 1650	30 sec.	60 sec.
	EN 13624	30 sec.	60 sec.
	EN 16615	30 sec.	30 sec.
Mycobacterium tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.	60 sec.
Enveloped viruses	DVV/RKI	30 sec.	30 sec.
<i>Adenovirus</i>	EN 14476	60 sec.	
<i>Rotavirus</i>	EN 14476	30 sec.	
<i>BVDV</i>	DVV/RKI	30 sec.	

Velox Wipes NA

Alcohol-free wipes for disinfection small surfaces and medical equipment.



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 13727	60 sec.	60 sec.
	EN 16615	60 sec.	3 min
	EN 13697	30 sec.	60 sec.
Yeast (<i>C. albicans</i>)	EN 13624	60 sec.	60 sec.
	EN 16615	60 sec.	60 sec.
	EN 13697	30 sec.	
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	5 min	10 min
Mycobacteria (<i>M. avium</i>, <i>M. terrae</i>)	EN 14348	15 min	15 min
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	60 sec.	60 sec.
BVDV	EN 14476	60 sec.	60 sec.

Velox Duo Wipes

Alcoholic wipes for cleaning and rapid disinfection of small surfaces and medical equipment.



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 13697	30 sec.	60 sec.
	EN 13727	30 sec.	30 sec.
	EN 16615	30 sec.	30 sec.
Yeast (<i>C. albicans</i>)	EN 13624	15 sec.	15 sec.
	EN 13697	30 sec.	
	EN 16615	30 sec.	30 sec.
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.	30 sec.
Enveloped viruses	EN 14476	30 sec.	60 sec.
Norovirus	EN 14476	30 sec.	30 sec.
Rotavirus	EN 14476	30 sec.	60 sec.
BVDV	EN 14476	30 sec.	30 sec.

Quatrodex Forte

Concentrate for surface disinfection and cleaning



Spectrum (effectiveness)	EN 14885	Time	
		15 min	30 min
Bacteria	EN 13727	0,5%*, 1%	
	EN 14561	1%	0,5%
	EN 13697	1%*	
Yeast (<i>C. albicans</i>)	EN 13624	0,5%	
	EN 14562	1%	0,5%
	EN 13697	1%*	
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	1%*, 4%	1%
	EN 14563	4%	1%
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	0,5%	
Adenovirus	EN 14476	1%	
Poliovirus	EN 14476	4%	
Norovirus	EN 14476	4%	
BVDV	EN 14476	0,5%	

* clean conditions

Quatrodex Extra

Concentrate for surface disinfection and cleaning



Spectrum (effectiveness)	EN 14885	Time		
		15 min	30 min	60 min
Bacteria	EN 13727	0,25%*, 0,5%		
	EN 14561	1%	0,5%	
	EN 16615	1%		
Bacteria (<i>P. aeruginosa</i>)	EN 13697	0,5%		
Yeast (<i>C. albicans</i>)	EN 13624	0,25%*, 0,5%		
	EN 13697	0,5%		
	EN 14562	1%	0,5%	
	EN 16615	1%		
Fungi (<i>A. brasiliensis</i>)	EN 13624			1,5%
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	0,5%*, 1%	0,5%	
	EN 14563	1%*, 2%	0,5%	
Enveloped viruses (<i>BVDV</i>)	EN 14476	0,5%		
Adenovirus	EN 14476	0,5%*, 1%	0,5%	
Norovirus	EN 14476	4%	4%	2%
Poliovirus	EN 14476	1%	0,5%*	
Rotavirus	EN 14476	1%		

* clean conditions

Current research for Medisept products

status for the day 21.02.2024

Quatrosdes Unit NF

Concentrate for suction devices disinfection and cleaning



Spectrum (effectiveness)	EN 14885	Time	
		15 min	30 min
Bacteria	EN 13727	0,5%*, 1%	
Yeast (<i>C. albicans</i>)	EN 13624	0,5%	
Mycobacteria tuberculosis (<i>M. tuberculosis</i>)	EN 14348		4%
Enveloped viruses	EN 14476	2%	

* clean conditions

Instrument disinfection

Viruton Forte

Concentrate for cleaning and disinfection of surgical instruments



Spectrum (effectiveness)	EN 14885	Time	
		15 min	30 min
Bacteria	EN 13727	0,5%*, 1%	
	EN 14561	1%	0,5%
	EN 13697	1%*	
Yeast (<i>C. albicans</i>)	EN 13624	0,5%	
	EN 14562	1%	0,5%
	EN 13697	1%*	
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	4%	1%
	EN 14563	4%	1%
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	0,5%	
BVDV	EN 14476	0,5%	
Adenovirus	EN 14476	1%	
Norovirus	EN 14476	4%	
Poliovirus	EN 14476	4%	

* clean conditions

Viruton Extra

Concentrate for cleaning and disinfection surgical instruments



Spectrum (effectiveness)	EN 14885	Time			
		15 min	30 min	60 min	120 min
Bacteria	EN 13727	0,5%			
	EN 14561	1%	0,5%		
	EN 13624	0,5%			
Fungi (<i>A. brasiliensis</i>)	EN 13624			1,5%	
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	0,5%*, 1%	0,5%		
	EN 14563	1%*, 2%	0,5%		
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	0,5%			
Enveloped viruses (<i>BVDV</i>)	EN 17111	0,5%			
Adenovirus	EN 14476	0,5%*	0,5%		
	EN 17111				4%
Norovirus	EN 14476	4%		2%	
	EN 17111				4%
Poliovirus	EN 14476	1%	0,5%*		

* clean conditions

Viruton Pre

Pre-disinfecting and cleaning RTU foam for surgical instruments



Spectrum (effectiveness)	EN 14885	Time
Bacteria	EN 13697	15 min*
	EN 13727	15 min
	EN 14561	15 min
Yeast (<i>C. albicans</i>)	EN 13624	15 min
	EN 13697	15 min*
	EN 14562	15 min
Mycobacteria (<i>M. avium</i>, <i>M. terrae</i>)	EN 14348	15 min
	EN 14563	15 min
BVDV	EN 14476	15 min
Adenovirus	EN 14476	15 min
Poliovirus	EN 14476	15 min

* clean conditions

Viruton Pulver

Powder for cleaning and disinfecting tools, endoscopes, incubators and other surfaces of medical devices



Spectrum (effectiveness)	EN 14885	Time			
		10 min	15 min	30 min	60 min
Bacteria (<i>S. aureus</i> , <i>P. aeruginosa</i> , <i>E. hirae</i>)	EN 13727		0,5%		
	EN 14561	1%		0,5%	
	EN 13697			0,5%	
Yeast (<i>C.albicans</i>)	EN 14562	1%		0,5%	
	EN 13624			0,5%	0,5%
Fungi (<i>A. brasiliensis</i>)	EN 13624	1%			0,5%
	EN 14562	2%	1,5%	1%	
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348		0,5%		
	EN 14563				0,5%
Mycobacteria (<i>M.avium</i> , <i>M.terrae</i>)	EN 14348		0,5%		
	EN 14563	1%			
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	2%		1%	
	EN 17111	0,5%			
<i>Poliovirus</i> , <i>Adenovirs</i> , <i>Norovirus</i>	EN 14476	2%		1%	
<i>Adenovirs</i> , <i>Norovirus</i>	EN 17111		0,5%		
Spores (<i>Cl. difficile</i>)	EN 17126	1%			
	EN 13704	1%, 1%*	2%	0,5%	
Spores (<i>Cl difficile</i> R027)	EN 17126	1%*, 2%			
Spores (<i>Cl. sporogenes</i>)	EN 13704	2%			
Spores (<i>B. subtilis</i>)	EN 13704	1%*		0,5%	
	EN 17126	4%		2%	2%
Spores (<i>B. cereus</i>)	EN 17126	4%		2%	2%

* clean conditions

Viruton Bohr

Ready-to-use formula for cleaning and disinfection of rotary instruments



Spectrum (effectiveness)	EN 14885	Time
Bacteria	EN 13727	15 min
	EN 14561	15 min
Yeast (<i>C.albicans</i>)	EN 14562	15 min
	EN 13624	15 min
Mycobacteria (<i>M. terrae</i> , <i>M. avium</i>)	EN 14563	15 min
	EN 14348	15 min
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	15 min
<i>BVDV</i>	EN 14476	15 min
<i>Adenovirus</i>	EN 14476	15 min
<i>Poliovirus</i>	EN 14476	15 min

Hand disinfection and hand care

MEDISEPT wipes for hand and surface disinfection

High-quality alcohol wipes for disinfection hands and any small surfaces resistant to alcohol.



Spectrum (effectiveness)	EN 14885	Time
Bacteria	EN 13727	30 sec.*; 60 sec.
	EN 16615	30 sec.
Yeast (<i>C.albicans</i>)	EN 1275	30 sec.*
	EN 16615	30 sec.
	EN 13624	30 sec.*
	EN 14348	30 sec.*
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.*
Enveloped viruses	EN 14476	30 sec.*
	EN 14476	30 sec.*
<i>BVDV</i>	EN 14476	30 sec.*
<i>Rotavirus</i>	EN 14476	45 sec.*
Hygienic hand disinfection	EN 1500	2 x 1,5 ml ic 2 x 15 sec.
Surgical hand disinfection	EN 12791	2 x 1,5 ml ic 90 sec.

* clean conditions

Velodes Soft

Liquid for hygienic and surgical hand disinfection



Spectrum (effectiveness)	EN 14885	Time
Bacteria	EN 13727	30 sec.*; 60 sec.
Fungi (<i>C. albicans</i>)	EN 1275	30 sec.*
	EN 13624	30 sec.*
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.*
Enveloped viruses (<i>Vaccinia</i>)	EN 14476	30 sec.*
<i>BVDV</i>	EN 14476	30 sec.*
<i>Rotavirus</i>	EN 14476	45 sec.*
Hygienic hand disinfection	EN 1500	2 x 1,5 ml in time 2 x 15 sec.
Surgical hand disinfection	EN 12791	2 x 1,5 ml in time 90 sek. + Prolonged effectiveness of surgical

* clean conditions

Current research for Medisept products

status for the day 21.02.2024

Velodes Silk

Liquid for hygienic and surgical hand disinfection and for surface disinfection



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 13727	30 sec.	30 sec.
Bacteria	EN 13697	30 sec.	60 sec.
Yeast (<i>C. albicans</i>)	EN 13624	15 sec.	15 sec.
Yeast (<i>C. albicans</i>)	EN 13697	30 sec.	
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.	30 sec.
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	30 sec.	60 sec.
Enveloped viruses (Vaccinia)	EN 14476	30 sec.	60 sec.
Norovirus	EN 14476	30 sec.	30 sec.
Rotavirus	EN 14476	30 sec.	60 sec.
BVDV	EN 14476	30 sec.	30 sec.
Hygienic hand disinfection	EN 1500	2 x 1,5 ml in time 2 x 15 sec.	
Surgical hand disinfection	EN 12791	2 x 1,5 ml in time 2 x 90 sec.	

Velodes Gel

Gel for hygienic and surgical hand disinfection



Spectrum (effectiveness)	EN 14885	Test conditions	
		Clean	Dirty
Bacteria	EN 13727	30 sec.	30 sec.
Fungi (<i>C. albicans</i>)	EN 13624	15 sec.	15 sec.
Mycobacteria tuberculosis (<i>M. terrae</i>)	EN 14348	30 sec.	30 sec.
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	30 sec.	60 sec.
Enveloped viruses (Vaccinia)	EN 14476	30 sec.	60 sec.
BVDV	EN 14476	30 sec.	30 sec.
Rotavirus	EN 14476	30 sec.	60 sec.
Norovirus	EN 14476	30 sec.	30 sec.
Hygienic hand disinfection	EN 1500	2 x 1,5 ml in time 2 x 15 sec.	
Surgical hand disinfection	EN 12791	2 x 1,5 ml in time 2 x 90 sec.	

NEW PRODUCTS

Velodes Skin

Liquid for hygienic and surgical hand disinfection



Spectrum (effectiveness)	EN 14885	Time
Bacteria	EN 13727	15 sek.
Yeast (<i>C. albicans</i>)	EN 13624	30 sek.
	EN 13697	3 min.
Fungi (<i>A. brasiliensis</i>)	EN 13624	30 sek.
	EN 13697	60 sek.
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	30 sek.
Enveloped viruses (Vaccinia)	EN 14476	15 sek.
Adenovirus	EN 14476	15 sek.
Norovirus	EN 14476	15 sek.
Poliovirus	EN 14476	30 sek.
Hygienic hand disinfection	EN 1500	3 ml in time 30 sek
Surgical hand disinfection	EN 12791	2 x 3 ml in time 120 sek.
ECBO (Bovine enterovirus) Veterinray	14675	30 min/10°C*

* mandatory conditions

Velox Oxy ETA

A ready-to-use sporidical preparation for disinfection and cleaning of non-invasive equipment and small surfaces in high-risk areas



Spectrum (effectiveness)	EN 14885	Dirty Test conditions
Bacteria	EN 13727	15 sec.
	EN 16615	30 sec.
	EN 13697	60 sec.*
Yeast (<i>C. albicans</i>)	EN 13624	15 sec.
	EN 16615	30 sec.
	EN 13697	60 sec.
Fungi (<i>A. brasiliensis</i>)	EN 13624	30 sec.
	EN 16615	30 sec.
	EN 13697	60 sec.
Mycobacteria (<i>M. terrae</i>, <i>M. avium</i>)	EN 14348	3 min.
Enveloped viruses (Vaccinia)	EN 14476	30 sec.
Adenovirus	EN 14476	30 sec.
Norovirus	EN 14476	15 sec.
Poliovirus	EN 14476	15 sec.
Spores (<i>B. subtilis</i>)	EN 13704	10 min.
Spores (<i>C. difficile</i> R027)	EN 17126	15 min.

* clean conditions

New performed studies are marked blue