



STATEMENT OF TEST RESULT

No.: MED GZME1610000845ME

Applicant: SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD.
Keji 12th Road South, Hi-tech Industrial Park, Nan Shan
Shenzhen 518057, P.R. China

Manufacturer: Same as applicant

Product Description: Defibrillator/Monitor

Model No.: BeneHeart D5, BeneHeart D6

Rating: 100-240 V~; 50/60 Hz; 1.8-0.8 A
Rechargeable battery (in Main unit): 14.8 Vd.c. 4500 mAh or 15.1 Vd.c. 5600 mAh
Battery charger BatteryFeed 20: 100-240 V~; 50/60 Hz; 1.0-0.5 A
AC docking station: 100-240 V~; 50/60 Hz; 1.8-0.8 A
DC docking station: 12 V d.c.; 15 A

Additional Information (if any): --

Sufficient samples of the product have been tested and found to be in conformity with

Test Standard: EN 1789: 2007 + A2: 2014

as shown in the
Test Report Number(s): GZME161000084501

Conclusion:

The submitted samples fulfilled the requirements of specified standard which can be evaluated.



2016-10-17

Copyright of this statement is owned by SGS-CSTC Standards Technical Services Co., Ltd. and may not be reproduced other than in full and with the prior approval of the General Manager. This verification is subjected to the governance of the General Conditions of Services, printed overleaf.

Member of SGS Group (Société Générale de Surveillance)