

Product Specifications

CL03.0204

HYDREX

Buffer pH 10.00

Tampão pH 10.00
Buffer pH 10.00
Tampon pH 10.00
Puffer pH 10.00
Tampón pH 10.00
Tampone pH 10.00

**For laboratory use, pH measurements, traceable to SRM from NIST
pH 10.00 ($\pm 0.02/20^{\circ}\text{C}$) H₃BO₃-KCl-NaOH**

**Traceability: Reference standards acc. NIST SRM 189c, 188, 185i,
186lg, 186llg, 187e**

Mol weight:
Density: 1,01 g/ml
HS Nr: 38229000
CAS Nr:
UN:
ADR:

Specifications:

pH	= 10.24 @ 0°C
pH	= 10.10 @ 10°C
pH	= 10.00 @ 20°C
pH	= 9.95 @ 25°C
pH	= 9.90 @ 30°C
pH	= 9.82 @ 40°C
pH	= 9.75 @ 50°C
pH	= 9.68 @ 60°C
pH	= 9.62 @ 70°C
pH	= 9.55 @ 80°C
pH	= 9.49 @ 90°C