

EU DECLARATION OF CONFORMITY
MDR 2017/745

TITANOX S.r.l.

Via Canove de' Biazzi - 26038 TORRE DE' PICENARDI (CR) - Tel: 0375394065
E-mail: info@titanox.it - SRN: IT-MF-000020325

Hereby declare under their own responsibility that

MEDICAL DEVICES CALLED
" X-RAY FILM VIEWER",

BASIC UDI-DI: 805930470M05H4

MANUFACTURED BY TITANOX S.R.L.,

ARE COMPLIANT WITH THE ESSENTIAL SAFETY AND PERFORMANCE
REQUIREMENTS SET FORTH IN ANNEX I OF THE MEDICAL DEVICES
REGULATION 2017/745.

such medical device is complying with all the applicable requirements of the Medical Devices Regulation 2017/745, in particular:

- That the medical device above belongs to Class I, according to Annex VIII of the Medical Devices Regulation 2017/745, rule 1;
- That the medical device above HAS NO MEASURING FUNCTION;
- That the medical device above IS NOT TO BE USED FOR CLINICAL INVESTIGATION;
- That the medical device above IS NOT MARKETED IN A STERILIZED PACKAGE;

We certify that the devices also comply:

DIR. 2014/30/UE – Electromagnetic compatibility (EMC)

DIR. 2014/35/UE – Low voltage (LVD)

DIR. 2011/65/UE – Directive ROHS II

DIR. 2015/863/UE – Modified Ann. II Directive ROHS II

Torre de' Picenardi 12.05.2022

.....
(place and date of issue)


Sole Administrator
Enrico Basatti

.....
(name and signature or stamping of the authorized person)

Annex list of devices covered by this declaration

Code	Description
M604043	X-RAY FILM VIEWER 40X43
M604043/M	X-RAY FILM VIEWER 40X43 1 PANEL WITH LIGHT DIMMER
M608043	X-RAY FILM VIEWER 80X43
M608043/GM	X-RAY FILM VIEWER 80X43 2 PANELS WITH LIGHT DIMMER
M612043	X-RAY FILM VIEWER 120X43
M612043/GM	X-RAY FILM VIEWER 120X43 3 PANELS WITH LIGHT DIMMER
M612043/V	X-RAY FILM VIEWER 120X43 VERTICAL
M616043	X-RAY FILM VIEWER 160X43