

Certificate of Conformity

VWR Catalogue Number	See List
Description	Blender Bags
Country of Origin	Manufactured in China
Date of Issue (yyyy-mm-dd)	2023-10-09

Catalogue Number	Type	Capacity [mL]	Dimension [mm]
129-0732	Plain bags	80	105×150
129-0728	Plain bags	400	300×190
129-0729	Plain bags	400	300×190
129-0731	Plain bags	3500	380×510
129-0733	Bags with lateral filter	400	300×190
129-0730	Bags with lateral filter	400	300×190
129-0734	Bags with lateral filter	400	300×190

Quality System Compliance

Products are manufactured under the **ISO 13485:2016** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Blending Test	:	Bags have to be resistant for 3 minutes blending (Frequency:times/second).
Welding strength	:	> 12N/15 mm.
Tensile Strength	:	> 25MP/10 mm.
Elongation	:	>600%.
Porosity	:	<250µm

Product Specifications

Material	<p>For Plain Bags: Mixture of 100% Virgin Raw material Low Density Polyethylene (LDPE) + High Density Polyethylene (HDPE).</p> <p>For Bags with lateral filter: 100% virgin raw material (PA + PE+Non-woven polyeste).</p> <p>Raw material used to produce these products are in accordance with FDA regulation 21 CFR § 177.1520, for use in contact with food applications</p>
Sterilization	<p>Product labeled as sterile is gamma irradiated and dose released upon ISO11137 recommended practices in effect at the time of validation. Products labeled sterile meet a minimum requirement of 10⁻³ SAL with a specified dose range of 10–40 kGy.</p>
ATP Assay	Not applicable.
DNase & RNase Free	Not applicable.
BSE/TSE	<p>No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).</p>
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	Not applicable.
BPA Statement	Bisphenol are not used in the manufacture of the raw material and are not expected to be present.
DEHP Statement	Not applicable.
RoHS	<p>No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed.</p>
REACH Statement	Not applicable.

Storage Conditions Room temperature.

Shelf Life 3 years.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.

Certificate of Conformity

VWR Blender Bag with Full Page Filter, Sterile

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog. No: 129-0734

Lot No: 12292020

Date of Manufacture: 29th Dec 2020

Date of Sterilization: 04th Jan 2021

Expiry Date: 29th Dec 2023

Capacity (ml)	Description	WxL (mm)	Packed	Pk
400	Bags with full page filter	300x190	10 bags of 50	500

Quality System Compliance: Products are manufactured under the ISO 13485:2016 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

Blending test: Bags have to be resistant for 3 minutes blending (frequency: 4 times/second)

Physical requirements:

Welding Strength: > 12N/15 mm

Tensile Strength: >25MP/10 mm

Elongation: >600 %

Material: 100% Virgin raw material (PA + PE+ Non-woven polyester)

Certificate of Conformity

Sterility: Products labeled as sterile is gamma irradiated and dose released upon ISO11137 recommended practices meets a minimum requirement of 10^{-3} SAL with a specified dose range of 10 - 40 kGy. Expiration date (Shelf life) is based on product and packaging shelf life, provided the product wrapper is not damaged.

BSE/TSE Statement: The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/ Bovine Spongiform Encephalopathy (BSE)

BPA Statement: Bisphenol-A (BPA) and Bisphenol-F (BPF) are not used in the manufacture or the formulation of the raw materials and are not expected to be present. However, tests have not been performed for these chemical materials.

DEHP Statement: Not Applicable.

Heavy metals - RoHS: Raw materials comply with the requirements of the Directive 2002/95/EC, as amended, and 2011/65/Eu concerning the limits of cadmium, lead, mercury, hexavalent chromium, Polybrominated Biphenyls (PBB) and Polybrominated Diphenyl Ethers (PBDE).

This document has been produced electronically and is valid without a signature

Date of Issue: **August 2021**

Product Specification Sheet

VWR PLAIN BLENDER BAGS STERILE

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No.	Description	Capacity (ml)	Height (mm)	Width (mm)	Thickness (mm)	Pk
129-0728	Plain Blender bags sterile	400	300±3	180±3	0.07±0.005	20bags of 25pcs
129-0729	Plain Blender bags sterile	400	300±3	180±3	0.07±0.005	10 bags of 50pcs
129-0731	Plain Blender bags sterile	3500	510±3	380±3	0.07±0.005	20 bags of 25pcs
129-0732	Plain Blender bags sterile	80	150±3	105±3	0.07±0.005	60 bags of 25pcs

Material: Natural LDPE (Low Density Polyethylene (PE)) + HDPE (High Density Polyethylene (PE))

Product Requirements

Absence of heavy material

Raw material is compliant with the EC/94/62 directive

Food contact

Raw material is compliant with regulation EC N°1935/2004

Based on ISO 11137

Sterilization of health care products -- Radiation, SAL10⁻³

Blender test

Bags have to be resistant for 3 minutes blending (frequency: 4 times / second)

Use temperature main body

-80°C / +75°C

Expiry: 3 years

Product Specification Sheet

Physical Requirements

Flexible, good mechanical resistance.

Packing Requirements

Bags are zipper bag wrap and then heat-seal.
Color change label when sterile.

Disclaimer: VWR wants to state that these specifications can only be regarded as general information and will not discharge the user from his obligation to ensure the product is suitable for the intended use - If there is any doubt ask us.

Date of issue: **June 2013**