

# Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016

This is to certify that:

ASAP International Sdn. Bhd.  
No.1 , Jalan Sitar 33/6  
Seksyen 33  
Shah Alam  
Selangor  
40400  
Malaysia

Holds Certificate No:

**MD 676150**

and operates a Quality Management System which complies with the requirements of ISO 13485:2016 for the following scope:

Manufacture, Marketing and Distribution of latex and nitrile examination gloves, face masks and sterile surgical gloves.

For and on behalf of BSI:



Graeme Tunbridge, Senior Vice President Medical Devices

Original Registration Date: 2017-12-13

Latest Revision Date: 2022-04-08

Effective Date: 2022-04-23

Expiry Date: 2025-04-22

Page: 1 of 1



...making excellence a habit.™

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated [online](#).

Printed copies can be validated at [www.bsi-global.com/ClientDirectory](http://www.bsi-global.com/ClientDirectory) or telephone +603 9212 9638.

Further clarifications regarding the scope of this certificate and the applicability of ISO 13485:2016 requirements may be obtained by consulting the organization.

This certificate is valid only if provided original copies are in complete set.

# Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 9001:2015

This is to certify that:

ASAP International Sdn. Bhd.  
No.1 , Jalan Sitar 33/6  
Seksyen 33  
Shah Alam  
Selangor  
40400  
Malaysia

Holds Certificate No:

**FM 679348**

and operates a Quality Management System which complies with the requirements of ISO 9001:2015 for the following scope:

Manufacture, Marketing and Distribution of latex and nitrile examination gloves, face masks and sterile surgical gloves.

For and on behalf of BSI:

Poon Cheong Yuen, Managing Director

Original Registration Date: 2017-11-13

Latest Revision Date: 2022-03-28

Effective Date: 2022-04-25

Expiry Date: 2025-04-24

Page: 1 of 1



...making excellence a habit.™

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated [online](#).

Printed copies can be validated at [www.bsi-global.com/ClientDirectory](http://www.bsi-global.com/ClientDirectory) or telephone +603 9212 9638.

Further clarifications regarding the scope of this certificate and the applicability of ISO 9001:2015 requirements may be obtained by consulting the organization.

This certificate is valid only if provided original copies are in complete set.

Malaysia Headquarters: Suite 29.01, Level 29, The Gardens North Tower, Mid Valley City, Lingkaran Syed Putra, 59200 Kuala Lumpur.

A Member of the BSI Group of Companies.



# ASAP INTERNATIONAL SDN. BHD. (1157980-X)

No. 1, Jalan Sitar 33/6, Seksyen 33, 40400 Shah Alam, Selangor, Malaysia.

Tel: +6 03 5191 0166

Fax: +6 03 5191 0702

Website: www.whyasap.com

Email: info@whyasap.com

## EU DECLARATION of CONFORMITY



Manufacturer:

**ASAP International Sdn. Bhd.**

No. 1, Jalan Sitar 33/6, Seksyen 33,  
40400 Shah Alam, Selangor, Malaysia.

SRN: MY-MF-000004500

Medical Devices &  
Basic UDI- DI:

- Powdered Latex Examination Gloves  
(955589090ASAPPPL5E)
- Powder Free Latex Examination Gloves  
(955589090ASAPPFL4G)
- Powder Free Nitrile Examination Gloves  
(955589090ASAPPFN4L)

Classification:

Class I Medical Devices according to Medical Device  
Regulation (EU) 2017/745, Annex VIII

Harmonised Standards:

EN455-1:2000, EN 455-2:2015, EN455-3:2015,  
EN455-4:2009



European Authorized  
Representative:

ASAP INNOVATIONS LIMITED

7 Saggart Lakes, Saggart,  
Dublin, D24 PY01, Ireland.

SRN: IE-AR-000002548

This EU declaration of conformity is issued under the sole responsibility of manufacturer, ASAP International Sdn. Bhd. We hereby declare that device covered by the present declaration is in conformity with the Medical Device Regulation (EU) 2017/745 and with the above-mentioned standards. All supporting documentation is retained at the premises of the manufacturer.

Place, Date of Declaration: Malaysia., May 26, 2021

Signature:

Mr. Chin Tze Weng  
Senior Global Business Development Manager



# EU Type-Examination Certificate

## Certificate number: 2777/11077-01/E03-01

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:

Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

**Product reference:** AS LPF  
**Description:** Latex Examination powder free gloves.

<b>Sizes:</b>		<b>Classification:</b>		
6	XS	<b>EN ISO 374-1:2016/Type B</b>	<b>Level</b>	<b>EN374-4:2013</b>
7	S	40% Sodium Hydroxide	1	-46.1%
8	M	65% Nitric Acid	2	30.9%
9	L	37% Formaldehyde	2	-93.8%
10	XL	40% Hydrofluoric Acid	2	X
		<b>EN ISO 374-5:2016</b>		
		Resistance to Bacteria and Fungi	Pass	
		Resistance to Virus	Pass	

Standards/Technical specifications applied:  
EN 420: 2003+A1: 2009; EN ISO 374-1:2016; EN ISO 374-5:2016

Technical reports/Approval documents:  
SATRA: CHM0265590/1750/CL/A, CHM0265590/1750/CL/B, CHM0265590/1750/SPT  
TUV: 7191154274-CHM16-01-RC

Signed on behalf of SATRA:



Hannah Coe



Geoff Graham

**Date of issue:** 17/04/2019

**Expiry date:** 14/08/2023

# TERMS AND CONDITIONS

The following conditions apply in addition to SATRA's standard terms and conditions of business and those given in the current certification agreement.

The certificate holder is licensed to mark the products detailed within this certificate in accordance with Annex V (Module B) of the Regulation (EU) 2016/425 of the European Parliament and of the council of 9th March 2016 on personal protective equipment once you have drawn up an EU declaration of product conformity. Please note:

1. Where the product is classified as category III then CE Marking of production is reliant on current compliance with Regulation 2016/425 module C2 or Module D. (Except that specifically produced to fit an individual user).
2. Full details of the certification and product are contained within the manufacturer's technical documentation.
3. Where a translation of this certificate exists, the English language version shall be considered as the authoritative text.
4. Certification is limited to production undertaken at the sites listed in the manufacturers technical documentation.
5. Ongoing manufactured product shall be consistent with the product(s) certified and listed on this certificate.
6. The Manufacturer shall inform SATRA of any changes to the certified product or technical documentation.
7. This certificate shall be kept together with the relevant technical documentation in a safe place by the client named on this certificate. Production of this certificate and other documentation may be required by a representative of the EC member state government.
8. This certificate relates only to the condition of the testable items at the time of the certification procedure and is subject to the expiry date shown.
9. SATRA Technology reserves the right to withdraw this certificate if it is found that a condition of manufacture, design, materials or packaging have been changed and therefore no longer comply with the requirements of Regulation 2016/425.

# EU Type-Examination Certificate

## Certificate number: 2777/10648-04/E04-01

This EU Type-Examination Certificate covers the following product group(s) supported by testing to the relevant standards/technical specifications and examination of the technical file documentation:  
Following the EU Type-Examination this product group has been shown to satisfy the applicable essential health and safety requirements of Annex II of the PPE Regulation (EU) 2016/425 as a Category III product.

**Product reference:**

AS NPF

**Description:**

Nitrile examination powder free gloves

**Sizes:**

6 (XS) – 10 (XL)

**Classification:**

EN ISO 374-1:2016/Type B	Level	EN374-4:2013
37% Formaldehyde	6	3.1%
40% Sodium Hydroxide	6	-25.6%
30% Hydrogen Peroxide	2	17.0%

**EN ISO 374-5:2016**

Resistance to Bacteria and Fungi	Pass
Resistance to Virus	Pass


Standards/Technical specifications applied:

EN 420: 2003+A1: 2009; EN ISO 374-1:2016; EN ISO 374-5:2016

Technical reports/Approval documents:

SATRA: CHM0265112/1749/EN/A, CHM0265112/1749/EN/B, CHM0265112/1749/SPT, CHM0272621/1826/JS, CHM0275215/1836/LH, CHM0275215/1836/LH/E, CHM0275215/1836/LH/D, CHM0275215/1836/LH/A/Final  
TUV: 7191143339-CHM16-01-RC

Signed on behalf of SATRA:



Hannah Coe



Geoff Graham

Date of issue: 17/04/2019

Expiry date: 25/06/2023

# TERMS AND CONDITIONS

The following conditions apply in addition to SATRA's standard terms and conditions of business and those given in the current certification agreement.

The certificate holder is licensed to mark the products detailed within this certificate in accordance with Annex V (Module B) of the Regulation (EU) 2016/425 of the European Parliament and of the council of 9th March 2016 on personal protective equipment once you have drawn up an EU declaration of product conformity. Please note:

1. Where the product is classified as category III then CE Marking of production is reliant on current compliance with Regulation 2016/425 module C2 or Module D. (Except that specifically produced to fit an individual user).
2. Full details of the certification and product are contained within the manufacturer's technical documentation.
3. Where a translation of this certificate exists, the English language version shall be considered as the authoritative text.
4. Certification is limited to production undertaken at the sites listed in the manufacturers technical documentation.
5. Ongoing manufactured product shall be consistent with the product(s) certified and listed on this certificate.
6. The Manufacturer shall inform SATRA of any changes to the certified product or technical documentation.
7. This certificate shall be kept together with the relevant technical documentation in a safe place by the client named on this certificate. Production of this certificate and other documentation may be required by a representative of the EC member state government.
8. This certificate relates only to the condition of the testable items at the time of the certification procedure and is subject to the expiry date shown.
9. SATRA Technology reserves the right to withdraw this certificate if it is found that a condition of manufacture, design, materials or packaging have been changed and therefore no longer comply with the requirements of Regulation 2016/425.

**Test Report No : CPSA/210663477-CA64942**  
**Company : ASAP International Sdn. Bhd.**  
**No.1, Jalan Sitar 33/6, Seksyen 33,**  
**40400 Shah Alam, Selangor,**  
**Malaysia.**

**TEST REPORT**

Sample Description : Powder Free Latex Examination Gloves,  
Device Basic UDI-DI: 955589090ASAPPFL4G  
Brand Name : ASAP  
Country of Origin : Malaysia  
Country of Destination : All countries  
Size : M  
Quantity Tested : 200 pieces  
Test Conducted : Freedom from holes  
Test Method : EN455 Part 1:2020  
Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained :-

Acceptable Quality Limit (AQL) : 1.5      Accept : 7      Found : 4

Result : Within AQL

SIGNED FOR AND ON BEHALF OF  
SGS (MALAYSIA) SDN BHD

\*\*\*\*End of Report\*\*\*\*



CHEE TUCK CHOON  
SECTION HEAD  
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.



**Test Report No : CPSA/210663478-CA64943**  
**Company : ASAP International Sdn. Bhd.**  
**No.1, Jalan Sitar 33/6, Seksyen 33,**  
**40400 Shah Alam, Selangor,**  
**Malaysia.**

**TEST REPORT**

Sample Description : Powder Free Latex Examination Gloves,  
 Device Basic UDI-DI: 955589090ASAPPFL4G  
 Brand Name : ASAP  
 Country of Origin : Malaysia  
 Country of Destination : All countries  
 Size : M  
 Quantity Tested : 13 pieces per test  
 Test Conducted : Force at Break During Shelf Life and After Challenge  
 Test Method : EN 455 Part 2:2015  
 Ageing : 70 ± 2 Deg C for 168 hrs  
 Testing Period : 21 May 2021 – 04 June 2021

SIZE	SAMPLE NO.	Force at Break, N	
		<u>BEFORE AGING</u>	<u>AFTER AGING</u>
M	1	10.7	8.8
	2	10.7	10.0
	3	10.3	9.1
	4	11.1	9.5
	5	9.8	9.1
	6	10.6	9.8
	7	10.1	9.1
	8	10.7	9.1
	9	10.0	9.4
	10	10.7	10.0
	11	10.7	9.1
	12	9.4	8.4
	13	8.9	9.0
Median		10.6	9.1
Requirement		≥ 6.0	≥ 6.0

\*\*\*\*End of Report\*\*\*\*

SIGNED FOR AND ON BEHALF OF  
 SGS (MALAYSIA) SDN BHD



CHEE TUCK CHOON  
 SECTION HEAD  
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

**Test Report No : CPSA/210663479-CA64944**  
**Company : ASAP International Sdn. Bhd.**  
**No.1, Jalan Sitar 33/6, Seksyen 33,**  
**40400 Shah Alam, Selangor,**  
**Malaysia.**

**TEST REPORT**

Sample Description : Powder Free Latex Examination Gloves,  
 Device Basic UDI-DI: 955589090ASAPPFL4G  
 Brand Name : ASAP  
 Country of Origin : Malaysia  
 Country of Destination : All countries  
 Size : M  
 Quantity Tested : 13 pieces  
 Test Conducted : Dimensions  
 Test Method : EN 455 Part 2:2015  
 Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width Median: 95±10mm	95	94	95	95	94	95	94	95	95	95	95	94	95	95
Length Median: ≥ 240mm	251	248	251	252	252	254	250	252	252	251	251	250	251	251

SIGNED FOR AND ON BEHALF OF  
 SGS (MALAYSIA) SDN BHD

\*\*\*\*End of Report\*\*\*\*



CHEE TUCK CHOON  
 SECTION HEAD  
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

**Test Report No** : CPSA/210663480-CA64945  
**Company** : ASAP International Sdn. Bhd.  
 No.1, Jalan Sitar 33/6, Seksyen 33,  
 40400 Shah Alam, Selangor,  
 Malaysia

**TEST REPORT**

Sample Description : Powder Free Latex Examination Gloves,  
 Device Basic UDI-DI: 955589090ASAPPFL4G  
 Brand Name : ASAP  
 Country of Origin : Malaysia  
 Country of Destination : All countries  
 Size : M  
 Quantity Tested : 8 pieces  
 Test Conducted : Aqueous Extractable Protein Content  
 Test Method : EN455 Part 3:2015  
 Testing Period : 21 May 2021 – 04 June 2021

**EXTRACTABLE PROTEIN CONTENT**

Analysis had been carried out as per your request. We report the following results:

SIZE	AQUEOUS EXTRACTABLE PROTEIN CONTENT, µg/g
M	23

Note: Detection limit = 10 µg/g

SIGNED FOR AND ON BEHALF OF  
 SGS (MALAYSIA) SDN BHD

\*\*\*End of Report\*\*\*



CHEE TUCK CHOON  
 SECTION HEAD  
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

**Test Report No : CPSA/210663481-CA64946**  
**Company : ASAP International Sdn. Bhd.**  
**No.1, Jalan Sitar 33/6, Seksyen 33,**  
**40400 Shah Alam, Selangor,**  
**Malaysia.**

**TEST REPORT**

Sample Description : Powder Free Latex Examination Gloves,  
Device Basic UDI-DI: 955589090ASAPPFL4G  
Brand Name : ASAP  
Country of Origin : Malaysia  
Country of Destination : All countries  
Size : M  
Quantity Tested : 5 pieces  
Test Conducted : Powder Content  
Test Method : EN455 Part 3:2015  
Testing Period : 21 May 2021 – 04 June 2021

On testing the samples, the following results were obtained:-

**SIZE**

**Average Powder Mass per Glove**

M

0.12 mg

SIGNED FOR AND ON BEHALF OF  
SGS (MALAYSIA) SDN BHD

\*\*\*\*End of Report\*\*\*\*



CHEE TUCK CHOON  
SECTION HEAD  
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

**Test Report No** : CPSA/210663482-CA64947  
**Company** : ASAP International Sdn. Bhd.  
No.1, Jalan Sitar 33/6, Seksyen 33,  
40400 Shah Alam, Selangor,  
Malaysia

**TEST REPORT**

Sample Description : Powder Free Nitrile Examination Gloves,  
Device Basic UDI-DI: 955589090ASAPPFN4L  
Brand Name : ASAP  
Country of Origin : Malaysia  
Country of Destination : All countries  
Size : M  
Quantity Tested : 200 pieces  
Test Conducted : Freedom from holes  
Test Method : EN455 Part 1:2020  
Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained:-

Acceptable Quality Limit (AQL) : 1.5      Accept : 7      Found : 3

Result : Within AQL

SIGNED FOR AND ON BEHALF OF  
SGS (MALAYSIA) SDN BHD

\*\*\*\* End of Report \*\*\*\*



CHEE TUCK CHOON  
SECTION HEAD  
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

**Test Report No** : CPSA/210663483-CA64948  
**Company** : ASAP International Sdn. Bhd.  
 No.1, Jalan Sitar 33/6, Seksyen 33,  
 40400 Shah Alam, Selangor,  
 Malaysia

**TEST REPORT**

Sample Description : Powder Free Nitrile Examination Gloves,  
 Device Basic UDI-DI: 955589090ASAPPFN4L  
 Brand Name : ASAP  
 Country of Origin : Malaysia  
 Country of Destination : All countries  
 Size : M  
 Quantity Tested : 13 pieces per test  
 Test Conducted : Force at Break During Shelf Life and After Challenge  
 Test Method : EN 455 Part 2:2015  
 Ageing : 70 ± 2 Deg C for 168 hrs  
 Testing Period : 21 May 2021 – 04 June 2021

SIZE	SAMPLE NO.	Force at Break, N	
		<u>BEFORE AGING</u>	<u>AFTER AGING</u>
M	1	8.4	8.1
	2	8.6	8.3
	3	9.1	8.9
	4	7.5	9.8
	5	8.2	8.8
	6	8.6	9.9
	7	8.9	8.9
	8	8.7	8.2
	9	8.1	8.2
	10	7.9	8.6
	11	9.4	9.1
	12	8.9	9.3
	13	8.5	9.3
Median		8.6	8.9
Requirement		≥ 6.0	≥ 6.0

SIGNED FOR AND ON BEHALF OF  
 SGS (MALAYSIA) SDN BHD

\*\*\*\* End of Report \*\*\*\*



CHEE TUCK CHOON  
 SECTION HEAD  
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

**Test Report No** : CPSA/210663484-CA64949  
**Company** : ASAP International Sdn. Bhd.  
 No.1, Jalan Sitar 33/6, Seksyen 33,  
 40400 Shah Alam, Selangor,  
 Malaysia

**TEST REPORT**

Sample Description : Powder Free Nitrile Examination Gloves,  
 Device Basic UDI-DI: 955589090ASAPPFN4L  
 Brand Name : ASAP  
 Country of Origin : Malaysia  
 Country of Destination : All countries  
 Size : M  
 Quantity Tested : 13 pieces  
 Test Conducted : Dimensions  
 Test Method : EN 455 Part 2:2015  
 Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained :-

Size	M	M	M	M	M	M	M	M	M	M	M	M	M	Median
Width Median: 95±10mm	96	96	96	96	96	96	96	96	96	95	95	96	96	96
Length Median: ≥ 240mm	245	240	246	244	241	245	241	246	242	240	248	249	245	245

SIGNED FOR AND ON BEHALF OF  
 SGS (MALAYSIA) SDN BHD

\*\*\*\* End of Report \*\*\*\*



CHEE TUCK CHOON  
 SECTION HEAD  
 IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.

**Test Report No** : CPSA/210663485-CA64950  
**Company** : ASAP International Sdn. Bhd.  
No.1, Jalan Sitar 33/6, Seksyen 33,  
40400 Shah Alam, Selangor,  
Malaysia

**TEST REPORT**

Sample Description : Powder Free Nitrile Examination Gloves,  
Device Basic UDI-DI: 955589090ASAPPFN4L  
Brand Name : ASAP  
Country of Origin : Malaysia  
Country of Destination : All countries  
Size : M  
Quantity Tested : 5 pieces  
Test Conducted : Powder Content  
Test Method : EN455 Part 3:2015  
Testing Period : 21 May 2021 – 04 June 2021

On testing the samples, the following results were obtained:-

**SIZE**

**Average Powder Mass per Glove**

M

0.08 mg

SIGNED FOR AND ON BEHALF OF  
SGS (MALAYSIA) SDN BHD

\*\*\*\* End of Report \*\*\*\*



CHEE TUCK CHOON  
SECTION HEAD  
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.